

LB 99926


245.1

INSTRUMENT ID : CV1

Linear

 μ Abs.:

37574



A= 0.0000e+000

B= 2.6696e-004

C= -1.7546e-002

Rho= 0.9999798

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.010	-0.010	28	0.000	28				
0.05	0.050	0.052	0.002	260	0.0 %	260				
0.2	0.200	0.199	-0.001	810	0.0 %	810				
2.5	2.500	2.502	0.002	9437	0.0 %	9437				
5.0	5.000	5.040	0.040	18944	0.0 %	18944				
7.5	7.500	7.455	-0.045	27991	0.0 %	27991				
10.0	10.000	10.013	0.013	37574	0.0 %	37574				

LB99926 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 19 Dec 2018 13:32:50

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	28	-	0.0000	245.1	PPB	19 Dec 2018 13:35:15	S	Std
0.05 - 1	50.05	260	-	0.0500	245.1	PPB	19 Dec 2018 13:37:33	S	Std
0.2 - 1	50.2	810	-	0.2000	245.1	PPB	19 Dec 2018 13:39:52	S	Std
2.5 - 1	52.15	9437	-	2.5000	245.1	PPB	19 Dec 2018 13:42:09	S	Std
5.0 - 1	55	18944	-	5.0000	245.1	PPB	19 Dec 2018 13:44:27	S	Std
7.5 - 1	57.5	27991	-	7.5000	245.1	PPB	19 Dec 2018 13:46:43	S	Std
10.0 - 1	510	37574	-	10.0000	245.1	PPB	19 Dec 2018 13:49:00	S	Std
ICV33 - 1	ICV33	14402	3.8272	-245.1		PPB	19 Dec 2018 13:54:28	U	SMPL
ICB33 - 1	ICB33	-7	-0.0194	-245.1		PPB	19 Dec 2018 13:56:43	U	SMPL
CCV50 - 1	CCV50	18217	4.8456	-245.1		PPB	19 Dec 2018 13:59:00	U	SMPL
CCB50 - 1	CCB50	-12	-0.0207	-245.1		PPB	19 Dec 2018 14:01:15	U	SMPL
CRA - 1	CRA	241	0.0468	-245.1		PPB	19 Dec 2018 14:03:31	U	SMPL
HighStd - 1	HighStd	37393	9.9648	-245.1		PPB	19 Dec 2018 14:05:47	U	SMPL
ChkStd - 1	ChkStd	26243	6.9883	-245.1		PPB	19 Dec 2018 14:08:04	U	SMPL
PB115778BL - 1	PBW	-158	-0.0597	-245.1		PPB	19 Dec 2018 14:10:21	U	SMPL
PB115778BS - 1	LCSW	14294	3.7984	-245.1		PPB	19 Dec 2018 14:17:32	U	SMPL
J6417-01 - 1	121318-TIPC-F-D-B	-164	-0.0613	-245.1		PPB	19 Dec 2018 14:19:49	U	SMPL
J6417-02 - 1	121318-TIPC-F-D-S	-96	-0.0432	-245.1		PPB	19 Dec 2018 14:22:06	U	SMPL
J6417-03 - 1	121318-TIPC-F-U-B	-87	-0.0408	-245.1		PPB	19 Dec 2018 14:24:24	U	SMPL
J6417-04 - 1	121318-TIPC-F-U-S	-87	-0.0408	-245.1		PPB	19 Dec 2018 14:26:41	U	SMPL
J6417-04DUP - 1	121318-TIPC-F-U-SDUP	-79	-0.0386	-245.1		PPB	19 Dec 2018 14:28:57	U	SMPL
CCV51 - 1	CCV51	19267	5.1259	-245.1		PPB	19 Dec 2018 14:31:13	U	SMPL
CCB51 - 1	CCB51	-3	-0.0183	-245.1		PPB	19 Dec 2018 14:33:30	U	SMPL
J6417-04MS - 1	121318-TIPC-F-U-SMS	15461	4.1099	-245.1		PPB	19 Dec 2018 14:35:47	U	SMPL
J6417-04MSD - 1	121318-TIPC-F-U-SMSD	15405	4.0950	-245.1		PPB	19 Dec 2018 14:38:03	U	SMPL
J6478-01 - 1	121718-JETT-E-D-B	-116	-0.0485	-245.1		PPB	19 Dec 2018 14:40:19	U	SMPL
J6478-02 - 1	121718-JETT-E-D-S	-31	-0.0258	-245.1		PPB	19 Dec 2018 14:42:36	U	SMPL
J6478-03 - 1	121718-JETT-E-U-B	-26	-0.0245	-245.1		PPB	19 Dec 2018 14:44:53	U	SMPL
J6478-04 - 1	121718-JETT-E-U-S	-25	-0.0242	-245.1		PPB	19 Dec 2018 14:47:11	U	SMPL
J6417-04LX5 - 1		22	-0.0117	-245.1		PPB	19 Dec 2018 14:49:28	U	SMPL
J6417-04A - 1		15388	4.0904	-245.1		PPB	19 Dec 2018 14:51:45	U	SMPL
CCV52 - 1	CCV52	18977	5.0485	-245.1		PPB	19 Dec 2018 14:54:04	U	SMPL
CCB52 - 1	CCB52	-11	-0.0205	-245.1		PPB	19 Dec 2018 14:56:19	U	SMPL