



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: sweta

Run Number: LB99927

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.6	12/19/2018	11:25
2	J6472-04		1	9	>100.0	>212.0	>212.0	12/19/2018	11:55
3	J6472-04DUP		1	9	>100.0	>212.0	>212.0	12/19/2018	12:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 99927

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp121918

WorkList ID : 120585

Date : 12/19/2018 10:51:09 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/21/2018	Water	J6472-04 C	Flash Point	Cool 4 deg C	PSEG01	E31	OILY-WATER	12/18/2018	1010A

Date/Time 12/19/18 10:55 AM
Received by: KS
Relinquished by: CP

Date/Time 12/19/18 12:25 PM
Received by: CP
Relinquished by: KS