



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 12/20/2018

OVENTEMP IN Celsius(°C): 108
Time IN: 18:15
In Date: 12/19/2018
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:05
Out Date: 12/20/2018
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1

QC:LB99937

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
J6481-01	VB-1	1	1.13	9.55	8.58	88.5
J6481-02	SW-PROP-2	2	1.15	9.95	7.99	77.7
J6481-03	PROP-SW-1	3	1.19	9.73	8.05	80.3
J6481-04	HOUSE-WALL	4	1.11	9.94	8.84	87.5
J6482-01	R20-OS	5	1.00	2.00	2.00	100.0
J6482-02	R20-OS-A	6	1.00	2.00	2.00	100.0
J6482-03	R20-OS-B	7	1.00	2.00	2.00	100.0
J6482-04	R20-OS	8	1.00	2.00	2.00	100.0
J6482-05	R20-CONCRETE	9	1.00	2.00	2.00	100.0
J6482-06	DPC-91468-OL-01	10	1.00	2.00	2.00	100.0
J6482-07	DPC-91468-OD-01	11	1.00	2.00	2.00	100.0
J6482-08	DPC-91468-OD-02	12	1.00	2.00	2.00	100.0
J6482-09	DPC-91468-OS-01	13	1.00	2.00	2.00	100.0
J6482-10	DPC-91468-OS-01-A	14	1.00	2.00	2.00	100.0
J6482-11	DPC-91468-OS-01-B	15	1.00	2.00	2.00	100.0
J6482-12	DPC-91468-OS-01	16	1.00	2.00	2.00	100.0
J6490-01	72-11929	20	1.00	2.00	2.00	100.0
J6493-01	SOIL-DRUM	17	1.13	9.77	8.82	89.0
J6493-02	SOIL-DRUM	18	1.13	9.77	8.82	89.0
J6493-03	OILY-DEBRIS	19	1.00	2.00	2.00	100.0

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$