



Analytical Summary Report

Analysis Method: 7196A
Parameter: Hexavalent Chromium
Run Number: LB99938

ANALYST: Indiana
SUPERVISOR REVIEW BY: apatel
pH Meter ID: WC pH Meter-1

Table with 2 columns: Reagent/Standard, Lot/Log #. Rows include Calibration Std. hexchrome 0.1 ppm, Hexavalent Chromium LCS Std, Hexavalent Chromium ICV Std, etc.

Intercept: 0.0001 Slope: 0.7712 Regression: 0.999984

Table with 12 columns: Seq, Lab ID, True Value (mg/l), DF, Initial Vol (ml/gm), Final Vol (ml), pH HN03, pH H2SO4, Absorb. at 540 (Backgrnd, Color), Absorbance Difference, Anal Date, Anal Time. Contains 7 rows of calibration data.



Analytical Summary Report

Analysis Method: 7196A

ANALYST: Indiana

Parameter: Hexavalent Chromium

SUPERVISOR REVIEW BY: apatel

Run Number: LB99938

pH Meter ID: WC pH Meter-1

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol (ml/gm)	Final Vol (ml)	pH HN03	pH H2SO4	Absorb. at 540		Absorbance Difference	Result	Anal Date	Anal Time
								Backgrnd	Color				
1	ICV	0.5	1	100	100		2.17	0.000	0.389	0.389	0.504	12/18/2018	15:03
2	ICB		1	100	100		2.08	0.000	0.000	0.000	0.000	12/18/2018	15:04
3	CCV1	0.5	1	100	100		1.81	0.000	0.390	0.390	0.506	12/18/2018	16:00
4	CCB1		1	100	100		2.36	0.000	0.000	0.000	0.000	12/18/2018	16:00
5	LB99938BL		1	100	100		1.84	0.000	0.000	0.000	0.000	12/18/2018	16:01
6	LB99938BS	0.5	1	100	100		2.10	0.000	0.375	0.375	0.486	12/18/2018	16:01
7	J6465-10		1	100	100		2.22	0.008	0.008	0.000	0.000	12/18/2018	16:02
8	J6465-10DUP		1	100	100		2.38	0.009	0.009	0.000	0.000	12/18/2018	16:02
9	J6465-10MS	1	2	100	100		1.59	0.009	0.368	0.359	0.465	12/18/2018	16:03
10	J6465-10MSD	1	2	100	100		1.78	0.010	0.367	0.357	0.463	12/18/2018	16:03
11	CCV2	0.5	1	100	100		2.25	0.000	0.376	0.376	0.487	12/18/2018	16:04
12	CCB2		1	100	100		2.34	0.000	0.000	0.000	0.000	12/18/2018	16:04

LB99938

### WORKLIST(Hardcopy Internal Chain)

WorkList Name : Hex121718      WorkList ID : 120483      Date : 12/17/2018 11:48:46 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/18/2018	Water	J6465-10	Hexavalent Chromium	Cool 4 deg C	LANG03		SO-FB02_121718	12/17/2018	7196A

Date/Time 12/18/18  
Received by: ISB  
Relinquished by: CS

Date/Time 12/18/18 15:30  
Received by: CS  
Relinquished by: ISB