

LB99942


7471B

INSTRUMENT ID : CY3

Linear

 μ Abs. :

51392



A= 0.0000e+000

B= 1.9511e-004

C= -3.1390e-002

Rho= 0.9999984

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.002	-0.002	149	0.000	149				
0.2	0.200	0.193	-0.007	1152	0.0 %	1152				
2.5	2.500	2.513	0.013	13041	0.0 %	13041				
5.0	5.000	4.996	-0.004	25767	0.0 %	25767				
7.5	7.500	7.504	0.004	38619	0.0 %	38619				
10.0	10.000	9.996	-0.004	51392	0.0 %	51392				

LB99942 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 19 Dec 2018 17:53:44

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	50	149	-	0.0000	7471B	PPB	19 Dec 2018 17:54:45	S	Std
0.2 - 1	50.2	1152	-	0.2000	7471B	PPB	19 Dec 2018 17:57:03	S	Std
2.5 - 1	52.5	13041	-	2.5000	7471B	PPB	19 Dec 2018 17:59:21	S	Std
5.0 - 1	55	25767	-	5.0000	7471B	PPB	19 Dec 2018 18:01:39	S	Std
7.5 - 1	57.5	38619	-	7.5000	7471B	PPB	19 Dec 2018 18:03:56	S	Std
10.0 - 1	510	51392	-	10.0000	7471B	PPB	19 Dec 2018 18:06:15	S	Std
ICV34 - 1	ICV34	19974	3.8658	-	7471B	PPB	19 Dec 2018 18:09:10	U	SMPL
ICB34 - 1	ICB34	55	-0.0207	-	7471B	PPB	19 Dec 2018 18:11:25	U	SMPL
CCV57 - 1	CCV57	26413	5.1222	-	7471B	PPB	19 Dec 2018 18:13:42	U	SMPL
CCB57 - 1	CCB57	45	-0.0226	-	7471B	PPB	19 Dec 2018 18:15:58	U	SMPL
CRA - 1	CRA	1114	0.1860	-	7471B	PPB	19 Dec 2018 18:18:15	U	SMPL
HighStd - 1	HighStd	51149	9.9485	-	7471B	PPB	19 Dec 2018 18:20:31	U	SMPL
ChkStd - 1	ChkStd	35915	6.9762	-	7471B	PPB	19 Dec 2018 18:22:47	U	SMPL
PB115828BL - 1	PBS	-48	-0.0408	-	7471B	PPB	19 Dec 2018 18:25:06	U	SMPL
PB115828BS - 1	LCSS	20659	3.9995	-	7471B	PPB	19 Dec 2018 18:27:24	U	SMPL
J6446-01 - 1	P001-SS022-0612-02	4892	0.9231	-	7471B	PPB	19 Dec 2018 18:29:41	U	SMPL
J6446-02 - 1	P001-SS025-1218-02	5007	0.9455	-	7471B	PPB	19 Dec 2018 18:31:59	U	SMPL
J6446-03 - 1	P001-SS026-0006-01	4818	0.9087	-	7471B	PPB	19 Dec 2018 18:34:17	U	SMPL
J6446-04 - 1	P001-SS026-0612-01	4358	0.8189	-	7471B	PPB	19 Dec 2018 18:36:35	U	SMPL
J6446-05 - 1	P001-SS026-1218-01	1923	0.3438	-	7471B	PPB	19 Dec 2018 18:38:51	U	SMPL
CCV58 - 1	CCV58	26192	5.0791	-	7471B	PPB	19 Dec 2018 18:41:07	U	SMPL
CCB58 - 1	CCB58	71	-0.0175	-	7471B	PPB	19 Dec 2018 18:43:23	U	SMPL
J6446-06 - 1	P001-SS026-1824-01	1540	0.2691	-	7471B	PPB	19 Dec 2018 18:45:40	U	SMPL
J6446-08 - 1	P001-SS021-0006-01	7742	1.4792	-	7471B	PPB	19 Dec 2018 18:47:56	U	SMPL
J6446-09 - 1	P001-SS021-0612-01	7025	1.3393	-	7471B	PPB	19 Dec 2018 18:50:12	U	SMPL
J6446-10 - 1	P001-SS021-1218-01	6878	1.3106	-	7471B	PPB	19 Dec 2018 18:52:29	U	SMPL
J6446-11 - 1	P001-SS021-1824-01	5255	0.9939	-	7471B	PPB	19 Dec 2018 18:54:46	U	SMPL
J6446-12 - 1	P001-SS022-0006-01	3740	0.6983	-	7471B	PPB	19 Dec 2018 18:57:03	U	SMPL
J6446-13 - 1	P001-SS022-0612-01	4240	0.7959	-	7471B	PPB	19 Dec 2018 18:59:21	U	SMPL
J6446-13DUP - 1	P001-SS022-0612-01DUP	4290	0.8057	-	7471B	PPB	19 Dec 2018 19:01:39	U	SMPL
J6446-14 - 1	J6446-13MS	23023	4.4607	-	7471B	PPB	19 Dec 2018 19:03:56	U	SMPL
J6446-15 - 1	J6446-13MSD	23128	4.4812	-	7471B	PPB	19 Dec 2018 19:06:13	U	SMPL
CCV59 - 1	CCV59	26349	5.1097	-	7471B	PPB	19 Dec 2018 19:08:30	U	SMPL
CCB59 - 1	CCB59	53	-0.0210	-	7471B	PPB	19 Dec 2018 19:10:48	U	SMPL
J6446-16 - 1	P001-SS022-1218-01	4729	0.8913	-	7471B	PPB	19 Dec 2018 19:13:05	U	SMPL
J6446-17 - 1	P001-SS022-1824-01	3682	0.6870	-	7471B	PPB	19 Dec 2018 19:15:22	U	SMPL
J6446-18 - 1	P001-SS023-0006-01	2081	0.3746	-	7471B	PPB	19 Dec 2018 19:17:38	U	SMPL
J6446-19 - 1	P001-SS023-0612-01	3930	0.7354	-	7471B	PPB	19 Dec 2018 19:19:55	U	SMPL
J6446-20 - 1	P001-SS023-1218-01	5410	1.0242	-	7471B	PPB	19 Dec 2018 19:22:12	U	SMPL
PB115829BL - 1	PBS	134	-0.0052	-	7471B	PPB	19 Dec 2018 19:24:29	U	SMPL
PB115829BS - 1	LCSS	20255	3.9207	-	7471B	PPB	19 Dec 2018 19:26:46	U	SMPL
J6445-01 - 1	EB06_1-3	5019	0.9479	-	7471B	PPB	19 Dec 2018 19:29:03	U	SMPL
J6445-02 - 1	EB06_12-14	2348	0.4267	-	7471B	PPB	19 Dec 2018 19:31:20	U	SMPL
J6445-03 - 1	EB02_1.5-3.5	13086	2.5219	-	7471B	PPB	19 Dec 2018 19:33:37	U	SMPL
CCV60 - 1	CCV60	26298	5.0997	-	7471B	PPB	19 Dec 2018 19:35:54	U	SMPL
CCB60 - 1	CCB60	36	-0.0244	-	7471B	PPB	19 Dec 2018 19:38:11	U	SMPL
J6445-04 - 1	EB02_6-8	453	0.0570	-	7471B	PPB	19 Dec 2018 19:40:28	U	SMPL
J6445-05 - 1	EB02_12-14	341	0.0351	-	7471B	PPB	19 Dec 2018 19:42:45	U	SMPL
J6445-06 - 1	EB05_1.5-3.5	2085	0.3754	-	7471B	PPB	19 Dec 2018 19:44:62	U	SMPL
J6445-07 - 1	EB05_12-14	206	0.0088	-	7471B	PPB	19 Dec 2018 19:46:79	U	SMPL
J6447-01 - 1	P001-SS023-1824-01	2789	0.5128	-	7471B	PPB	19 Dec 2018 19:48:96	U	SMPL
J6447-02 - 1	P001-SS024-0006-01	4320	0.8115	-	7471B	PPB	19 Dec 2018 19:51:13	U	SMPL
J6447-03 - 1	P001-SS024-0612-01	4425	0.8320	-	7471B	PPB	19 Dec 2018 19:53:30	U	SMPL
J6447-04 - 1	P001-SS024-1218-01	4363	0.8199	-	7471B	PPB	19 Dec 2018 19:55:47	U	SMPL
J6447-05 - 1	P001-SS024-1824-01	3263	0.6053	-	7471B	PPB	19 Dec 2018 19:58:04	U	SMPL
J6447-06 - 1	P001-SS025-0006-01	5269	0.9967	-	7471B	PPB	19 Dec 2018 20:00:21	U	SMPL
CCV61 - 1	CCV61	26640	5.1665	-	7471B	PPB	19 Dec 2018 20:02:38	U	SMPL
CCB61 - 1	CCB61	45	-0.0226	-	7471B	PPB	19 Dec 2018 20:04:55	U	SMPL
J6447-07 - 1	P001-SS025-0612-01	5182	0.9797	-	7471B	PPB	19 Dec 2018 20:07:12	U	SMPL
J6447-08 - 1	P001-SS025-1218-01	4651	0.8761	-	7471B	PPB	19 Dec 2018 20:09:29	U	SMPL
J6447-08DUP - 1	P001-SS025-1218-01DUP	4811	0.9073	-	7471B	PPB	19 Dec 2018 20:11:46	U	SMPL
J6447-09 - 1	J6447-08MS	24381	4.7257	-	7471B	PPB	19 Dec 2018 20:14:03	U	SMPL
J6447-10 - 1	J6447-08MSD	24212	4.6927	-	7471B	PPB	19 Dec 2018 20:16:20	U	SMPL
J6447-11 - 1	P001-SS025-1824-01	2256	0.4088	-	7471B	PPB	19 Dec 2018 20:18:37	U	SMPL
PB115830BL - 1	PBS	66	-0.0185	-	7471B	PPB	19 Dec 2018 20:20:54	U	SMPL
PB115830BS - 1	LCSS	19634	3.7995	-	7471B	PPB	19 Dec 2018 20:23:11	U	SMPL
J6194-07 - 1	OK-01-121818-A	2034	0.3655	-	7471B	PPB	19 Dec 2018 20:25:28	U	SMPL
J6194-09 - 1	OK-01-121818-B	2791	0.5132	-	7471B	PPB	19 Dec 2018 20:27:45	U	SMPL
CCV62 - 1	CCV62	26356	5.1111	-	7471B	PPB	19 Dec 2018 20:29:62	U	SMPL
CCB62 - 1	CCB62	38	-0.0240	-	7471B	PPB	19 Dec 2018 20:31:79	U	SMPL
J6196-07 - 1	EO-01-121818-A	1313	0.2248	-	7471B	PPB	19 Dec 2018 20:33:96	U	SMPL
J6196-09 - 1	EO-01-121818-B	6450	1.2271	-	7471B	PPB	19 Dec 2018 20:36:13	U	SMPL
J6196-11 - 1	EO-01-121818-C	8605	1.6476	-	7471B	PPB	19 Dec 2018 20:38:30	U	SMPL
J6199-09 - 1	SU-02-121818-A	2323	0.4219	-	7471B	PPB	19 Dec 2018 20:40:47	U	SMPL
J6199-11 - 1	SU-02-121818-B	4450	0.8369	-	7471B	PPB	19 Dec 2018 20:42:64	U	SMPL
J6465-01 - 1	EB01_12-14	252	0.0178	-	7471B	PPB	19 Dec 2018 20:44:81	U	SMPL
J6465-01DUP - 1	EB01_12-14DUP	277	0.0227	-	7471B	PPB	19 Dec 2018 20:46:98	U	SMPL
J6465-01MS - 1	EB01_12-14MS	17443	3.3720	-	7471B	PPB	19 Dec 2018 20:49:15	U	SMPL
J6465-01MSD - 1	EB01_12-14MSD	17633	3.4091	-	7471B	PPB	19 Dec 2018 20:51:32	U	SMPL
J6465-02 - 1	EB01_0-2	2057	0.3700	-	7471B	PPB	19 Dec 2018 20:53:49	U	SMPL
CCV63 - 1	CCV63	26475	5.1343	-	7471B	PPB	19 Dec 2018 20:56:06	U	SMPL

LB99942 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 19 Dec 2018 17:53:44

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB63 - 1	CCB63	71	-0.0175		-7471B	PPB	19 Dec 2018 21:06:31	U	SMPL
J6465-03 - 1	EB03_12-14	158	-0.0006		-7471B	PPB	19 Dec 2018 21:08:50	U	SMPL
J6465-04 - 1	EB03_1.5-3.5	11050	2.1246		-7471B	PPB	19 Dec 2018 21:11:07	U	SMPL
J6465-05 - 1	EB08_12-14	98	-0.0123		-7471B	PPB	19 Dec 2018 21:13:24	U	SMPL
J6465-06 - 1	EB08_1.5-3.5	4571	0.8605		-7471B	PPB	19 Dec 2018 21:15:42	U	SMPL
J6469-01 - 1	DRUM-1	509	0.0679		-7471B	PPB	19 Dec 2018 21:17:59	U	SMPL
J6475-01 - 1	SLUDGE	18511	3.5804		-7471B	PPB	19 Dec 2018 21:20:16	U	SMPL
J6477-01 - 1	SB-A1	1874	0.3343		-7471B	PPB	19 Dec 2018 21:22:34	U	SMPL
J6477-02 - 1	SB-A2	2102	0.3787		-7471B	PPB	19 Dec 2018 21:24:52	U	SMPL
CCV64 - 1	CCV64	26854	5.2082		-7471B	PPB	19 Dec 2018 21:27:10	U	SMPL
CCB64 - 1	CCB64	50	-0.0216		-7471B	PPB	19 Dec 2018 21:29:27	U	SMPL
J6446-13LX5 - 1		985	0.1608		-7471B	PPB	19 Dec 2018 21:31:45	U	SMPL
J6446-13A - 1		22444	4.3478		-7471B	PPB	19 Dec 2018 21:34:02	U	SMPL
J6447-08LX5 - 1		960	0.1559		-7471B	PPB	19 Dec 2018 21:36:19	U	SMPL
J6447-08A - 1		22678	4.3934		-7471B	PPB	19 Dec 2018 21:38:38	U	SMPL
J6465-01LX5 - 1		68	-0.0181		-7471B	PPB	19 Dec 2018 21:40:55	U	SMPL
J6465-01A - 1		18056	3.4916		-7471B	PPB	19 Dec 2018 21:43:13	U	SMPL
CCV65 - 1	CCV65	26899	5.2170		-7471B	PPB	19 Dec 2018 21:45:30	U	SMPL
CCB65 - 1	CCB65	47	-0.0222		-7471B	PPB	19 Dec 2018 21:47:48	U	SMPL