

LB 99931


ISM02.4

INSTRUMENT ID: C41

Linear

 μ Abs.:

47645



A= 0.0000e+000

B= 2.1127e-004

C= -1.0390e-001

Rho= 0.9999780

Accept=Accepted

slope
 $y = 2.1127e-004 x - 1.0390e-001$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	-0.010	-0.010	446	0.000	446				
0.05	0.050	0.048	-0.002	718	0.0 %	718				
0.2	0.200	0.198	-0.002	1431	0.0 %	1431				
2.5	2.500	2.483	-0.017	12243	0.0 %	12243				
5.0	5.000	5.039	0.039	24341	0.0 %	24341				
7.5	7.500	7.530	0.030	36134	0.0 %	36134				
10.0	10.000	9.962	-0.038	47645	0.0 %	47645				

47645
 14
 11
 10
 103
 12/19/18

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.05	446	-		0 ISM02.4	PPB	12/19/2018 15:13	Std
	0.05 S0.05	718	-		0.05 ISM02.4	PPB	12/19/2018 15:16	Std
	0.2 S0.2	1431	-		0.2 ISM02.4	PPB	12/19/2018 15:18	Std
	2.5 S2.5	12243	-		2.5 ISM02.4	PPB	12/19/2018 15:20	Std
	5 S5	24341	-		5 ISM02.4	PPB	12/19/2018 15:22	Std
	7.5 S7.5	36134	-		7.5 ISM02.4	PPB	12/19/2018 15:25	Std
	10 S10.0	47645	-		10 ISM02.4	PPB	12/19/2018 15:27	Std
ICV	ICV	19142	3.9402	-	ISM02.4	PPB	12/19/2018 15:30	SMPL
ICB	ICB	395	-0.0205	-	ISM02.4	PPB	12/19/2018 15:32	SMPL
CCV053	CCV053	24143	4.9968	-	ISM02.4	PPB	12/19/2018 15:34	SMPL
CCB053	CCB053	386	-0.0224	-	ISM02.4	PPB	12/19/2018 15:37	SMPL
PB115824BL	PBS824	462	-0.0063	-	ISM02.4	PPB	12/19/2018 15:41	SMPL
J6266-01	MCOAQ0	1017	0.111	-	ISM02.4	PPB	12/19/2018 15:44	SMPL
J6266-02	MCOAQ2	748	0.0541	-	ISM02.4	PPB	12/19/2018 15:46	SMPL
J6266-03	MCOAQ4	953	0.0974	-	ISM02.4	PPB	12/19/2018 15:48	SMPL
J6266-04	MCOAQ4D	970	0.101	-	ISM02.4	PPB	12/19/2018 15:51	SMPL
J6266-05	MCOAQ4S	12788	2.5978	-	ISM02.4	PPB	12/19/2018 15:53	SMPL
J6266-06	MCOAQ6	1139	0.1367	-	ISM02.4	PPB	12/19/2018 15:55	SMPL
J6266-07	MCOAQ8	1555	0.2246	-	ISM02.4	PPB	12/19/2018 15:57	SMPL
J6266-08	MCOAR0	1577	0.2293	-	ISM02.4	PPB	12/19/2018 16:00	SMPL
J6266-09	MCOAR4	1149	0.1388	-	ISM02.4	PPB	12/19/2018 16:02	SMPL
J6266-10	MCOAS2	1713	0.258	-	ISM02.4	PPB	12/19/2018 16:04	SMPL
J6266-11	MCOAS4	522	0.0064	-	ISM02.4	PPB	12/19/2018 16:06	SMPL
J6266-12	MCOAS7	500	0.0017	-	ISM02.4	PPB	12/19/2018 16:09	SMPL
PB115825BL	PBS825	441	-0.0107	-	ISM02.4	PPB	12/19/2018 16:14	SMPL
J6267-01	MCOAS9	696	0.0431	-	ISM02.4	PPB	12/19/2018 16:16	SMPL
J6267-02	MCOAS9D	700	0.044	-	ISM02.4	PPB	12/19/2018 16:18	SMPL
J6267-03	MCOAS9S	12719	2.5832	-	ISM02.4	PPB	12/19/2018 16:20	SMPL
J6267-04	MCOAT6	1199	0.1494	-	ISM02.4	PPB	12/19/2018 16:23	SMPL
J6267-05	MCOAT7	976	0.1023	-	ISM02.4	PPB	12/19/2018 16:25	SMPL
PB115826BL	PBS826	441	-0.0107	-	ISM02.4	PPB	12/19/2018 16:30	SMPL
CCV054	CCV054	24180	5.0046	-	ISM02.4	PPB	12/19/2018 16:33	SMPL
CCB054	CCB054	387	-0.0221	-	ISM02.4	PPB	12/19/2018 16:35	SMPL
J6289-01	MCOBT4	2346	0.3917	-	ISM02.4	PPB	12/19/2018 16:37	SMPL
J6289-02	MCOBT4D	2451	0.4139	-	ISM02.4	PPB	12/19/2018 16:40	SMPL
J6289-03	MCOBT4S	14467	2.9525	-	ISM02.4	PPB	12/19/2018 16:42	SMPL
PB115827BL	PBS827	434	-0.0122	-	ISM02.4	PPB	12/19/2018 16:47	SMPL
J6315-01	MBE7Z4	132882	27.9701	-	ISM02.4	PPB	12/19/2018 16:49	SMPL
J6315-02	MBE7Z4D	128961	27.1417	-	ISM02.4	PPB	12/19/2018 16:51	SMPL
J6315-03	MBE7Z4S	135864	28.6001	-	ISM02.4	PPB	12/19/2018 16:54	SMPL
J6315-04	MBE8B5	957979	HIGH	-	ISM02.4	PPB	12/19/2018 16:56	
J6315-05	MBE8B6	862411	HIGH	-	ISM02.4	PPB	12/19/2018 16:58	
J6315-01DLX5	MBE7Z4	25752	5.3367	-	ISM02.4	PPB	12/19/2018 17:00	SMPL
J6315-02DLX5	MBE7Z4D	25456	5.2742	-	ISM02.4	PPB	12/19/2018 17:03	SMPL
J6315-03DLX5	MBE7Z4S	27557	5.7181	-	ISM02.4	PPB	12/19/2018 17:05	SMPL
J6315-04DLX10	MBE8B5	284179	59.9346	-	ISM02.4	PPB	12/19/2018 17:07	SMPL

INSTRUMENT ID : CV1

J6315-05DLX10	MBE8B6	210001	44.263 -	ISM02.4	PPB	12/19/2018 17:10	SMPL
CCV055	CCV055	23948	4.9556 -	ISM02.4	PPB	12/19/2018 17:12	SMPL
CCB055	CCB055	298	-0.0409 -	ISM02.4	PPB	12/19/2018 17:14	SMPL
J6315-04DL2X100	MBE8B5	31383	6.5264 -	ISM02.4	PPB	12/19/2018 17:17	SMPL
J6315-05DL2X100	MBE8B6	22641	4.6795 -	ISM02.4	PPB	12/19/2018 17:19	SMPL
CCV056	CCV056	24320	5.0342 -	ISM02.4	PPB	12/19/2018 17:21	SMPL
CCB056	CCB056	382	-0.0232 -	ISM02.4	PPB	12/19/2018 17:24	SMPL