


7470A

LB 99948
INSTRUMENT ID: CV1

Linear

 μ Abs.:

39768



A= 0.0000e+000

B= 2.5259e-004

C= -1.2277e-002

Rho= 0.9999023

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.005	0.005	68	0.000	68				
0.2	0.200	0.205	0.005	859	0.0 %	859				
2.5	2.500	2.466	-0.034	9812	0.0 %	9812				
5.0	5.000	5.078	0.078	20154	0.0 %	20154				
7.5	7.500	7.413	-0.087	29397	0.0 %	29397				
10.0	10.000	10.033	0.033	39768	0.0 %	39768				

LB99948 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 Dec 2018 10:35:26

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	68	-	0.0000	7470A	PPB	20 Dec 2018 10:44:26	S	Std
0.2 - 1	50.2	859	-	0.2000	7470A	PPB	20 Dec 2018 10:46:35	S	Std
2.5 - 1	52.5	9812	-	2.5000	7470A	PPB	20 Dec 2018 10:48:44	S	Std
5.0 - 1	55	20154	-	5.0000	7470A	PPB	20 Dec 2018 10:50:52	S	Std
7.5 - 1	57.5	29397	-	7.5000	7470A	PPB	20 Dec 2018 10:53:00	S	Std
10.0 - 1	510	39768	-	10.0000	7470A	PPB	20 Dec 2018 10:55:08	S	Std
ICV35 - 1	ICV35	14805	3.7273	-	7470A	PPB	20 Dec 2018 10:57:39	U	SMPL
ICB35 - 1	ICB35	-15	-0.0161	-	7470A	PPB	20 Dec 2018 10:59:45	U	SMPL
CCV66 - 1	CCV66	18959	4.7766	-	7470A	PPB	20 Dec 2018 11:01:51	U	SMPL
CCB66 - 1	CCB66	-1	-0.0125	-	7470A	PPB	20 Dec 2018 11:03:58	U	SMPL
CRA - 1	CRA	819	0.1946	-	7470A	PPB	20 Dec 2018 11:06:05	U	SMPL
HighStd - 1	HighStd	38172	9.6296	-	7470A	PPB	20 Dec 2018 11:08:12	U	SMPL
ChkStd - 1	ChkStd	27085	6.8291	-	7470A	PPB	20 Dec 2018 11:10:19	U	SMPL
PB115865BL - 1	PBW	38	-0.0027	-	7470A	PPB	20 Dec 2018 11:12:28	U	SMPL
PB115865BS - 1	LCSW	14804	3.7271	-	7470A	PPB	20 Dec 2018 11:18:38	U	SMPL
J6194-08 - 1	OK-01-121818-A	16	-0.0082	-	7470A	PPB	20 Dec 2018 11:20:46	U	SMPL
J6194-08DUP - 1	OK-01-121818-ADUP	70	0.0054	-	7470A	PPB	20 Dec 2018 11:22:54	U	SMPL
J6194-08MS - 1	OK-01-121818-AMS	14671	3.6935	-	7470A	PPB	20 Dec 2018 11:27:47	U	SMPL
J6194-08MSD - 1	OK-01-121818-AMSD	14782	3.7215	-	7470A	PPB	20 Dec 2018 11:29:56	U	SMPL
J6194-10 - 1	OK-01-121818-B	-30	-0.0199	-	7470A	PPB	20 Dec 2018 11:32:03	U	SMPL
CCV67 - 1	CCV67	19385	4.8842	-	7470A	PPB	20 Dec 2018 11:34:09	U	SMPL
CCB67 - 1	CCB67	11	-0.0095	-	7470A	PPB	20 Dec 2018 11:36:16	U	SMPL
J6196-08 - 1	EO-01-121818-A	41	-0.0019	-	7470A	PPB	20 Dec 2018 11:38:23	U	SMPL
J6196-10 - 1	EO-01-121818-B	60	0.0029	-	7470A	PPB	20 Dec 2018 11:40:29	U	SMPL
J6196-12 - 1	EO-01-121818-C	102	0.0135	-	7470A	PPB	20 Dec 2018 11:42:36	U	SMPL
J6199-10 - 1	SU-02-121818-A	138	0.0226	-	7470A	PPB	20 Dec 2018 11:44:43	U	SMPL
J6199-12 - 1	SU-02-121818-B	37	-0.0029	-	7470A	PPB	20 Dec 2018 11:46:51	U	SMPL
J6466-01 - 1	OILY-DEBRIS	99	0.0127	-	7470A	PPB	20 Dec 2018 11:48:59	U	SMPL
J6469-02 - 1	DRUM-1	64	0.0039	-	7470A	PPB	20 Dec 2018 11:51:07	U	SMPL
J6470-02 - 1	OD-3	79	0.0077	-	7470A	PPB	20 Dec 2018 11:53:15	U	SMPL
J6472-01 - 1	OILY-DEBRIS	93	0.0112	-	7470A	PPB	20 Dec 2018 11:55:24	U	SMPL
J6475-02 - 1	SLUDGE	137	0.0223	-	7470A	PPB	20 Dec 2018 11:57:31	U	SMPL
CCV68 - 1	CCV68	19042	4.7975	-	7470A	PPB	20 Dec 2018 11:59:38	U	SMPL
CCB68 - 1	CCB68	25	-0.0060	-	7470A	PPB	20 Dec 2018 12:01:45	U	SMPL
PB115866BL - 1	PBW	54	0.0014	-	7470A	PPB	20 Dec 2018 12:03:52	U	SMPL
PB115866BS - 1	LCSW	14784	3.7220	-	7470A	PPB	20 Dec 2018 12:06:00	U	SMPL
J6462-07 - 1	OILY-WATER	31	-0.0044	-	7470A	PPB	20 Dec 2018 12:08:07	U	SMPL
J6462-07DUP - 1	OILY-WATERDUP	86	0.0094	-	7470A	PPB	20 Dec 2018 12:10:15	U	SMPL
J6462-07MS - 1	OILY-WATERMS	14710	3.7033	-	7470A	PPB	20 Dec 2018 12:14:53	U	SMPL
J6462-07MSD - 1	OILY-WATERMSD	14706	3.7023	-	7470A	PPB	20 Dec 2018 12:17:00	U	SMPL
J6472-04 - 1	OILY-WATER	182	0.0337	-	7470A	PPB	20 Dec 2018 12:19:09	U	SMPL
PB115865TB - 1	PB115820TB	127	0.0198	-	7470A	PPB	20 Dec 2018 12:21:17	U	SMPL
PB115866TB - 1	PB115822TB	94	0.0115	-	7470A	PPB	20 Dec 2018 12:23:26	U	SMPL
CCV69 - 1	CCV69	19058	4.8016	-	7470A	PPB	20 Dec 2018 12:25:33	U	SMPL
CCB69 - 1	CCB69	13	-0.0090	-	7470A	PPB	20 Dec 2018 12:27:41	U	SMPL
J6194-08LX5 - 1		69	0.0052	-	7470A	PPB	20 Dec 2018 12:29:48	U	SMPL
J6194-08A - 1		14470	3.6427	-	7470A	PPB	20 Dec 2018 12:31:56	U	SMPL
J6462-07LX5 - 1		36	-0.0032	-	7470A	PPB	20 Dec 2018 12:34:03	U	SMPL
J6462-07A - 1		14109	3.5515	-	7470A	PPB	20 Dec 2018 12:36:11	U	SMPL
CCV70 - 1	CCV70	19246	4.8491	-	7470A	PPB	20 Dec 2018 12:38:18	U	SMPL
CCB70 - 1	CCB70	10	-0.0098	-	7470A	PPB	20 Dec 2018 12:40:26	U	SMPL