



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : jignesh

Run Number: LB99957

Slope : 99.0

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.1	4.01	12/19/2018	18:00
2	CAL2	1	Water			20.2	7.00	12/19/2018	18:02
3	CAL3	1	Water			20.1	10.03	12/19/2018	18:04
4	ICV	1	Water			20.1	7.01	12/19/2018	18:06
5	CCV1	1	Water			20.3	2.00	12/19/2018	18:08
6	J6482-04	1	Solid	20.02	20	19.2	5.33	12/19/2018	18:10
7	J6482-04DUP	1	Solid	20.03	20	19.3	5.34	12/19/2018	18:12
8	J6482-12	1	Solid	20.02	20	20.1	5.18	12/19/2018	18:14
9	J6493-02	1	Solid	20.04	20	21.4	7.49	12/19/2018	18:16
10	J6494-01	1	Solid	20.04	20	20.3	9.08	12/19/2018	18:18
11	CCV2	1	Water			20.1	12.00	12/19/2018	18:20

WORKLIST(Hardcopy Internal Chain)

UB 99957

WorkList Name : corrosivity121918

WorkList ID : 120601

Date : 12/19/2018 4:16:45 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/19/2018	Solid	J6482-04	Corrosivity	Cool 4 deg C	PSEG01	E41	R20-OS	12/19/2018	9045D
12/19/2018	Solid	J6482-12	Corrosivity	Cool 4 deg C	PSEG01	E41	DPC-91468-OS-01	12/19/2018	9045D
12/19/2018	Solid	J6493-02	Corrosivity	Cool 4 deg C	PSEG01	E41	SOIL-DRUM	12/19/2018	9045D
12/14/2018	Solid	J6494-01	Corrosivity	Cool 4 deg C	CLEA11	D22	OWS	12/14/2018	9045D

Date/Time 12/19/18 5:55
 Received by: SP
 Relinquished by: CP

Date/Time 12/19/18 6:30
 Received by: CP
 Relinquished by: SP