

2B99959

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : As

12/20/2018 13:57

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	98.339	0.0	0.080	
ICB1	0.219	0.0	0.004	
CCV1	244.155	0.0	0.193	
CCB1	0.081	0.0	0.004	
PB115886BL	0.176	0.0	0.004	
PB115886BS	99.445	0.0	0.081	
LOWPB115886	10.082	0.0	0.012	
HIGHPB115886	498.352	0.0	0.390	
J6482-01	0.940	0.0	0.005	
J6482-01DUP	-0.296	0.0	0.004	
J6482-01MS	41.294	0.0	0.036	
J6482-01MSD	41.218	0.0	0.036	
J6482-09	19.764	0.0	0.019	
J6493-01	6.570	0.0	0.009	
CCV2	236.338	0.0	0.187	
CCB2	0.328	0.0	0.004	

N 16
Mean 81.063
SD 137.1578
CV% 169.20

Aquakem v. 7.2AQ1

Results from time period:

Thu Dec 20 13:32:33 2018

Thu Dec 20 13:57:00 2018

✓

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-0.239	µg/l	12/20/2018 12:33:31	
5.0PPBCN	A	Total CN	P	5.5361	µg/l	12/20/2018 12:33:32	
10PPBCN	A	Total CN	P	10.0636	µg/l	12/20/2018 12:33:33	
50PPBCN	A	Total CN	P	50.2902	µg/l	12/20/2018 12:33:34	
100PPBCN	A	Total CN	P	101.1613	µg/l	12/20/2018 12:33:35	
250PPBCN	A	Total CN	P	246.9114	µg/l	12/20/2018 12:33:36	
500PPBCN	A	Total CN	P	501.2764	µg/l	12/20/2018 12:33:37	
ICV1	S	Total CN	P	98.3388	µg/l	12/20/2018 13:32:33	
ICB1	S	Total CN	P	0.2187	µg/l	12/20/2018 13:32:34	
CCV1	S	Total CN	P	244.1552	µg/l	12/20/2018 13:32:35	
CCB1	S	Total CN	P	0.0811	µg/l	12/20/2018 13:32:36	
PB115886BL	S	Total CN	P	0.1762	µg/l	12/20/2018 13:32:37	
PB115886BS	S	Total CN	P	99.4453	µg/l	12/20/2018 13:32:38	
LOWPB115886	S	Total CN	P	10.0821	µg/l	12/20/2018 13:32:39	
HIGHPB115886	S	Total CN	P	498.3518	µg/l	12/20/2018 13:32:40	
J6482-01	S	Total CN	P	0.9398	µg/l	12/20/2018 13:32:41	
J6482-01DUP	S	Total CN	P	-0.2961	µg/l	12/20/2018 13:32:42	
J6482-01MS	S	Total CN	P	41.2936	µg/l	12/20/2018 13:32:43	
J6482-01MSD	S	Total CN	P	41.2176	µg/l	12/20/2018 13:39:12	
J6482-09	S	Total CN	P	19.764	µg/l	12/20/2018 13:39:13	
J6493-01	S	Total CN	P	6.5697	µg/l	12/20/2018 13:39:14	
CCV2	S	Total CN	P	236.3382	µg/l	12/20/2018 13:39:17	
CCB2	S	Total CN	P	0.3283	µg/l	12/20/2018 13:39:18	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/20/2018 12:35


Test Total CN

Accepted 12/20/2018 12:34

Factor 1291
Bias 0.004

Coeff. of det. 0.999937

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.2390	0.0000	
2	5.0PPBCN	0.008	5.5361	5.0000	
3	10PPBCN	0.012	10.0636	10.0000	
4	50PPBCN	0.043	50.2902	50.0000	
5	100PPBCN	0.082	101.1613	100.0000	
6	250PPBCN	0.195	246.9114	250.0000	
7	500PPBCN	0.392	501.2764	500.0000	