



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : sweta

Parameter: pH

Supervisor Review By : jignesh

Run Number: LB99960

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.1	4.01	12/20/2018	13:50
2	CAL2	1	Water			22.0	7.00	12/20/2018	13:52
3	CAL3	1	Water			22.1	10.04	12/20/2018	13:54
4	ICV	1	Water			22.2	7.00	12/20/2018	13:56
5	CCV1	1	Water			22.1	2.00	12/20/2018	13:58
6	J6497-01	1	Water			16.9	5.75	12/20/2018	14:00
7	J6497-01DUP	1	Water			17.0	5.76	12/20/2018	14:02
8	J6497-02	1	Water			13.0	6.47	12/20/2018	14:04
9	J6497-03	1	Water			12.0	6.40	12/20/2018	14:06
10	J6497-04	1	Water			12.6	6.39	12/20/2018	14:08
11	J6508-01	1	Water			19.5	5.44	12/20/2018	14:10
12	J6508-02	1	Water			19.5	5.25	12/20/2018	14:12
13	J6508-03	1	Water			19.8	5.35	12/20/2018	14:14
14	CCV2	1	Water			22.0	12.00	12/20/2018	14:16

WORKLIST(Hardcopy Internal Chain)

LB 99960

WorkList Name : pH122018

WorkList ID : 120648

Date : 12/20/2018 12:18:53 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/19/2018	Water	J6497-01	pH	Cool 4 deg C	LOCK01	E41	MW-1	12/19/2018	SM 4500-H B
12/19/2018	Water	J6497-02	pH	Cool 4 deg C	LOCK01	E41	MW-2	12/19/2018	SM 4500-H B
12/19/2018	Water	J6497-03	pH	Cool 4 deg C	LOCK01	E41	MW-3	12/19/2018	SM 4500-H B
12/19/2018	Water	J6497-04	pH	Cool 4 deg C	LOCK01	E41	MW-4	12/19/2018	SM 4500-H B
12/19/2018	Water	J6508-01	pH	Cool 4 deg C	STER07	D62	CELL-A	12/19/2018	SM 4500-H B
12/19/2018	Water	J6508-02	pH	Cool 4 deg C	STER07	D62	CELL-B-C	12/19/2018	SM 4500-H B
12/19/2018	Water	J6508-03	pH	Cool 4 deg C	STER07	D62	CELL-D	12/19/2018	SM 4500-H B

Date/Time 12/20/18 1:40
 Received by: SP
 Relinquished by: CP

Date/Time 12/20/18 2:20
 Received by: CP
 Relinquished by: SP