

LB99967


ISM02.4

INSTRUMENT ID: CVL

Linear

 μ Abs.:

43321



A= 0.0000e+000

B= 2.3486e-004

C= -1.1990e-001

Rho= 0.9999316

Accept=Accepted

slope
y = 1.0125 x

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.025	0.025	618	0.000	618					1
0.2	0.200	0.215	0.015	1426	0.0 %	1426					7
2.5	2.500	2.511	0.011	11200	0.0 %	11200					0
5.0	5.000	4.918	-0.082	21452	0.0 %	21452					-2
7.5	7.500	7.476	-0.024	32344	0.0 %	32344					0
10.0	10.000	10.054	0.054	43321	0.0 %	43321					1

AB
12/20/18

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	618	-		0 ISM02.4	PPB	12/20/2018 14:49	Std
	0.2 S0.2	1426	-		0.2 ISM02.4	PPB	12/20/2018 14:51	Std
	2.5 S2.5	11200	-		2.5 ISM02.4	PPB	12/20/2018 14:53	Std
	5 S5	21452	-		5 ISM02.4	PPB	12/20/2018 14:58	Std
	7.5 S7.5	32344	-		7.5 ISM02.4	PPB	12/20/2018 15:03	Std
	10 S10.0	43321	-		10 ISM02.4	PPB	12/20/2018 15:08	Std
ICV	ICV	17706	4.0385 -		ISM02.4	PPB	12/20/2018 15:13	SMPL
ICB	ICB	549	0.009 -		ISM02.4	PPB	12/20/2018 15:15	SMPL
CCV071	CCV071	21495	4.9284 -		ISM02.4	PPB	12/20/2018 15:18	SMPL
CCB071	CCB071	539	0.0067 -		ISM02.4	PPB	12/20/2018 15:20	SMPL
J6366-01	METC10	19	-0.1154 -		ISM02.4	PPB	12/20/2018 15:24	SMPL
J6366-02	METC11	16	-0.1161 -		ISM02.4	PPB	12/20/2018 15:27	SMPL
J6366-03	METC11D	18	-0.1157 -		ISM02.4	PPB	12/20/2018 15:29	SMPL
J6366-04	METC11S	4674	0.9778 -		ISM02.4	PPB	12/20/2018 15:36	SMPL
PB115880BL	PBW880	597	0.0203 -		ISM02.4	PPB	12/20/2018 15:39	SMPL
CCV072	CCV072	21501	4.9298 -		ISM02.4	PPB	12/20/2018 15:41	SMPL
CCB072	CCB072	546	0.0083 -		ISM02.4	PPB	12/20/2018 15:43	SMPL