


7471B

LB99972
INSTRUMENT ID: CV1

Linear

 μ Abs.:

56723



A= 0.0000e+000

B= 1.7605e-004

C= 3.7591e-003

Rho= 0.9999714

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.008	0.008	26	0.000	26				
0.2	0.200	0.200	0.000	1115	0.0 %	1115				
2.5	2.500	2.508	0.008	14226	0.0 %	14226				
5.0	5.000	4.950	-0.050	28096	0.0 %	28096				
7.5	7.500	7.544	0.044	42830	0.0 %	42830				
10.0	10.000	9.990	-0.010	56723	0.0 %	56723				

LB99972 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 20 Dec 2018 17:21:10

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	26	-	0.0000	7471B	PPB	20 Dec 2018 17:22:42	S	Std
0.2 - 1	50.2	1115	-	0.2000	7471B	PPB	20 Dec 2018 17:25:00	S	Std
2.5 - 1	52.5	14226	-	2.5000	7471B	PPB	20 Dec 2018 17:27:18	S	Std
5.0 - 1	55	28096	-	5.0000	7471B	PPB	20 Dec 2018 17:29:35	S	Std
7.5 - 1	57.5	42830	-	7.5000	7471B	PPB	20 Dec 2018 17:31:53	S	Std
10.0 - 1	510	56723	-	10.0000	7471B	PPB	20 Dec 2018 17:34:10	S	Std
ICV37 - 1	ICV37	22099	3.8942	-	7471B	PPB	20 Dec 2018 17:36:56	U	SMPL
ICB37 - 1	ICB37	-85	-0.0112	-	7471B	PPB	20 Dec 2018 17:39:11	U	SMPL
CCV73 - 1	CCV73	27583	4.8596	-	7471B	PPB	20 Dec 2018 17:43:56	U	SMPL
CCB73 - 1	CCB73	-91	-0.0123	-	7471B	PPB	20 Dec 2018 17:46:12	U	SMPL
CRA - 1	CRA	1080	0.1939	-	7471B	PPB	20 Dec 2018 17:48:28	U	SMPL
HighStd - 1	HighStd	55274	9.7345	-	7471B	PPB	20 Dec 2018 17:50:44	U	SMPL
ChkStd - 1	ChkStd	39201	6.9050	-	7471B	PPB	20 Dec 2018 17:53:01	U	SMPL
PB115878BL - 1	PBS	-164	-0.0251	-	7471B	PPB	20 Dec 2018 17:55:19	U	SMPL
J6482-01 - 1	R20-OS	888	0.1601	-	7471B	PPB	20 Dec 2018 17:59:55	U	SMPL
J6482-09 - 1	DPC-91468-OS-01	591	0.1078	-	7471B	PPB	20 Dec 2018 18:02:12	U	SMPL
J6493-01 - 1	SOIL-DRUM	1781	0.3173	-	7471B	PPB	20 Dec 2018 18:04:29	U	SMPL
J6493-01DUP - 1	SOIL-DRUMDUP	1664	0.2967	-	7471B	PPB	20 Dec 2018 18:06:48	U	SMPL
J6493-01MS - 1	SOIL-DRUMMS	23061	4.0636	-	7471B	PPB	20 Dec 2018 18:09:04	U	SMPL
CCV74 - 1	CCV74	27652	4.8718	-	7471B	PPB	20 Dec 2018 18:11:20	U	SMPL
CCB74 - 1	CCB74	-124	-0.0181	-	7471B	PPB	20 Dec 2018 18:13:36	U	SMPL
J6493-01MSD - 1	SOIL-DRUMMSD	22475	3.9604	-	7471B	PPB	20 Dec 2018 18:15:53	U	SMPL
J6493-01LX5 - 1		215	0.0416	-	7471B	PPB	20 Dec 2018 18:18:09	U	SMPL
J6493-01A - 1		22952	4.0444	-	7471B	PPB	20 Dec 2018 18:20:26	U	SMPL
PB115878BS - 1	LCSS	21989	3.8748	-	7471B	PPB	20 Dec 2018 18:22:43	U	SMPL
CCV75 - 1	CCV75	27775	4.8934	-	7471B	PPB	20 Dec 2018 18:25:00	U	SMPL
CCB75 - 1	CCB75	-96	-0.0131	-	7471B	PPB	20 Dec 2018 18:27:18	U	SMPL