



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : jignesh

Run Number: LB99983

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.1	4.01	12/20/2018	16:30
2	CAL2	1	Water			22.3	7.00	12/20/2018	16:32
3	CAL3	1	Water			22.1	10.04	12/20/2018	16:34
4	ICV	1	Water			22.2	7.01	12/20/2018	16:36
5	CCV1	1	Water			22.1	2.00	12/20/2018	17:26
6	J6188-04	1	Solid	20.02	20	21.6	7.44	12/20/2018	17:28
7	J6188-04DUP	1	Solid	20.03	20	21.7	7.45	12/20/2018	17:30
8	J6191-02	1	Solid	20.02	20	21.3	8.27	12/20/2018	17:32
9	J6191-04	1	Solid	20.02	20	21.1	8.52	12/20/2018	17:34
10	J6192-04	1	Solid	20.04	20	20.8	8.34	12/20/2018	17:36
11	J6200-02	1	Solid	20.02	20	21.6	8.85	12/20/2018	17:38
12	J6499-02	1	Solid	20.03	20	22.2	9.45	12/20/2018	17:40
13	J6500-02	1	Solid	20.03	20	19.8	8.53	12/20/2018	17:42
14	J6501-02	1	Solid	20.04	20	22.3	8.48	12/20/2018	17:44
15	J6518-03	1	Solid	20.02	20	21.7	8.70	12/20/2018	17:46
16	CCV2	1	Water			22.1	12.00	12/20/2018	17:48
17	J6520-04	1	Solid	20.02	20	21.2	7.55	12/20/2018	17:50
18	CCV3	1	Water			22.1	2.00	12/20/2018	17:52

WORKLIST(Hardcopy Internal Chain)

B99983

WorkList Name : corrosivity122018

WorkList ID : 120662

Date : 12/20/2018 4:20:47 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/20/2018	Solid	J6188-04	Corrosivity	Cool 4 deg C	PSEG05	E51	BU-01-122018	12/20/2018	9045D
12/20/2018	Solid	J6191-02	Corrosivity	Cool 4 deg C	PSEG05	E51	NB-07-122018-A	12/20/2018	9045D
12/20/2018	Solid	J6191-04	Corrosivity	Cool 4 deg C	PSEG05	E51	NB-07-122018-B	12/20/2018	9045D
12/20/2018	Solid	J6192-04	Corrosivity	Cool 4 deg C	PSEG05	E51	HD-01-122018	12/20/2018	9045D
12/20/2018	Solid	J6200-02	Corrosivity	Cool 4 deg C	PSEG05	E51	TR-05-122018	12/20/2018	9045D
12/20/2018	Solid	J6499-02	Corrosivity	Cool 4 deg C	PSEG01	E42	72-11989	12/20/2018	9045D
12/19/2018	Solid	J6500-02	Corrosivity	Cool 4 deg C	SEAC01	E21	SOIL-CUTTINGS	12/19/2018	9045D
12/20/2018	Solid	J6501-02	Corrosivity	Cool 4 deg C	PSEG01	E41	VNJ-221	12/20/2018	9045D
12/20/2018	Solid	J6518-03	Corrosivity	Cool 4 deg C	PSEG01	E41	337-SOIL	12/20/2018	9045D
12/20/2018	Solid	J6520-04	Corrosivity	Cool 4 deg C	PSEG01	E51	ROLLOFF-165	12/20/2018	9045D

Date/Time 12/20/18 5:20
 Received by: SP
 Relinquished by: CP

Date/Time 12/20/18 6:00
 Received by: CP
 Relinquished by: SP