

LB99985


7470A

INSTRUMENT ID: CV2

Linear

μ Abs.:

38870



A= 0.0000e+000

B= 2.5963e-004

C= 8.4709e-003

Rho= 0.9997872

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.015	0.015	26	0.000	26				
0.2	0.200	0.208	0.008	768	0.0 %	768				
2.5	2.500	2.514	0.014	9650	0.0 %	9650				
5.0	5.000	5.018	0.018	19294	0.0 %	19294				
7.5	7.500	7.345	-0.155	28259	0.0 %	28259				
10.0	10.000	10.100	0.100	38870	0.0 %	38870				

LB99985 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Dec 2018 10:15:04

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	26	-	0.0000	7470A	PPB	21 Dec 2018 10:21:07	S	Std
0.2 - 1	50.2	768	-	0.2000	7470A	PPB	21 Dec 2018 10:23:15	S	Std
2.5 - 1	52.5	9650	-	2.5000	7470A	PPB	21 Dec 2018 10:25:24	S	Std
5.0 - 1	55	19294	-	5.0000	7470A	PPB	21 Dec 2018 10:27:33	S	Std
7.5 - 1	57.5	28259	-	7.5000	7470A	PPB	21 Dec 2018 10:29:41	S	Std
10.0 - 1	510	38870	-	10.0000	7470A	PPB	21 Dec 2018 10:31:49	S	Std
ICV38 - 1	ICV38	14717	3.8294	-	7470A	PPB	21 Dec 2018 10:34:26	U	SMPL
ICB38 - 1	ICB38	-89	-0.0146	-	7470A	PPB	21 Dec 2018 10:36:32	U	SMPL
CCV76 - 1	CCV76	18716	4.8676	-	7470A	PPB	21 Dec 2018 10:38:39	U	SMPL
CCB76 - 1	CCB76	-67	-0.0089	-	7470A	PPB	21 Dec 2018 10:40:46	U	SMPL
CRA - 1	CRA	761	0.2060	-	7470A	PPB	21 Dec 2018 10:42:53	U	SMPL
HighStd - 1	HighStd	37721	9.8018	-	7470A	PPB	21 Dec 2018 10:45:00	U	SMPL
ChkStd - 1	ChkStd	26260	6.8262	-	7470A	PPB	21 Dec 2018 10:47:07	U	SMPL
PB115896BL - 1	PBW	-144	-0.0289	-	7470A	PPB	21 Dec 2018 10:49:15	U	SMPL
PB115896BS - 1	LCSW	14406	3.7486	-	7470A	PPB	21 Dec 2018 10:51:23	U	SMPL
J6482-04 - 1	R20-OS	-40	-0.0019	-	7470A	PPB	21 Dec 2018 10:53:31	U	SMPL
J6482-04DUP - 1	R20-OSDUP	-6	0.0069	-	7470A	PPB	21 Dec 2018 10:55:39	U	SMPL
J6482-04MS - 1	R20-OSMS	14936	3.8862	-	7470A	PPB	21 Dec 2018 10:57:47	U	SMPL
J6482-04MSD - 1	R20-OSMSD	14570	3.7912	-	7470A	PPB	21 Dec 2018 10:59:56	U	SMPL
J6482-12 - 1	DPC-91468-OS-01	-91	-0.0152	-	7470A	PPB	21 Dec 2018 11:02:03	U	SMPL
CCV77 - 1	CCV77	18924	4.9216	-	7470A	PPB	21 Dec 2018 11:04:10	U	SMPL
CCB77 - 1	CCB77	-55	-0.0058	-	7470A	PPB	21 Dec 2018 11:06:17	U	SMPL
J6493-02 - 1	SOIL-DRUM	-25	0.0020	-	7470A	PPB	21 Dec 2018 11:08:24	U	SMPL
J6493-03 - 1	OILY-DEBRIS	-33	-0.0001	-	7470A	PPB	21 Dec 2018 11:10:31	U	SMPL
J6494-01 - 1	OWS	101	0.0347	-	7470A	PPB	21 Dec 2018 11:12:38	U	SMPL
PB115897BL - 1	PBW	-47	-0.0037	-	7470A	PPB	21 Dec 2018 11:14:45	U	SMPL
PB115897BS - 1	LCSW	14685	3.8211	-	7470A	PPB	21 Dec 2018 11:16:52	U	SMPL
J6465-10 - 1	SO-FB02_121718	-67	-0.0089	-	7470A	PPB	21 Dec 2018 11:19:00	U	SMPL
J6472-03 - 1	OILY-WATER	3456	0.9057	-	7470A	PPB	21 Dec 2018 11:21:08	U	SMPL
J6472-03DUP - 1	OILY-WATERDUP	3243	0.8504	-	7470A	PPB	21 Dec 2018 11:23:17	U	SMPL
J6472-03MS - 1	OILY-WATERMS	19750	5.1361	-	7470A	PPB	21 Dec 2018 11:25:26	U	SMPL
J6472-03MSD - 1	OILY-WATERMSD	19835	5.1581	-	7470A	PPB	21 Dec 2018 11:27:33	U	SMPL
CCV78 - 1	CCV78	18876	4.9092	-	7470A	PPB	21 Dec 2018 11:29:40	U	SMPL
CCB78 - 1	CCB78	-91	-0.0152	-	7470A	PPB	21 Dec 2018 11:31:48	U	SMPL
J6477-04 - 1	EB-2018.12.18	-85	-0.0136	-	7470A	PPB	21 Dec 2018 11:33:55	U	SMPL
PB115896TB - 1	PB115867TB	15	0.0124	-	7470A	PPB	21 Dec 2018 11:36:03	U	SMPL
J6482-04LX5 - 1		-26	0.0017	-	7470A	PPB	21 Dec 2018 11:38:10	U	SMPL
J6482-04A - 1		14728	3.8322	-	7470A	PPB	21 Dec 2018 11:40:18	U	SMPL
J6472-03LX5 - 1		532	0.1466	-	7470A	PPB	21 Dec 2018 11:42:25	U	SMPL
J6472-03A - 1		19733	5.1317	-	7470A	PPB	21 Dec 2018 11:44:33	U	SMPL
CCV79 - 1	CCV79	18745	4.8752	-	7470A	PPB	21 Dec 2018 11:46:41	U	SMPL
CCB79 - 1	CCB79	-87	-0.0141	-	7470A	PPB	21 Dec 2018 11:48:50	U	SMPL