

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB99993

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP70044
calibration std. phosphate 1 ppm	WP70480
calibration std. phosphate 0.5 ppm	WP70479
calibration std. phosphate 0.3 ppm	WP70478
calibration std. 0 ppm	WP70475
phosphate ICV std.	WP70482
phosphate CCV std.	WP70481
calibration std. phosphate 0.1 ppm	WP70477
calibration std. phosphate 0.05 ppm	WP70476
Combined reagent	WP70485
11 N sulfuric acid	WP66905
Phenolphthalein indicator	WP66907
Sodium hydroxide, 1N	WP66924

Intercept: -0.0008Slope: 0.651Regression: 0.999827

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	12/21/2018	12:31
2	CAL2	0.05	1	50	50	0.033	12/21/2018	12:31
3	CAL3	0.10	1	50	50	0.066	12/21/2018	12:32
4	CAL4	0.30	1	50	50	0.186	12/21/2018	12:32
5	CAL5	0.50	1	50	50	0.330	12/21/2018	12:33
6	CAL6	1.00	1	50	50	0.650	12/21/2018	12:34



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB99993

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.324	0.499	12/21/2018	12:35
2	ICB		1	50	50	0.000	0.001	12/21/2018	12:35
3	CCV1	0.50	1	50	50	0.329	0.507	12/21/2018	12:36
4	CCB1		1	50	50	0.000	0.001	12/21/2018	12:36
5	PB115937BL		1	50	50	0.000	0.001	12/21/2018	12:37
6	PB115937BS	0.50	1	50	50	0.319	0.491	12/21/2018	12:37
7	J6483-01		1	50	50	0.007	0.012	12/21/2018	12:38
8	J6483-02		1	50	50	0.008	0.014	12/21/2018	12:38
9	J6483-02DUP		1	50	50	0.010	0.017	12/21/2018	12:39
10	J6483-02MS	0.50	1	50	50	0.287	0.442	12/21/2018	12:39
11	J6483-02MSD	0.50	1	50	50	0.290	0.447	12/21/2018	12:40
12	J6530-01		1	50	50	0.626	0.963	12/21/2018	12:40
13	CCV2	0.50	1	50	50	0.328	0.505	12/21/2018	12:41
14	CCB2		1	50	50	0.000	0.001	12/21/2018	12:41