



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: sweta

Run Number: LB99995

Ambient Barometric Pressure (mmHg): 740

Barometric Scale ID: 051106

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	81.8	12/21/2018	09:40
2	J6496-02		1	9	>100.0	>212.0	>212.0	12/21/2018	10:10
3	J6496-02DUP		1	9	>100.0	>212.0	>212.0	12/21/2018	10:40
4	J6500-01		1	7	>100.0	>212.0	>212.0	12/21/2018	11:15
5	J6503-02		1	6	>100.0	>212.0	>212.0	12/21/2018	11:45
6	J6516-02		1	8	>100.0	>212.0	>212.0	12/21/2018	12:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB99995

WorkList Name : fp122118

WorkList ID : 120689

Date : 12/21/2018 9:00:40 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/25/2018	Water	J6496-02	Flash Point	Cool 4 deg C	PSEG01	E41	WATER-1	12/20/2018	1010A
01/02/2019	Water	J6500-01	Flash Point	Cool 4 deg C	SEAC01	E21	GROUNDWATER	12/19/2018	1010A
12/25/2018	Water	J6503-02	Flash Point	Cool 4 deg C	PSEG01	E21	TP-GROUNDWATER-1-1	12/20/2018	1010A
12/26/2018	Water	J6516-02	Flash Point	Cool 4 deg C	PSEG01	E51	EXCAVATION-WATER	12/20/2018	1010A

Date/Time 12/21/18 9:10 AM
 Received by: KS
 Relinquished by: CP

Date/Time 12/21/18 12:25 PM
 Received by: CP
 Relinquished by: KS