


7471B

LB 99998
INSTRUMENT ID: CY2

Linear

 μ Abs.:

56285



A= 0.0000e+000

B= 1.7804e-004

C= 2.4411e-002

Rho= 0.9998911

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.016	0.016	-47	0.000	-47				
0.2	0.200	0.209	0.009	1034	0.0 %	1034				
2.5	2.500	2.456	-0.044	13655	0.0 %	13655				
5.0	5.000	5.068	0.068	28331	0.0 %	28331				
7.5	7.500	7.406	-0.094	41463	0.0 %	41463				
10.0	10.000	10.045	0.045	56285	0.0 %	56285				

LB99998 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 21 Dec 2018 14:16:59

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-47	-	0.0000	7471B	PPB	21 Dec 2018 14:18:34	S	Std
0.2 - 1	50.2	1034	-	0.2000	7471B	PPB	21 Dec 2018 14:20:52	S	Std
2.5 - 1	52.5	13655	-	2.5000	7471B	PPB	21 Dec 2018 14:23:10	S	Std
5.0 - 1	55	28331	-	5.0000	7471B	PPB	21 Dec 2018 14:25:28	S	Std
7.5 - 1	57.5	41463	-	7.5000	7471B	PPB	21 Dec 2018 14:27:46	S	Std
10.0 - 1	510	56285	-	10.0000	7471B	PPB	21 Dec 2018 14:35:14	S	Std
ICV39 - 1	ICV39	21462	3.8454	-	7471B	PPB	21 Dec 2018 14:37:55	U	SMPL
ICB39 - 1	ICB39	-156	-0.0034	-	7471B	PPB	21 Dec 2018 14:40:10	U	SMPL
CCV80 - 1	CCV80	27402	4.9030	-	7471B	PPB	21 Dec 2018 14:42:27	U	SMPL
CCB80 - 1	CCB80	-143	-0.0010	-	7471B	PPB	21 Dec 2018 14:44:43	U	SMPL
CRA - 1	CRA	992	0.2010	-	7471B	PPB	21 Dec 2018 14:47:00	U	SMPL
HighStd - 1	HighStd	56962	10.1658	-	7471B	PPB	21 Dec 2018 14:49:17	U	SMPL
ChkStd - 1	ChkStd	37429	6.6882	-	7471B	PPB	21 Dec 2018 14:51:33	U	SMPL
PB115903BL - 1	PBS	-146	-0.0016	-	7471B	PPB	21 Dec 2018 14:53:52	U	SMPL
PB115903BS - 1	LCSS	21379	3.8307	-	7471B	PPB	21 Dec 2018 14:56:11	U	SMPL
J6188-03 - 1	BU-01-122018	1057	0.2126	-	7471B	PPB	21 Dec 2018 14:58:28	U	SMPL
J6191-01 - 1	NB-07-122018-A	1506	0.2925	-	7471B	PPB	21 Dec 2018 15:00:46	U	SMPL
J6191-03 - 1	NB-07-122018-B	2833	0.5288	-	7471B	PPB	21 Dec 2018 15:03:04	U	SMPL
J6192-03 - 1	HD-01-122018	1642	0.3167	-	7471B	PPB	21 Dec 2018 15:05:22	U	SMPL
J6200-01 - 1	TR-05-122018	1499	0.2913	-	7471B	PPB	21 Dec 2018 15:07:39	U	SMPL
CCV81 - 1	CCV81	27287	4.8825	-	7471B	PPB	21 Dec 2018 15:09:55	U	SMPL
CCB81 - 1	CCB81	-169	-0.0057	-	7471B	PPB	21 Dec 2018 15:12:11	U	SMPL
J6499-01 - 1	72-11989	35777	6.3940	-	7471B	PPB	21 Dec 2018 15:14:31	U	SMPL
J6500-02 - 1	SOIL-CUTTINGS	664	0.1426	-	7471B	PPB	21 Dec 2018 15:16:48	U	SMPL
J6501-01 - 1	VNJ-221	8347	1.5105	-	7471B	PPB	21 Dec 2018 15:19:06	U	SMPL
J6518-01 - 1	337-SOIL	8227	1.4891	-	7471B	PPB	21 Dec 2018 15:21:23	U	SMPL
J6518-01DUP - 1	337-SOILDUP	9485	1.7131	-	7471B	PPB	21 Dec 2018 15:23:40	U	SMPL
J6518-01MS - 1	337-SOILMS	32036	5.7280	-	7471B	PPB	21 Dec 2018 15:25:58	U	SMPL
J6518-01MSD - 1	337-SOILMSD	32673	5.8414	-	7471B	PPB	21 Dec 2018 15:28:16	U	SMPL
J6520-01 - 1	ROLLOFF-165	27320	4.8884	-	7471B	PPB	21 Dec 2018 15:30:35	U	SMPL
J6518-01LX5 - 1		1523	0.2956	-	7471B	PPB	21 Dec 2018 15:32:53	U	SMPL
J6518-01A - 1		30286	5.4164	-	7471B	PPB	21 Dec 2018 15:35:10	U	SMPL
CCV82 - 1	CCV82	27326	4.8894	-	7471B	PPB	21 Dec 2018 15:37:28	U	SMPL
CCB82 - 1	CCB82	-135	0.0004	-	7471B	PPB	21 Dec 2018 15:39:48	U	SMPL