

LB134086


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

46102



A= 0.0000e+000

B= 2.1317e-004

C= -2.3290e-002

Rho= 0.9994276

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	-0.016	-0.016	34	0.000	34					0.0
0.2	0.200	0.167	-0.033	895	0.0 %	895					-17
2.5	2.500	2.447	-0.053	11588	0.0 %	11588					-2
5.0	5.000	5.091	0.091	23991	0.0 %	23991					2
7.5	7.500	7.707	0.207	36263	0.0 %	36263					4
10.0	10.000	9.804	-0.196	46102	0.0 %	46102					-2

LB134086 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Dec 2024 12:42:17

Sample ID	Extended ID	μ Abs.	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	34	-	0.0000	7471B	PPB	26 Dec 2024 12:48:17	S	Std
0.2 - 1	502	895	-	0.2000	7471B	PPB	26 Dec 2024 12:50:34	S	Std
2.5 - 1	525	11588	-	2.5000	7471B	PPB	26 Dec 2024 12:52:51	S	Std
5.0 - 1	55	23991	-	5.0000	7471B	PPB	26 Dec 2024 12:55:08	S	Std
7.5 - 1	525	36263	-	7.5000	7471B	PPB	26 Dec 2024 12:57:27	S	Std
10.0 - 1	510	46102	-	10.0000	7471B	PPB	26 Dec 2024 13:05:16	S	Std
ICV31 - 1	ICV31	19272	4.0849	-	7471B	PPB	26 Dec 2024 13:10:45	U	SMPL
ICB31 - 1	ICB31	-193	-0.0644	-	7471B	PPB	26 Dec 2024 13:13:01	U	SMPL
CCV67 - 1	CCV67	23993	5.0912	-	7471B	PPB	26 Dec 2024 13:15:19	U	SMPL
CCB67 - 1	CCB67	-237	-0.0738	-	7471B	PPB	26 Dec 2024 13:17:34	U	SMPL
CRA - 1	CRA	882	0.1647	-	7471B	PPB	26 Dec 2024 13:19:54	U	SMPL
HighStd - 1	HighStd	46678	9.9269	-	7471B	PPB	26 Dec 2024 13:22:09	U	SMPL
ChkStd - 1	ChkStd	32810	6.9707	-	7471B	PPB	26 Dec 2024 13:24:25	U	SMPL
PB165862BL - 1	PBS	-324	-0.0924	-	7471B	PPB	26 Dec 2024 13:26:51	U	SMPL
PB165862BS - 1	LCSS	16761	3.5496	-	7471B	PPB	26 Dec 2024 13:29:13	U	SMPL
P5318-01 - 1	AU-06-122024	2553	0.5209	-	7471B	PPB	26 Dec 2024 13:31:30	U	SMPL
P5382-01 - 1	COMP-1	11493	2.4266	-	7471B	PPB	26 Dec 2024 13:40:17	U	SMPL
P5382-01DUP - 1	COMP-1DUP	11353	2.3968	-	7471B	PPB	26 Dec 2024 13:42:33	U	SMPL
P5382-01MS - 1	COMP-1MS	27020	5.7365	-	7471B	PPB	26 Dec 2024 13:47:26	U	SMPL
P5382-01MSD - 1	COMP-1MSD	28626	6.0788	-	7471B	PPB	26 Dec 2024 13:49:56	U	SMPL
CCV68 - 1	CCV68	23863	5.0635	-	7471B	PPB	26 Dec 2024 13:52:17	U	SMPL
CCB68 - 1	CCB68	-218	-0.0698	-	7471B	PPB	26 Dec 2024 13:54:38	U	SMPL
P5382-02 - 1	COMP-2	16494	3.4927	-	7471B	PPB	26 Dec 2024 13:56:57	U	SMPL
P5382-03 - 1	COMP-3	17928	3.7984	-	7471B	PPB	26 Dec 2024 13:59:13	U	SMPL
P5383-01 - 1	OK-02-12232024	109	-0.0001	-	7471B	PPB	26 Dec 2024 14:01:33	U	SMPL
P5382-01LX5 - 1		2613	0.5337	-	7471B	PPB	26 Dec 2024 14:06:42	U	SMPL
P5382-01A - 1		30328	6.4416	-	7471B	PPB	26 Dec 2024 14:09:46	U	SMPL
CCV69 - 1	CCV69	24064	5.1063	-	7471B	PPB	26 Dec 2024 14:12:01	U	SMPL
CCB69 - 1	CCB69	-170	-0.0595	-	7471B	PPB	26 Dec 2024 14:14:23	U	SMPL