

7471A


LB89720

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39744



A= 0.0000e+000

B= 2.5315e-004

C= -2.9448e-002

Rho= 0.9999543

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.023	-0.023	27	0.000	27				
0.2	0.200	0.227	0.027	1012	0.0 %	1012				
2.5	2.500	2.541	0.041	10152	0.0 %	10152				
5.0	5.000	4.943	-0.057	19644	0.0 %	19644				
7.5	7.500	7.480	-0.020	29665	0.0 %	29665	0			
10.0	10.000	10.032	0.032	39744	0.0 %	39744				

LB89720 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 25 Aug 2017 15:11:10

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	27	-	0.00007471A	PPB		25 Aug 2017 15:15:15	S	Std
0.2 - 1	50.2	1012	-	0.20007471A	PPB		25 Aug 2017 15:17:33	S	Std
2.5 - 1	52.5	10152	-	2.50007471A	PPB		25 Aug 2017 15:19:51	S	Std
5.0 - 1	55	19644	-	5.00007471A	PPB		25 Aug 2017 15:22:08	S	Std
7.5 - 1	57.5	29665	-	7.50007471A	PPB		25 Aug 2017 15:24:25	S	Std
10.0 - 1	510	39744	-	10.00007471A	PPB		25 Aug 2017 15:29:18	S	Std
ICV30 - 1	ICV30	16536	4.1566	-7471A	PPB		25 Aug 2017 15:32:10	U	SMPL
ICB30 - 1	ICB30	-166	-0.0715	-7471A	PPB		25 Aug 2017 15:34:25	U	SMPL
CCV59 - 1	CCV59	19619	4.9371	-7471A	PPB		25 Aug 2017 15:36:41	U	SMPL
CCB59 - 1	CCB59	-163	-0.0707	-7471A	PPB		25 Aug 2017 15:38:57	U	SMPL
CRA - 1	CRA	928	0.2055	-7471A	PPB		25 Aug 2017 15:41:13	U	SMPL
HighStd - 1	HighStd	40311	10.1752	-7471A	PPB		25 Aug 2017 15:43:29	U	SMPL
ChkStd - 1	ChkStd	27510	6.9347	-7471A	PPB		25 Aug 2017 15:45:45	U	SMPL
PB101808BL - 1	PBS	-234	-0.0887	-7471A	PPB		25 Aug 2017 15:48:03	U	SMPL
PB101808BS - 1	LCSS	14568	3.6584	-7471A	PPB		25 Aug 2017 15:50:21	U	SMPL
I4943-01 - 1	OR-01-082317	1055	0.2376	-7471A	PPB		25 Aug 2017 15:52:38	U	SMPL
I4944-01 - 1	4-1-DRUMS-COMP	3642	0.8925	-7471A	PPB		25 Aug 2017 15:54:56	U	SMPL
I4945-09 - 1	SP-1-4-COMP	471	0.0898	-7471A	PPB		25 Aug 2017 15:57:13	U	SMPL
I4946-01 - 1	VNJ-211	967	0.2153	-7471A	PPB		25 Aug 2017 15:59:31	U	SMPL
I4946-01DUP - 1	VNJ-211DUP	973	0.2169	-7471A	PPB		25 Aug 2017 16:01:46	U	SMPL
CCV60 - 1	CCV60	19646	4.9439	-7471A	PPB		25 Aug 2017 16:04:02	U	SMPL
CCB60 - 1	CCB60	-138	-0.0644	-7471A	PPB		25 Aug 2017 16:06:19	U	SMPL
I4946-01MS - 1	VNJ-211MS	16841	4.2338	-7471A	PPB		25 Aug 2017 16:08:35	U	SMPL
I4946-01MSD - 1	VNJ-211MSD	17100	4.2994	-7471A	PPB		25 Aug 2017 16:10:50	U	SMPL
I4947-01 - 1	ELIZABETH-RO-COMP-A	33	-0.0211	-7471A	PPB		25 Aug 2017 16:13:07	U	SMPL
I4947-06 - 1	ELIZABETH-RO-COMP-B	149	0.0083	-7471A	PPB		25 Aug 2017 16:15:24	U	SMPL
I4948-01 - 1	MANHOLE-A-COMP	1395	0.3237	-7471A	PPB		25 Aug 2017 16:17:41	U	SMPL
I4948-04 - 1	TP-2-COMP	1474	0.3437	-7471A	PPB		25 Aug 2017 16:19:57	U	SMPL
I4946-01LX5 - 1		188	0.0181	-7471A	PPB		25 Aug 2017 16:22:15	U	SMPL
I4946-01A - 1		17281	4.3452	-7471A	PPB		25 Aug 2017 16:24:32	U	SMPL
CCV61 - 1	CCV61	20124	5.0649	-7471A	PPB		25 Aug 2017 16:26:50	U	SMPL
CCB61 - 1	CCB61	-43	-0.0403	-7471A	PPB		25 Aug 2017 16:29:06	U	SMPL