

SOP ID : M3010A-Digestion-15

SDG No : N/A

Matrix : WATER

Pipette ID: ICP-A

Balance ID : N/A

Filter paper ID : N/A

pH Strip ID : M4462

Hood ID : #3

Start Digest Date: 05/07/2020 Time : 08:31 Temp : 96 °C

End Digest Date: 05/07/2020 Time : 12:45 Temp : 96 °C

Digestion tube ID: M4534

Block thermometer ID: MET-DIG. #2

Dig Technician Signature: *D.P.*Supervisor Signature: *S.T.*

Block ID/Temp : 1. N/A 2. 96°C 3. N/A 4. N/A

Standardized Name	MLS USED	STD REF. # FROM LOG
LFS-1	0.25	M4477
LFS-2	0.25	M4482
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
CONC: HNO3	3.00	M4545
1:1 HCL	5.00	MP56729
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

N/A

Date / Time	Received By	Relinquished By	Location
5/7/20 13:05	<i>S.T.</i>	<i>D.P.</i>	Metals Lab
	Analysis Group	Preparation Group	

Lab Sample ID	Client Sample ID	pH	Initial Vol (ml)	Final Vol (ml)	Color Before	Color After	Clarity Before	Clarity After	Comment	Prep Pos
L2510-01	B-1	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	2-16
L2510-01DUP	B-1DUP	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	17
L2510-01MS	B-1MS	<2	50	25	Colorless	Colorless	Clear	Clear	M4477,M4482	18
L2510-01MSD	B-1MSD	<2	50	25	Colorless	Colorless	Clear	Clear	M4477,M4482	19
L2510-02	G-1	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	20
L2516-01	S-7	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	21
L2516-02	S-8	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	22
PB128706BL	PBW706	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	23
PB128706BS	LCS706	<2	50	25	Colorless	Colorless	Clear	Clear	M4477,M4482	24
PB128706TB	LEB706	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	25