

SOP ID : M3010A-Digestion-15

SDG No : N/A

Matrix : WATER

Pipette ID: ICP-A

Balance ID : N/A

Filter paper ID : N/A

pH Strip ID : M4588

Hood ID : #3

Start Digest Date: 08/03/2020 Time : 08:02 Temp : 95 °C

End Digest Date: 08/03/2020 Time : 12:27 Temp : 95 °C

Digestion tube ID: M4578

Block thermometer ID: MET-DIG. #2

Dig Technician Signature: Supervisor Signature: 

Block ID/Temp : 1. N/A 2. 95°C 3. N/A 4. N/A

Standard Name	MLS USED	STD REF. # FROM LOG
LFS-1	0.25	M4566
LFS-2	0.25	M4484
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
CONC: HNO3	3.00	M4586
1:1 HCL	5.00	MP56729
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

N/A

Date / Time	Received By	Relinquished By	Location
8/3/20 12:47			Metals Lab
	Analysis Group	Preparation Group	

Lab Sample ID	Client Sample ID	pH	Initial Vol (ml)	Final Vol (ml)	Color Before	Color After	Clarity Before	Clarity After	Comment	Prep Pos
L3489-15	FB	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	2-26
L3523-01	20200447-FIELD-BLANK	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	27
L3523-02	20200448-SITE-WATER	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	28
L3523-02DUP	20200448-SITE-WATERDU	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	29
L3523-02MS	20200448-SITE-WATERMS	<2	50	25	Colorless	Colorless	Clear	Clear	M4566,M4484	30
L3523-02MSD	20200448-SITE-WATERMS	<2	50	25	Colorless	Colorless	Clear	Clear	M4566,M4484	31
PB130706BL	PBW706	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	32
PB130706BS	LCS706	<2	50	25	Colorless	Colorless	Clear	Clear	M4566,M4484	33