

SOP ID : M200.7-Trace Elements-20

SDG No : N/A

Matrix : WATER

Pipette ID: ICP-A

Balance ID : N/A

Filter paper ID : N/A

pH Strip ID : M4588

Hood ID : #3

Start Digest Date: 10/02/2020 Time : 10:22 Temp : 95 °C

End Digest Date: 10/02/2020 Time : 13:54 Temp : 95 °C

Digestion tube ID: M4633

Block thermometer ID: MET-DIG. #2

Dig Technician Signature: D.P.

Supervisor Signature: 

Block ID/Temp : 1. N/A 2. 95°C 3. N/A 4. N/A

Standardized Name	MLS USED	STD REF. # FROM LOG
LFS-1	0.25	M4567
LFS-2	0.25	M4573
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
1:1 HNO3	1.00	MP57471
1:1 HCL	0.50	MP59660
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

N/A

Date / Time	Received By	Relinquished By	Location
10/2/20 14:14		D.P.	Metals Lab
	Analysis Group	Preparation Group	

Lab Sample ID	Client Sample ID	pH	Initial Vol (ml)	Final Vol (ml)	Color Before	Color After	Clarity Before	Clarity After	Comment	Prep Pos
L4163-01	JS9600	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	2-16
L4187-01	CELLS-B/C	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	17
L4233-01	55-ST-29	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	18
L4233-01DUP	55-ST-29DUP	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	19
L4233-01MS	55-ST-29MS	<2	50	25	Colorless	Colorless	Clear	Clear	M4567,M4573	20
L4233-01MSD	55-ST-29MSD	<2	50	25	Colorless	Colorless	Clear	Clear	M4567,M4573	21
PB132054BL	PBW054	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	22
PB132054BS	LCS054	<2	50	25	Colorless	Colorless	Clear	Clear	M4567,M4573	23