

SOP ID : M3010A-Digestion-15

SDG No : N/A

Matrix : WATER

Pipette ID: ICP-A

Balance ID : N/A

Filter paper ID : N/A

pH Strip ID : M4588

Hood ID : #3

Block ID/Temp : 1. N/A 2. N/A 3. 95°C 4. N/A

Start Digest Date: 02/17/2021 Time : 12:41 Temp : 95 °C

End Digest Date: 02/17/2021 Time : 15:59 Temp : 98 °C

Digestion tube ID: M4713

Block thermometer ID: MET-DIG. #3

Dig Technician Signature: *D.P.*

Supervisor Signature: *JD*

Standard Name	MLS USED	STD REF. # FROM LOG
LFS-1	0.25	M4647
LFS-2	0.25	M4651
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
CONC: HNO3	3.00	M4754
1:1 HCL	5.00	MP59661
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

N/A

Date / Time	Received By	Relinquished By	Location
2/17/21 16:19	<i>JD</i>	<i>D.P.</i>	Metals Lab
	Analysis Group	Preparation Group	

Lab Sample ID	Client Sample ID	pH	Initial Vol (ml)	Final Vol (ml)	Color Before	Color After	Clarity Before	Clarity After	Comment	Prep Pos
M1404-01	NWB-1579-1582	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	3-28
M1431-01	IDW-BR2-AQ-BR-021621	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	29
M1432-03	GWBR2-SC-93-021621	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	30
M1432-03DUP	GWBR2-SC-93-021621DUP	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	31
M1432-03MS	GWBR2-SC-93-021621MS	<2	50	25	Colorless	Colorless	Clear	Clear	M4647,M4651	32
M1432-03MS	GWBR2-SC-93-021621MS	<2	50	25	Colorless	Colorless	Clear	Clear	M4647,M4651	33
PB134706BL	PBW706	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	34
PB134706BS	LCS706	<2	50	25	Colorless	Colorless	Clear	Clear	M4647,M4651	35