

SOP ID : M1312-SPLP-9

SDG No : N/A

Weigh By : JP

Balance ID : WC SC-4

pH Meter ID : WC PH METER-1

Extraction By : JP

Filter By : JP

Pipette ID : WC

Tumbler ID : T-1 /T-2

TCLP Filter ID : 113075

Start Prep Date : 06/14/2021 Time : 15:30

End Prep Date : 06/08/2021 Time : 08:35

Combination Ratio : 20

ZHE Cleaning Batch : N/A

Initial Room Temperature: 21 °C

Final Room Temperature: 21 °C

TCLP Technician Signature : JP

Supervisor By : al

Standard Name	MLS USED	STD REF. # FROM LOG
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE US	Lot Number
SPLP FLUID	N/A	WP90944
N/A	N/A	N/A
DI WATER	W2606, P.H 5.17	N/A
pH Strips	N/A	W1931,W1934,W2755,W2350
pH Strips	N/A	W1937,W1938,W1939,W1940,W1941,W1942
1L Amber bottle	N/A	19948
120ml Plastic bottle	N/A	210210
N/A	N/A	N/A

## Extraction Conformance/Non-Conformance Comments:

Matrix spikes are added after filtration and before preservation. M2678-15 IS USED FOR MS-MSD. Tumbler T-1/T-2 CHECKED, 30 RPM.

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
06/15/21 11:00	JP / T-2	RJ / Ext
	Preparation Group	Analysis Group