

SOP ID : M200.7-Trace Elements-21

SDG No : N/A

Matrix : WATER

Pipette ID : ICP-A

Balance ID : N/A

Filter paper ID : N/A

pH Strip ID : M4588

Hood ID : #3


Block ID/Temp : 1. N/A 2. N/A 3. 95°C 4. N/A

Start Digest Date: 08/06/2021 Time : 10:10 Temp : 95 °C

End Digest Date: 08/06/2021 Time : 14:45 Temp : 95 °C

Digestion tube ID: M4845

Block thermometer ID: MET-DIG.#3

Dig Technician Signature: Supervisor Signature: 

Standard Name	MLS USED	STD REF. # FROM LOG
LFS-1	0.25	M4812
LFS-2	0.25	M4820
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
1:1 HNO3	1.00	MP64100
1:1 HCL	0.50	MP63865
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

N/A

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
08/06/21 15:10	BH/Dig Lab	SO (metablab)
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	pH	Initial Vol (ml)	Final Vol (ml)	Color Before	Color After	Clarity Before	Clarity After	Comment	Prep Pos
M3262-01	001-WILLETS-PT-BLVD(AUG)	<2	50	25	Colorless	Colorless	Cloudy	Cloudy	C	3
M3262-02	002-35TH-AVE(AUG)	<2	50	25	Colorless	Colorless	Cloudy	Cloudy	C	4
M3262-02DUP	002-35TH-AVE(AUG)DUP	<2	50	25	Colorless	Colorless	Cloudy	Cloudy	C	5
M3262-02MS	002-35TH-AVE(AUG)MS	<2	50	25	Colorless	Colorless	Cloudy	Cloudy	M4812,M4820	6
M3262-02MS	002-35TH-AVE(AUG)MSD	<2	50	25	Colorless	Colorless	Cloudy	Cloudy	M4812,M4820	7
M3281-03	TWP01	<2	50	25	Black	Black	Cloudy	Cloudy	F	8
M3281-04	TWP01	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	9
M3281-05	TWP02	<2	50	25	Black	Black	Cloudy	Cloudy	G	10
M3281-06	TWP02	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	11
M3281-07	TWP03	<2	50	25	Black	Black	Cloudy	Cloudy	E	12
M3281-08	TWP03	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	13
M3285-01	GRAB	<2	50	25	Black	Yellow	Cloudy	Cloudy	A	14
M3302-01	32220008-3	<2	50	25	Colorless	Colorless	Clear	Clear	B	15
M3303-01	315220012-1	<2	50	25	Yellow	Yellow	Cloudy	Cloudy	B	16
M3311-01	POOL-WATER	<2	50	25	Colorless	Colorless	Clear	Clear	B	17
PB138233BL	PBW233	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	3-1
PB138233BS	LCS233	<2	50	25	Colorless	Colorless	Clear	Clear	M4812,M4820	2