

211000080

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : As

12/28/2018 9:48

Test: *Reactive*
Total CN

Sample Id	<i>ug/L</i> Result	Dil. 1 +	Response	Errors
ICV1	95.714	0.0	0.090	
ICB1	-0.640	0.0	0.004	
CCV1	241.189	0.0	0.219	
CCB1	-0.768	0.0	0.004	
PB116013BL	-0.651	0.0	0.004	
J6575-02	-0.588	0.0	0.004	
J6575-02DUP	-0.609	0.0	0.004	
PB116012BL	-0.676	0.0	0.004	
J6195-12	-0.333	0.0	0.004	
J6195-12DUP	-0.254	0.0	0.004	
J6195-14	-0.739	0.0	0.004	
J6579-03	-0.563	0.0	0.004	
CCV2	257.558	0.0	0.234	
CCB2	-0.450	0.0	0.004	

N	14
Mean	42.014
SD	91.5512
CV%	217.91

Aquakem v. 7.2AQ1

Results from time period:

Fri Dec 28 09:29:55 2018

Fri Dec 28 09:33:48 2018

Sample Id	Sam/Ctr/c/	Test short nan	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Reactive CN	P	-0.9065	µg/l	12/28/2018 8:01:14	
5.0PPBCN	A	Reactive CN	P	4.1588	µg/l	12/28/2018 8:01:15	
10PPBCN	A	Reactive CN	P	9.0998	µg/l	12/28/2018 8:01:16	
50PPBCN	A	Reactive CN	P	49.9548	µg/l	12/28/2018 8:01:17	
100PPBCN	A	Reactive CN	P	102.2571	µg/l	12/28/2018 8:01:18	
250PPBCN	A	Reactive CN	P	251.713	µg/l	12/28/2018 8:01:19	
500PPBCN	A	Reactive CN	P	498.7231	µg/l	12/28/2018 8:01:20	
ICV1	S	Reactive CN	P	95.7136	µg/l	12/28/2018 9:29:55	
ICB1	S	Reactive CN	P	-0.6396	µg/l	12/28/2018 9:29:56	
CCV1	S	Reactive CN	P	241.1893	µg/l	12/28/2018 9:29:57	
CCB1	S	Reactive CN	P	-0.7675	µg/l	12/28/2018 9:29:58	
PB116013BL	S	Reactive CN	P	-0.6506	µg/l	12/28/2018 9:29:59	
J6575-02	S	Reactive CN	P	-0.5878	µg/l	12/28/2018 9:30:00	
J6575-02DUP	S	Reactive CN	P	-0.6091	µg/l	12/28/2018 9:30:01	
PB116012BL	S	Reactive CN	P	-0.6761	µg/l	12/28/2018 9:30:02	
J6195-12	S	Reactive CN	P	-0.3331	µg/l	12/28/2018 9:30:03	
J6195-12DUP	S	Reactive CN	P	-0.254	µg/l	12/28/2018 9:30:04	
J6195-14	S	Reactive CN	P	-0.7387	µg/l	12/28/2018 9:30:05	
J6579-03	S	Reactive CN	P	-0.5628	µg/l	12/28/2018 9:33:46	
CCV2	S	Reactive CN	P	257.5579	µg/l	12/28/2018 9:33:47	
CCB2	S	Reactive CN	P	-0.4497	µg/l	12/28/2018 9:33:48	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/28/2018 8:02


Test Total CN

Accepted 12/28/2018 8:02

Factor 1122
Bias 0.004

Coeff. of det. 0.999942

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.9065	0.0000	
2	5.0PPBCN	0.008	4.1588	5.0000	
3	10PPBCN	0.012	9.0998	10.0000	
4	50PPBCN	0.049	49.9548	50.0000	
5	100PPBCN	0.095	102.2571	100.0000	
6	250PPBCN	0.229	251.7130	250.0000	
7	500PPBCN	0.449	498.7231	500.0000	