

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AB

1/14/2019 16:40

Test: ~~Total~~ ^{Reactive} CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.891	0.0	0.088	
ICB1	1.533	0.0	0.004	
CCV1	236.798	0.0	0.211	
CCB1	1.457	0.0	0.004	
PB116359BL	1.419	0.0	0.004	
K1015-06	1.448	0.0	0.004	
K1015-08	1.244	0.0	0.004	
K1126-02	1.395	0.0	0.004	
K1128-02	1.217	0.0	0.004	
K1128-04	1.407	0.0	0.004	
K1128-04DUP	1.341	0.0	0.004	
PB116373BL	1.057	0.0	0.004	
K1127-03	1.277	0.0	0.004	
K1139-22	1.295	0.0	0.004	
CCV2	243.008	0.0	0.216	
CCB2	1.241	0.0	0.004	
K1139-22DUP	1.401	0.0	0.004	
CCV3	242.735	0.0	0.216	
CCB3	1.477	0.0	0.004	

N	19
Mean	44.192
SD	90.1716
CV%	204.05

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

1/14/2019 14:32


Test Total CN

Accepted 1/14/2019 14:32

Factor 1138
Bias 0.003

Coeff. of det. 0.999860

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	0.2513	0.0000	
2	5.0PPBCN	0.008	5.6379	5.0000	
3	10PPBCN	0.012	10.4481	10.0000	
4	50PPBCN	0.047	50.3557	50.0000	
5	100PPBCN	0.091	100.9159	100.0000	
6	250PPBCN	0.218	245.2504	250.0000	
7	500PPBCN	0.444	502.1407	500.0000	

Aquakem v. 7.2AQ1

Results from time period:

Mon Jan 14 16:34:35 2019

Mon Jan 14 16:40:53 2019

Sample Id	Sam/Ctr/c/	Test short name	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	0.2513	µg/l	1/14/2019 14:31:50	
5.OPPBCN	A	Reactive CN	P	5.6379	µg/l	1/14/2019 14:31:51	
10PPBCN	A	Reactive CN	P	10.4481	µg/l	1/14/2019 14:31:52	
50PPBCN	A	Reactive CN	P	50.3557	µg/l	1/14/2019 14:31:53	
100PPBCN	A	Reactive CN	P	100.9159	µg/l	1/14/2019 14:31:54	
250PPBCN	A	Reactive CN	P	245.2504	µg/l	1/14/2019 14:31:55	
500PPBCN	A	Reactive CN	P	502.1407	µg/l	1/14/2019 14:31:56	
ICV1	S	Reactive CN	P	96.8909	µg/l	1/14/2019 16:34:35	
ICB1	S	Reactive CN	P	1.533	µg/l	1/14/2019 16:34:36	
CCV1	S	Reactive CN	P	236.7976	µg/l	1/14/2019 16:34:37	
CCB1	S	Reactive CN	P	1.4568	µg/l	1/14/2019 16:34:38	
PB116359BL	S	Reactive CN	P	1.4187	µg/l	1/14/2019 16:34:39	
K1015-06	S	Reactive CN	P	1.4476	µg/l	1/14/2019 16:34:40	
K1015-08	S	Reactive CN	P	1.2444	µg/l	1/14/2019 16:34:41	
K1126-02	S	Reactive CN	P	1.3948	µg/l	1/14/2019 16:34:42	
K1128-02	S	Reactive CN	P	1.2166	µg/l	1/14/2019 16:34:43	
K1128-04	S	Reactive CN	P	1.4067	µg/l	1/14/2019 16:34:44	
K1128-04DUP	S	Reactive CN	P	1.3413	µg/l	1/14/2019 16:34:45	
PB116373BL	S	Reactive CN	P	1.0572	µg/l	1/14/2019 16:40:46	
K1127-03	S	Reactive CN	P	1.2774	µg/l	1/14/2019 16:40:47	
K1139-22	S	Reactive CN	P	1.2945	µg/l	1/14/2019 16:40:48	
CCV2	S	Reactive CN	P	243.0077	µg/l	1/14/2019 16:40:49	
CCB2	S	Reactive CN	P	1.2414	µg/l	1/14/2019 16:40:50	
K1139-22DUP	S	Reactive CN	P	1.4009	µg/l	1/14/2019 16:40:51	
CCV3	S	Reactive CN	P	242.7348	µg/l	1/14/2019 16:40:52	
CCB3	S	Reactive CN	P	1.4766	µg/l	1/14/2019 16:40:53	