

2B100507

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

1/28/2019 12:37

Test: ~~Total~~ ^{Detective} CN

Mg/L

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	94.923	0.0	0.088	
ICB1	0.755	0.0	0.004	
CCV1	245.558	0.0	0.221	
CCB1	0.702	0.0	0.004	
PB116661BL	0.479	0.0	0.004	
K1011-10	-0.353	0.0	0.003	
K1011-10DUP	0.387	0.0	0.004	
K1011-12	0.441	0.0	0.004	
K1250-03	0.015	0.0	0.004	
K1254-02	0.885	0.0	0.004	
K1254-04	0.342	0.0	0.004	
K1255-02	1.376	0.0	0.005	
K1261-02	0.406	0.0	0.004	
K1261-04	0.431	0.0	0.004	
CCV2	248.707	0.0	0.224	
CCB2	1.118	0.0	0.005	
PB116664BL	0.864	0.0	0.004	
K1250-05	0.741	0.0	0.004	
K1250-05DUP	0.978	0.0	0.004	
CCV3	251.196	0.0	0.226	
CCB3	0.706	0.0	0.004	

N 21
Mean 40.508
SD 89.3908
CV% 220.68

AS 01/28/19
PB116662 BL

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

1/28/2019 11:15


Test Total CN

Accepted 1/28/2019 11:14

Factor 1129
Bias 0.004

Coeff. of det. 0.999735

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	0.0287	0.0000	
2	5.0PPBCN	0.008	4.6490	5.0000	
3	10PPBCN	0.012	9.2069	10.0000	
4	50PPBCN	0.047	48.9852	50.0000	
5	100PPBCN	0.091	98.4493	100.0000	
6	250PPBCN	0.231	256.4998	250.0000	
7	500PPBCN	0.444	497.1811	500.0000	

Aquakem v. 7.2AQ1

Results from time period:

Mon Jan 28 10:46:55 2019

Mon Jan 28 12:36:31 2019



Sample Id	Sam/Ctr/c/	Test short nan	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	0.0287	µg/l	1/28/2019 10:46:55	
5.OPPBCN	A	Reactive CN	P	4.649	µg/l	1/28/2019 10:46:56	
10PPBCN	A	Reactive CN	P	9.2069	µg/l	1/28/2019 10:46:57	
50PPBCN	A	Reactive CN	P	48.9852	µg/l	1/28/2019 10:46:58	
100PPBCN	A	Reactive CN	P	98.4493	µg/l	1/28/2019 10:46:59	
250PPBCN	A	Reactive CN	P	256.4998	µg/l	1/28/2019 10:47:00	
500PPBCN	A	Reactive CN	P	497.1811	µg/l	1/28/2019 10:47:01	
ICV1	S	Reactive CN	P	94.9233	µg/l	1/28/2019 12:11:03	
ICB1	S	Reactive CN	P	0.7551	µg/l	1/28/2019 12:11:04	
CCV1	S	Reactive CN	P	245.5582	µg/l	1/28/2019 12:11:05	
CCB1	S	Reactive CN	P	0.702	µg/l	1/28/2019 12:11:06	
PB116661BL	S	Reactive CN	P	0.4788	µg/l	1/28/2019 12:11:07	
K1011-10	S	Reactive CN	P	-0.3525	µg/l	1/28/2019 12:11:08	
K1011-10DUP	S	Reactive CN	P	0.387	µg/l	1/28/2019 12:11:09	
K1011-12	S	Reactive CN	P	0.4411	µg/l	1/28/2019 12:11:10	
K1250-03	S	Reactive CN	P	0.0152	µg/l	1/28/2019 12:11:11	
K1254-02	S	Reactive CN	P	0.885	µg/l	1/28/2019 12:11:12	
K1254-04	S	Reactive CN	P	0.3417	µg/l	1/28/2019 12:11:13	
K1255-02	S	Reactive CN	P	1.3761	µg/l	1/28/2019 12:18:10	
K1261-02	S	Reactive CN	P	0.4059	µg/l	1/28/2019 12:18:11	
K1261-04	S	Reactive CN	P	0.4313	µg/l	1/28/2019 12:18:12	
CCV2	S	Reactive CN	P	248.7066	µg/l	1/28/2019 12:18:13	
CCB2	S	Reactive CN	P	1.1184	µg/l	1/28/2019 12:18:14	
PB116662BL	S	Reactive CN	P	0.8643	µg/l	1/28/2019 12:18:15	
K1250-05	S	Reactive CN	P	0.7411	µg/l	1/28/2019 12:18:16	
K1250-05DUP	S	Reactive CN	P	0.9778	µg/l	1/28/2019 12:18:17	
CCV3	S	Reactive CN	P	251.1958	µg/l	1/28/2019 12:18:18	
CCB3	S	Reactive CN	P	0.706	µg/l	1/28/2019 12:18:19	