

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

1/30/2019 11:04

Test: Phenolics-

Sample Id	Result ^{mg/d}	Dil. 1 +	Response	Errors
ICV1	1.039	0.0	0.126	
ICB1	-0.000	0.0	0.014	
CCV1	1.007	0.0	0.122	
CCB1	0.000	0.0	0.015	
PB116638BL	0.002	0.0	0.015	
PB116638BS	1.042	0.0	0.126	
K1230-01	0.009	0.0	0.015	
K1230-01DUP	0.007	0.0	0.015	
K1260-01	0.025	0.0	0.017	
K1260-02	-0.006	0.0	0.014	
K1260-03	-0.001	0.0	0.014	
K1230-01MS	1.020	0.0	0.124	
K1230-01MSD	1.024	0.0	0.124	
CCV2	1.009	0.0	0.123	
CCB2	-0.004	0.0	0.014	

N	15
Mean	0.412
SD	0.5174
CV%	125.71

Aquakem v. 7.2AQ1

Results from time period:

Wed Jan 30 10:05:51 2019

Wed Jan 30 10:55:03 2019

AS

Sample Id	Sam/Ctr/c/	Test short nam	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0505	mg/l	1/30/2019 10:05:51	
0.1PPM	A	Phenolics-	P	0.1043	mg/l	1/30/2019 10:05:52	
0.25PPM	A	Phenolics-	P	0.2522	mg/l	1/30/2019 10:05:53	
0.50PPM	A	Phenolics-	P	0.5055	mg/l	1/30/2019 10:05:54	
1.0PPM	A	Phenolics-	P	0.9788	mg/l	1/30/2019 10:05:55	
2.0PPM	A	Phenolics-	P	2.0087	mg/l	1/30/2019 10:05:56	
ICV1	S	Phenolics-	P	1.0395	mg/l	1/30/2019 10:49:37	
ICB1	S	Phenolics-	P	-0.0005	mg/l	1/30/2019 10:49:38	
CCV1	S	Phenolics-	P	1.007	mg/l	1/30/2019 10:49:39	
CCB1	S	Phenolics-	P	0.0001	mg/l	1/30/2019 10:49:40	
PB116638BL	S	Phenolics-	P	0.0021	mg/l	1/30/2019 10:49:41	
PB116638BS	S	Phenolics-	P	1.0422	mg/l	1/30/2019 10:49:42	
K1230-01	S	Phenolics-	P	0.009	mg/l	1/30/2019 10:49:43	
K1230-01DUP	S	Phenolics-	P	0.007	mg/l	1/30/2019 10:49:44	
K1260-01	S	Phenolics-	P	0.0246	mg/l	1/30/2019 10:49:47	
K1260-02	S	Phenolics-	P	-0.0059	mg/l	1/30/2019 10:49:48	
K1260-03	S	Phenolics-	P	-0.0007	mg/l	1/30/2019 10:54:59	
K1230-01MS	S	Phenolics-	P	1.0199	mg/l	1/30/2019 10:55:00	
K1230-01MSD	S	Phenolics-	P	1.0245	mg/l	1/30/2019 10:55:01	
CCV2	S	Phenolics-	P	1.0092	mg/l	1/30/2019 10:55:02	
CCB2	S	Phenolics-	P	-0.0038	mg/l	1/30/2019 10:55:03	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

1/30/2019 10:14


Test Phenolics-

Accepted 1/30/2019 10:14

Factor 9.329
Bias 0.015

Coeff. of det. 0.999792

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.020	0.0505	0.0500	
2	Phenol-2	0.026	0.1043	0.1000	
3	Phenol-2	0.042	0.2522	0.2500	
4	Phenol-2	0.069	0.5055	0.5000	
5	Phenol-2	0.119	0.9788	1.0000	
6	Phenol-2	0.230	2.0087	2.0000	