

=====

Test results

Aquakem 7.2AQ1

2B101990
Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

4/10/2019 11:26

Test: Total CN

Mg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	93.318	0.0	0.090	
ICB1	0.445	0.0	0.004	
CCV1	235.539	0.0	0.221	
CCB1	0.103	0.0	0.004	
PB118754BL	0.425	0.0	0.004	
PB118754BS	93.517	0.0	0.090	
MIDPB118754	236.838	0.0	0.222	
K2241-10	1.433	0.0	0.005	
K2241-11	8.567	0.0	0.012	
CCV2	237.928	0.0	0.223	
CCB2	0.155	0.0	0.004	

N 11
Mean 82.570
SD 105.2211
CV% 127.43

Aquakem v. 7.2AQ1

Results from time period:

Wed Apr 10 11:15:03 2019

Wed Apr 10 11:15:13 2019

✓AS

Sample Id	Sam/Ctr/c	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	0.1827	µg/l	4/10/2019 8:21:05	
5.0PPBCN	A	Total CN	P	5.2169	µg/l	4/10/2019 8:21:06	
10PPBCN	A	Total CN	P	9.9602	µg/l	4/10/2019 8:21:07	
50PPBCN	A	Total CN	P	49.8699	µg/l	4/10/2019 8:21:08	
100PPBCN	A	Total CN	P	99.3044	µg/l	4/10/2019 8:21:09	
250PPBCN	A	Total CN	P	250.6301	µg/l	4/10/2019 8:21:10	
500PPBCN	A	Total CN	P	499.8358	µg/l	4/10/2019 8:21:11	
ICV1	S	Total CN	P	93.3182	µg/l	4/10/2019 11:15:03	
ICB1	S	Total CN	P	0.445	µg/l	4/10/2019 11:15:04	
CCV1	S	Total CN	P	235.5392	µg/l	4/10/2019 11:15:05	
CCB1	S	Total CN	P	0.103	µg/l	4/10/2019 11:15:06	
PB118754BL	S	Total CN	P	0.425	µg/l	4/10/2019 11:15:07	
PB118754BS	S	Total CN	P	93.5165	µg/l	4/10/2019 11:15:08	
MIDPB118754	S	Total CN	P	236.8377	µg/l	4/10/2019 11:15:09	
K2241-10	S	Total CN	P	1.4326	µg/l	4/10/2019 11:15:10	
K2241-11	S	Total CN	P	8.5673	µg/l	4/10/2019 11:15:11	
CCV2	S	Total CN	P	237.9285	µg/l	4/10/2019 11:15:12	
CCB2	S	Total CN	P	0.1549	µg/l	4/10/2019 11:15:13	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

4/10/2019 8:23


Test Total CN

Accepted 4/10/2019 8:23

Factor 1086
Bias 0.004

Coeff. of det. 0.999995

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	0.1827	0.0000	
2	5.0PPBCN	0.009	5.2169	5.0000	
3	10PPBCN	0.013	9.9602	10.0000	
4	50PPBCN	0.050	49.8699	50.0000	
5	100PPBCN	0.095	99.3044	100.0000	
6	250PPBCN	0.235	250.6301	250.0000	
7	500PPBCN	0.464	499.8358	500.0000	