

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092

Reviewed by : AS

Instrument ID : Konelab

4/15/2019 13:57

Test: Total CN

ug/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.601	0.0	0.093	
ICB1	0.329	0.0	0.004	
CCV1	239.100	0.0	0.228	
CCB1	-0.014	0.0	0.004	
PB118888BL	0.159	0.0	0.004	
PB118888BS	97.231	0.0	0.095	
K2397-17MS	5.366	0.0	0.009	
K2397-17MSD	5.102	0.0	0.009	
K2397-17	1.731	0.0	0.005	
K2397-17DUP	1.541	0.0	0.005	
CCV2	239.043	0.0	0.228	
CCB2	0.180	0.0	0.004	

N	12
Mean	57.114
SD	92.3530
CV%	161.70

Aquakem v. 7.2AQ1

Results from time period:

Mon Apr 15 13:29:53 2019

Mon Apr 15 13:32:41 2019

AS

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	0.3284	µg/l	4/15/2019 8:02:58	
5.0PPBCN	A	Total CN	P	4.8036	µg/l	4/15/2019 8:02:59	
10PPBCN	A	Total CN	P	10.0187	µg/l	4/15/2019 8:03:00	
50PPBCN	A	Total CN	P	49.2268	µg/l	4/15/2019 8:03:01	
100PPBCN	A	Total CN	P	99.6245	µg/l	4/15/2019 8:03:02	
250PPBCN	A	Total CN	P	251.6878	µg/l	4/15/2019 8:03:03	
500PPBCN	A	Total CN	P	499.3101	µg/l	4/15/2019 8:03:04	
ICV1	S	Total CN	P	95.601	µg/l	4/15/2019 13:29:53	
ICB1	S	Total CN	P	0.3291	µg/l	4/15/2019 13:29:54	
CCV1	S	Total CN	P	239.0996	µg/l	4/15/2019 13:29:55	
CCB1	S	Total CN	P	-0.0142	µg/l	4/15/2019 13:29:56	
PB118888BL	S	Total CN	P	0.1592	µg/l	4/15/2019 13:29:57	
PB118888BS	S	Total CN	P	97.2307	µg/l	4/15/2019 13:29:58	
K2397-17MS	S	Total CN	P	5.3662	µg/l	4/15/2019 13:29:59	
K2397-17MSD	S	Total CN	P	5.1015	µg/l	4/15/2019 13:30:00	
K2397-17	S	Total CN	P	1.731	µg/l	4/15/2019 13:30:01	
K2397-17DUP	S	Total CN	P	1.5406	µg/l	4/15/2019 13:30:02	
CCV2	S	Total CN	P	239.0431	µg/l	4/15/2019 13:30:03	
CCB2	S	Total CN	P	0.18	µg/l	4/15/2019 13:32:41	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092

Reviewed by : AS

Instrument ID : Konelab

4/15/2019 13:04


Test Total CN

Accepted 4/15/2019 8:03

Factor 1068
Bias 0.004

Coeff. of det. 0.999979

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	0.3284	0.0000	
2	5.0PPBCN	0.008	4.8036	5.0000	
3	10PPBCN	0.013	10.0187	10.0000	
4	50PPBCN	0.050	49.2268	50.0000	
5	100PPBCN	0.097	99.6245	100.0000	
6	250PPBCN	0.239	251.6878	250.0000	
7	500PPBCN	0.471	499.3101	500.0000	