

LB102988

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

5/24/2019 13:54

Test: Total CN

ug/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.220	0.0	0.103	
ICB1	0.808	0.0	0.004	
CCV1	237.748	0.0	0.247	
CCB1	0.829	0.0	0.004	
PB120052BL	0.577	0.0	0.004	
PB120052BS	98.107	0.0	0.104	
MIDPB120052	239.773	0.0	0.249	
K2965-07	0.647	0.0	0.004	
K2980-07	0.571	0.0	0.004	
K2965-07DUP	0.286	0.0	0.004	
CCV2	251.336	0.0	0.261	
CCB2	0.615	0.0	0.004	

N 12
Mean 77.377
SD 106.3525
CV% 137.45

Aquakem v. 7.2AQ1

Results from time period:

Fri May 24 13:35:57 2019

Fri May 24 13:53:57 2019

AS

Sample Id	Sam/Ctr/c	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	0.4736	µg/l	5/24/2019 10:40:21	
5.0PPBCN	A	Total CN	P	5.3182	µg/l	5/24/2019 10:40:22	
10PPBCN	A	Total CN	P	10.6601	µg/l	5/24/2019 10:40:23	
50PPBCN	A	Total CN	P	50.5291	µg/l	5/24/2019 10:40:24	
100PPBCN	A	Total CN	P	100.2561	µg/l	5/24/2019 10:40:25	
250PPBCN	A	Total CN	P	245.7668	µg/l	5/24/2019 10:40:26	
500PPBCN	A	Total CN	P	501.9961	µg/l	5/24/2019 10:40:27	
ICV1	S	Total CN	P	97.2198	µg/l	5/24/2019 13:35:57	
ICB1	S	Total CN	P	0.808	µg/l	5/24/2019 13:35:58	
CCV1	S	Total CN	P	237.7479	µg/l	5/24/2019 13:35:59	
CCB1	S	Total CN	P	0.8289	µg/l	5/24/2019 13:36:00	
PB120052BL	S	Total CN	P	0.5771	µg/l	5/24/2019 13:36:01	
PB120052BS	S	Total CN	P	98.1075	µg/l	5/24/2019 13:36:02	
MIDPB120052	S	Total CN	P	239.7732	µg/l	5/24/2019 13:36:03	
K2965-07	S	Total CN	P	0.6474	µg/l	5/24/2019 13:36:04	
K2980-07	S	Total CN	P	0.5706	µg/l	5/24/2019 13:36:06	
K2965-07DUP	S	Total CN	P	0.2864	µg/l	5/24/2019 13:53:55	
CCV2	S	Total CN	P	251.3364	µg/l	5/24/2019 13:53:56	
CCB2	S	Total CN	P	0.6153	µg/l	5/24/2019 13:53:57	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

5/24/2019 13:53


Test Total CN

Accepted 5/24/2019 10:42

Factor 975.2
Bias 0.004

Coeff. of det. 0.999888

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	0.4736	0.0000	
2	5.0PPBCN	0.009	5.3182	5.0000	
3	10PPBCN	0.014	10.6601	10.0000	
4	50PPBCN	0.055	50.5291	50.0000	
5	100PPBCN	0.106	100.2561	100.0000	
6	250PPBCN	0.256	245.7668	250.0000	
7	500PPBCN	0.518	501.9961	500.0000	