

dB103053

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

5/29/2019 12:07

Test: *Reactive*
~~Total~~ CN

mg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	99.329	0.0	0.106	
ICB1	0.370	0.0	0.004	
CCV1	239.973	0.0	0.251	
CCB1	0.774	0.0	0.004	
PB120156BL	0.541	0.0	0.004	
K3074-06	0.584	0.0	0.004	
K3074-10	0.483	0.0	0.004	
K3054-08	0.322	0.0	0.004	
K3054-08DUP	0.633	0.0	0.004	
CCV2	253.930	0.0	0.265	
CCB2	0.622	0.0	0.004	

N 11
Mean 54.324
SD 99.7367
CV% 183.60

✓AS

Aquakem v. 7.2AQ1

Results from time period:

Wed May 29 11:15:55 2019

Wed May 29 12:06:16 2019

Sample Id	Sam/Ctr/c	Test short nam	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	0.5157	µg/l	5/29/2019 11:15:55	
5.OPPBCN	A	Reactive CN	P	5.1871	µg/l	5/29/2019 11:15:56	
10PPBCN	A	Reactive CN	P	10.6668	µg/l	5/29/2019 11:15:57	
50PPBCN	A	Reactive CN	P	49.3869	µg/l	5/29/2019 11:15:58	
100PPBCN	A	Reactive CN	P	100.5367	µg/l	5/29/2019 11:15:59	
250PPBCN	A	Reactive CN	P	247.5362	µg/l	5/29/2019 11:16:00	
500PPBCN	A	Reactive CN	P	501.1706	µg/l	5/29/2019 11:16:01	
ICV1	S	Reactive CN	P	99.3288	µg/l	5/29/2019 11:50:00	
ICB1	S	Reactive CN	P	0.37	µg/l	5/29/2019 11:50:01	
CCV1	S	Reactive CN	P	239.9726	µg/l	5/29/2019 11:50:02	
CCB1	S	Reactive CN	P	0.7742	µg/l	5/29/2019 11:50:03	
PB120156BL	S	Reactive CN	P	0.5409	µg/l	5/29/2019 11:50:04	
K3074-06	S	Reactive CN	P	0.5843	µg/l	5/29/2019 11:50:07	
K3074-10	S	Reactive CN	P	0.4832	µg/l	5/29/2019 11:50:08	
K3054-08	S	Reactive CN	P	0.3216	µg/l	5/29/2019 12:06:13	
K3054-08DUP	S	Reactive CN	P	0.6335	µg/l	5/29/2019 12:06:14	
CCV2	S	Reactive CN	P	253.9297	µg/l	5/29/2019 12:06:15	
CCB2	S	Reactive CN	P	0.6225	µg/l	5/29/2019 12:06:16	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

5/29/2019 11:26


Test Total CN

Accepted 5/29/2019 11:26

Factor 969.6
Bias 0.003

Coeff. of det. 0.999957

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	0.5157	0.0000	
2	5.0PPBCN	0.009	5.1871	5.0000	
3	10PPBCN	0.014	10.6668	10.0000	
4	50PPBCN	0.054	49.3869	50.0000	
5	100PPBCN	0.107	100.5367	100.0000	
6	250PPBCN	0.258	247.5362	250.0000	
7	500PPBCN	0.520	501.1706	500.0000	