

LB103328


7470A

INSTRUMENT ID : CV1

Linear

 μ Abs. :

42113



A= 0.0000e+000

B= 2.3626e-004

C= -3.5087e-002

Rho= 0.9998124

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.038	-0.038	-12	0.000	-12				
0.2	0.200	0.160	-0.040	826	0.0 %	826				
2.5	2.500	2.508	0.008	10763	0.0 %	10763				
5.0	5.000	5.142	0.142	21911	0.0 %	21911				
7.5	7.500	7.514	0.014	31952	0.0 %	31952				
10.0	10.000	9.915	-0.085	42113	0.0 %	42113				

LB103328 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 Jun 2019 09:52:42

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-D	-12	-	0.00007470A	PPB		11 Jun 2019 10:09:25	S	Std
0.2 - 1	50-2	826	-	0.20007470A	PPB		11 Jun 2019 10:11:39	S	Std
2.5 - 1	52-5	10763	-	2.50007470A	PPB		11 Jun 2019 10:13:47	S	Std
5.0 - 1	55-0	21911	-	5.00007470A	PPB		11 Jun 2019 10:15:56	S	Std
7.5 - 1	57-5	31952	-	7.50007470A	PPB		11 Jun 2019 10:18:04	S	Std
10.0 - 1	510-0	42113	-	10.00007470A	PPB		11 Jun 2019 10:20:11	S	Std
ICV85 - 1	ICV85	16094	3.7673	-7470A	PPB		11 Jun 2019 10:23:14	U	SMPL
ICB85 - 1	ICB85	-119	-0.0632	-7470A	PPB		11 Jun 2019 10:25:21	U	SMPL
CCV23 - 1	CCV23	22091	5.1841	-7470A	PPB		11 Jun 2019 10:27:27	U	SMPL
CCB23 - 1	CCB23	-50	-0.0469	-7470A	PPB		11 Jun 2019 10:29:33	U	SMPL
CRA - 1	CRA	879	0.1726	-7470A	PPB		11 Jun 2019 10:31:39	U	SMPL
HighStd - 1	HighStd	42891	10.0984	-7470A	PPB		11 Jun 2019 10:33:46	U	SMPL
ChkStd - 1	ChkStd	30117	7.0804	-7470A	PPB		11 Jun 2019 10:35:53	U	SMPL
PB120512BL - 1	PBW	-124	-0.0644	-7470A	PPB		11 Jun 2019 10:38:00	U	SMPL
PB120512BS - 1	LCSW	16659	3.9008	-7470A	PPB		11 Jun 2019 10:40:07	U	SMPL
K3193-01MS - 1	P-4-OWMS	25013	5.8745	-7470A	PPB		11 Jun 2019 10:50:20	U	SMPL
K3193-01MSD - 1	P-4-OWMSD	23649	5.5522	-7470A	PPB		11 Jun 2019 10:52:29	U	SMPL
K3193-02 - 1	P-1-OW	606	0.1081	-7470A	PPB		11 Jun 2019 10:54:38	U	SMPL
CCV24 - 1	CCV24	21285	4.9937	-7470A	PPB		11 Jun 2019 10:56:46	U	SMPL
CCB24 - 1	CCB24	-122	-0.0639	-7470A	PPB		11 Jun 2019 10:58:53	U	SMPL
K3249-02 - 1	BP-EB02-20190605	-117	-0.0627	-7470A	PPB		11 Jun 2019 11:01:00	U	SMPL
K3249-04 - 1	BPS1-TT-MW302S-20190606	-94	-0.0573	-7470A	PPB		11 Jun 2019 11:03:07	U	SMPL
K3249-05 - 1	BPS1-TT-MW302I1-20190606	-96	-0.0578	-7470A	PPB		11 Jun 2019 11:05:13	U	SMPL
K3249-06 - 1	BPS1-TT-MW305S-20190606	-87	-0.0556	-7470A	PPB		11 Jun 2019 11:07:20	U	SMPL
K3249-08 - 1	BPS1-TT-MW305I-20190606	-78	-0.0535	-7470A	PPB		11 Jun 2019 11:09:27	U	SMPL
K3249-09 - 1	BP-DUP03-20190606	-111	-0.0613	-7470A	PPB		11 Jun 2019 11:11:34	U	SMPL
K3249-11 - 1	BPS1-TT-MW305D-20190606	-91	-0.0566	-7470A	PPB		11 Jun 2019 11:13:42	U	SMPL
K3249-12 - 1	BP-HN-MW24S-20190606	-61	-0.0495	-7470A	PPB		11 Jun 2019 11:15:50	U	SMPL
K3249-14 - 1	BP-DUP04-20190606	-70	-0.0516	-7470A	PPB		11 Jun 2019 11:17:58	U	SMPL
K3249-16 - 1	BP-DUP05-20190606	-51	-0.0471	-7470A	PPB		11 Jun 2019 11:20:05	U	SMPL
CCV25 - 1	CCV25	21948	5.1504	-7470A	PPB		11 Jun 2019 11:22:11	U	SMPL
CCB25 - 1	CCB25	-82	-0.0545	-7470A	PPB		11 Jun 2019 11:24:18	U	SMPL
K3249-18 - 1	BP-HN-MW24IR-20190606	-72	-0.0521	-7470A	PPB		11 Jun 2019 11:26:25	U	SMPL
K3249-20 - 1	BPS1-TT-MW303S-20190606	-50	-0.0469	-7470A	PPB		11 Jun 2019 11:28:31	U	SMPL
PB120513BL - 1	PBW	-15	-0.0386	-7470A	PPB		11 Jun 2019 11:30:38	U	SMPL
PB120513BS - 1	LCSW	16641	3.8965	-7470A	PPB		11 Jun 2019 11:32:45	U	SMPL
K3190-03 - 1	BP-TT-MW201D-20190603-F	-86	-0.0554	-7470A	PPB		11 Jun 2019 11:34:52	U	SMPL
K3190-03DUP - 1	BP-TT-MW201D-20190603-FDUP	-12	-0.0379	-7470A	PPB		11 Jun 2019 11:36:59	U	SMPL
K3190-03MS - 1	BP-TT-MW201D-20190603-FMS	15667	3.6664	-7470A	PPB		11 Jun 2019 11:39:07	U	SMPL
K3190-03MSD - 1	BP-TT-MW201D-20190603-FMSD	14918	3.4894	-7470A	PPB		11 Jun 2019 11:41:15	U	SMPL
K3190-05 - 1	BP-TT-MW201D1-20190603-F	-109	-0.0608	-7470A	PPB		11 Jun 2019 11:43:23	U	SMPL
K3190-09 - 1	BP-TT-AOC22-MW04-20190603-F	-45	-0.0457	-7470A	PPB		11 Jun 2019 11:45:30	U	SMPL
CCV26 - 1	CCV26	21444	5.0313	-7470A	PPB		11 Jun 2019 11:47:36	U	SMPL
CCB26 - 1	CCB26	-91	-0.0566	-7470A	PPB		11 Jun 2019 11:49:43	U	SMPL
K3190-11 - 1	BP-DUP01-20190603-F	-33	-0.0429	-7470A	PPB		11 Jun 2019 11:51:49	U	SMPL
K3193-04 - 1	P-4-OW	139	-0.0022	-7470A	PPB		11 Jun 2019 11:53:56	U	SMPL
K3193-05 - 1	P-1-OW	22	-0.0299	-7470A	PPB		11 Jun 2019 11:56:02	U	SMPL
K3249-03 - 1	BP-EB02-20190605-F	-78	-0.0535	-7470A	PPB		11 Jun 2019 11:58:09	U	SMPL
K3249-07 - 1	BPS1-TT-MW305S-20190606-F	-64	-0.0502	-7470A	PPB		11 Jun 2019 12:00:16	U	SMPL
K3249-10 - 1	BP-DUP03-20190606-F	-60	-0.0493	-7470A	PPB		11 Jun 2019 12:02:23	U	SMPL
K3249-13 - 1	BP-HN-MW24S-20190606-F	-64	-0.0502	-7470A	PPB		11 Jun 2019 12:04:31	U	SMPL
K3249-15 - 1	BP-DUP04-20190606-F	-32	-0.0426	-7470A	PPB		11 Jun 2019 12:06:39	U	SMPL
K3249-17 - 1	BP-DUP05-20190606-F	-74	-0.0526	-7470A	PPB		11 Jun 2019 12:08:47	U	SMPL
K3249-19 - 1	BP-HN-MW24IR-20190606-F	-45	-0.0457	-7470A	PPB		11 Jun 2019 12:10:55	U	SMPL
CCV27 - 1	CCV27	21167	4.9658	-7470A	PPB		11 Jun 2019 12:13:01	U	SMPL
CCB27 - 1	CCB27	-114	-0.0620	-7470A	PPB		11 Jun 2019 12:15:09	U	SMPL
PB120514BL - 1	PBW	-26	-0.0412	-7470A	PPB		11 Jun 2019 12:17:17	U	SMPL
PB120514BS - 1	LCSW	16258	3.8060	-7470A	PPB		11 Jun 2019 12:19:24	U	SMPL
K3163-06 - 1	SU-04-060719-A	-73	-0.0523	-7470A	PPB		11 Jun 2019 12:21:31	U	SMPL
K3163-06DUP - 1	SU-04-060719-ADUP	-11	-0.0377	-7470A	PPB		11 Jun 2019 12:23:37	U	SMPL
K3163-06MS - 1	SU-04-060719-AMS	17083	4.0010	-7470A	PPB		11 Jun 2019 12:25:45	U	SMPL
K3163-06MSD - 1	SU-04-060719-AMSD	16833	3.9419	-7470A	PPB		11 Jun 2019 12:27:52	U	SMPL
K3163-08 - 1	SU-04-060719-B	-77	-0.0533	-7470A	PPB		11 Jun 2019 12:30:00	U	SMPL
K3262-01 - 1	STOCK-PILE-1	0	-0.0351	-7470A	PPB		11 Jun 2019 12:32:08	U	SMPL
PB120515BL - 1	PBW	-54	-0.0478	-7470A	PPB		11 Jun 2019 12:34:16	U	SMPL
PB120515BS - 1	LCSW	16599	3.8866	-7470A	PPB		11 Jun 2019 12:36:24	U	SMPL
CCV28 - 1	CCV28	21283	4.9932	-7470A	PPB		11 Jun 2019 12:38:31	U	SMPL
CCB28 - 1	CCB28	-113	-0.0618	-7470A	PPB		11 Jun 2019 12:40:38	U	SMPL
K3236-02 - 1	N48970	-19	-0.0396	-7470A	PPB		11 Jun 2019 12:42:45	U	SMPL
K3236-02DUP - 1	N48970DUP	-5	-0.0363	-7470A	PPB		11 Jun 2019 12:44:52	U	SMPL
K3236-02MS - 1	N48970MS	15553	3.6395	-7470A	PPB		11 Jun 2019 12:47:00	U	SMPL
K3236-02MSD - 1	N48970MSD	17350	4.0640	-7470A	PPB		11 Jun 2019 12:49:07	U	SMPL
K3264-06 - 1	WASTE-SLUDGE	-68	-0.0512	-7470A	PPB		11 Jun 2019 12:51:14	U	SMPL
PB120514TB - 1	PB120436TB	-14	-0.0384	-7470A	PPB		11 Jun 2019 12:53:22	U	SMPL
PB120515TB - 1	PB120438TB	5	-0.0339	-7470A	PPB		11 Jun 2019 12:55:29	U	SMPL
K3193-01A - 1		23783	5.5839	-7470A	PPB		11 Jun 2019 12:59:45	U	SMPL
K3190-03LX5 - 1		-473	-0.1468	-7470A	PPB		11 Jun 2019 13:01:52	U	SMPL
CCV29 - 1	CCV29	21253	4.9862	-7470A	PPB		11 Jun 2019 13:04:02	U	SMPL
CCB29 - 1	CCB29	-113	-0.0618	-7470A	PPB		11 Jun 2019 13:06:09	U	SMPL
K3190-03A - 1		15717	3.6782	-7470A	PPB		11 Jun 2019 13:08:17	U	SMPL
K3163-06LX5 - 1		-125	-0.0646	-7470A	PPB		11 Jun 2019 13:10:24	U	SMPL
K3163-06A - 1		16710	3.9128	-7470A	PPB		11 Jun 2019 13:12:32	U	SMPL
K3236-02LX5 - 1		-120	-0.0634	-7470A	PPB		11 Jun 2019 13:14:40	U	SMPL
K3236-02A - 1		15610	3.6529	-7470A	PPB		11 Jun 2019 13:16:49	U	SMPL
CCV30 - 1	CCV30	21964	5.1541	-7470A	PPB		11 Jun 2019 13:19:03	U	SMPL
CCB30 - 1	CCB30	-89	-0.0561	-7470A	PPB		11 Jun 2019 13:21:20	U	SMPL
K3193-01 - 1	P-4-OW	18252	4.2771	-7470A	PPB		11 Jun 2019 15:08:42	U	SMPL
K3193-01LX5 - 1		4353	0.9934	-7470A	PPB		11 Jun 2019 15:21:03	U	SMPL
K3193-01DUP - 1	P-4-OWDUP	17752	4.1590	-7470A	PPB		11 Jun 2019 15:26:44	U	SMPL
CCV31 - 1	CCV31	21073	4.9436	-7470A	PPB		11 Jun 2019 15:28:51	U	SMPL
CCB31 - 1	CCB31	-131	-0.0660	-7470A	PPB		11 Jun 2019 15:31:00	U	SMPL