

213103379

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

6/13/2019 14:55

Test: Reactive
~~Total~~ CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.635	0.0	0.105	
ICB1	-0.521	0.0	0.004	
CCV1	242.845	0.0	0.255	
CCB1	-0.695	0.0	0.004	
PB120569BL	-0.703	0.0	0.004	
K3156-04	-0.706	0.0	0.004	
K3156-04DUP	-1.024	0.0	0.003	
CCV2	245.181	0.0	0.257	
CCB2	-0.627	0.0	0.004	

N	9
Mean	64.598
SD	106.6927
CV%	165.16

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

6/13/2019 7:28


Test Ammonia-N

Accepted 6/13/2019 7:28

Factor 4.719
Bias 0.050

Coeff. of det. 0.999781

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	NH3-2PPM	0.073	0.1073	0.1000	
2	NH3-2PPM	0.092	0.1964	0.2000	
3	NH3-2PPM	0.135	0.4016	0.4000	
4	NH3-2PPM	0.258	0.9811	1.0000	
5	NH3-2PPM	0.336	1.3468	1.3333	
6	NH3-2PPM	0.474	2.0001	2.0000	

Aquakem v. 7.2AQ1

Results from time period:

Thu Jun 13 14:23:26 2019

Thu Jun 13 14:23:34 2019

✓AS

Sample Id	Sam/Ctr/c/	Test short nar	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Reactive CN	P	-0.6915	µg/l	6/13/2019 12:26:24	
5.0PPBCN	A	Reactive CN	P	4.0538	µg/l	6/13/2019 12:26:25	
10PPBCN	A	Reactive CN	P	9.1836	µg/l	6/13/2019 12:26:26	
50PPBCN	A	Reactive CN	P	50.0109	µg/l	6/13/2019 12:26:27	
100PPBCN	A	Reactive CN	P	101.0117	µg/l	6/13/2019 12:26:28	
250PPBCN	A	Reactive CN	P	253.2182	µg/l	6/13/2019 12:26:29	
500PPBCN	A	Reactive CN	P	498.2132	µg/l	6/13/2019 12:26:30	
ICV1	S	Reactive CN	P	97.6348	µg/l	6/13/2019 14:23:26	
ICB1	S	Reactive CN	P	-0.5207	µg/l	6/13/2019 14:23:27	
CCV1	S	Reactive CN	P	242.845	µg/l	6/13/2019 14:23:28	
CCB1	S	Reactive CN	P	-0.6947	µg/l	6/13/2019 14:23:29	
PB120569BL	S	Reactive CN	P	-0.7031	µg/l	6/13/2019 14:23:30	
K3156-04	S	Reactive CN	P	-0.7057	µg/l	6/13/2019 14:23:31	
K3156-04DUP	S	Reactive CN	P	-1.0242	µg/l	6/13/2019 14:23:32	
CCV2	S	Reactive CN	P	245.1812	µg/l	6/13/2019 14:23:33	
CCB2	S	Reactive CN	P	-0.6268	µg/l	6/13/2019 14:23:34	