

LB103785

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : At

Instrument ID : Konelab

7/3/2019 8:44

Test: Phenolics-

Sample Id	Result ^{mg/L}	Dil. 1 +	Response	Errors
ICV1	0.999	0.0	0.129	
ICB1	-0.004	0.0	0.015	
CCV1	1.014	0.0	0.131	
CCB1	0.001	0.0	0.016	
PB121087BL	-0.003	0.0	0.016	
PB121087BS	1.011	0.0	0.130	
K3589-01	-0.004	0.0	0.015	
K3589-01DUP	-0.008	0.0	0.015	
K3589-01MS	0.826	0.0	0.109	
K3589-01MSD	0.830	0.0	0.110	
DOC-01	1.003	0.0	0.129	
DOC-02	1.007	0.0	0.130	
DOC-03	1.001	0.0	0.129	
DOC-04	1.016	0.0	0.131	
CCV2	1.016	0.0	0.131	
CCB2	-0.006	0.0	0.015	

N	16
Mean	0.606
SD	0.4917
CV%	81.12

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

Instrument ID : Konelab

7/3/2019 8:03


Test Phenolics-

Accepted 7/3/2019 7:58

Factor 8.847
Bias 0.016

Coeff. of det. 0.999550

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.022	0.0541	0.0500	
2	Phenol-2	0.028	0.1034	0.1000	
3	Phenol-2	0.044	0.2519	0.2500	
4	Phenol-2	0.073	0.5090	0.5000	
5	Phenol-2	0.125	0.9687	1.0000	
6	Phenol-2	0.243	2.0129	2.0000	

LB103785

Aquakem v. 7.2AQ1

Results from time period:

Wed Jul 03 07:55:05 2019

Wed Jul 03 08:41:59 2019

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0541	mg/l	7/3/2019 7:55:05	
0.1PPM	A	Phenolics-	P	0.1034	mg/l	7/3/2019 7:55:06	
0.25PPM	A	Phenolics-	P	0.2519	mg/l	7/3/2019 7:55:07	
0.50PPM	A	Phenolics-	P	0.509	mg/l	7/3/2019 7:55:08	
1.0PPM	A	Phenolics-	P	0.9687	mg/l	7/3/2019 7:55:09	
2.0PPM	A	Phenolics-	P	2.0129	mg/l	7/3/2019 7:55:10	
ICV1	S	Phenolics-	P	0.9988	mg/l	7/3/2019 8:36:04	
ICB1	S	Phenolics-	P	-0.0041	mg/l	7/3/2019 8:36:05	
CCV1	S	Phenolics-	P	1.0138	mg/l	7/3/2019 8:36:06	
CCB1	S	Phenolics-	P	0.0015	mg/l	7/3/2019 8:36:07	
PB121087BL	S	Phenolics-	P	-0.0033	mg/l	7/3/2019 8:36:08	
PB121087BS	S	Phenolics-	P	1.0106	mg/l	7/3/2019 8:36:09	
K3589-01	S	Phenolics-	P	-0.0042	mg/l	7/3/2019 8:36:10	
K3589-01DUP	S	Phenolics-	P	-0.0079	mg/l	7/3/2019 8:36:11	
K3589-01MS	S	Phenolics-	P	0.8256	mg/l	7/3/2019 8:36:12	
K3589-01MSD	S	Phenolics-	P	0.8302	mg/l	7/3/2019 8:36:13	
DOC-01	S	Phenolics-	P	1.0034	mg/l	7/3/2019 8:36:14	
DOC-02	S	Phenolics-	P	1.0071	mg/l	7/3/2019 8:36:15	
DOC-03	S	Phenolics-	P	1.0009	mg/l	7/3/2019 8:41:54	
DOC-04	S	Phenolics-	P	1.0158	mg/l	7/3/2019 8:41:55	
CCV2	S	Phenolics-	P	1.0156	mg/l	7/3/2019 8:41:56	
CCB2	S	Phenolics-	P	-0.0062	mg/l	7/3/2019 8:41:57	