

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

7/10/2019 13:11

Test: Reactive
Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.603	0.0	0.095	
ICB1	0.707	0.0	0.001	
CCV1	245.187	0.0	0.239	
CCB1	0.561	0.0	0.001	
PB121265BL	0.342	0.0	0.001	
K3626-02	0.649	0.0	0.001	
K3626-02DUP	0.290	0.0	0.001	
K3626-04	0.637	0.0	0.001	
CCV2	247.994	0.0	0.241	
CCB2	0.810	0.0	0.002	

N	10
Mean	59.478
SD	103.1558
CV%	173.44

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AB

Instrument ID : Konelab

7/10/2019 13:13


Test Total CN

Accepted 7/10/2019 9:31

Factor 1031
Bias 0.001

Coeff. of det. 0.999982

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.001	0.5796	0.0000	
2	5.0PPBCN	0.005	4.7786	5.0000	
3	10PPBCN	0.010	10.0545	10.0000	
4	50PPBCN	0.049	50.0476	50.0000	
5	100PPBCN	0.098	100.4250	100.0000	
6	250PPBCN	0.242	248.4064	250.0000	
7	500PPBCN	0.487	500.7082	500.0000	

Aquakem v. 7.2AQ1

Results from time period:

Wed Jul 10 13:11:31 2019

Wed Jul 10 13:11:40 2019

AS

Sample Id	Sam/Ctr/c/	Test short nan	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	0.5796	µg/l	7/10/2019 9:23:47	
5.OPPBCN	A	Reactive CN	P	4.7786	µg/l	7/10/2019 9:23:48	
10PPBCN	A	Reactive CN	P	10.0545	µg/l	7/10/2019 9:23:49	
50PPBCN	A	Reactive CN	P	50.0476	µg/l	7/10/2019 9:23:50	
100PPBCN	A	Reactive CN	P	100.425	µg/l	7/10/2019 9:23:51	
250PPBCN	A	Reactive CN	P	248.4064	µg/l	7/10/2019 9:23:52	
500PPBCN	A	Reactive CN	P	500.7082	µg/l	7/10/2019 9:23:53	
ICV1	S	Reactive CN	P	97.6025	µg/l	7/10/2019 13:11:31	
ICB1	S	Reactive CN	P	0.7067	µg/l	7/10/2019 13:11:32	
CCV1	S	Reactive CN	P	245.1865	µg/l	7/10/2019 13:11:33	
CCB1	S	Reactive CN	P	0.5612	µg/l	7/10/2019 13:11:34	
PB121265BL	S	Reactive CN	P	0.3421	µg/l	7/10/2019 13:11:35	
K3626-02	S	Reactive CN	P	0.6487	µg/l	7/10/2019 13:11:36	
K3626-02DUP	S	Reactive CN	P	0.2904	µg/l	7/10/2019 13:11:37	
K3626-04	S	Reactive CN	P	0.6371	µg/l	7/10/2019 13:11:38	
CCV2	S	Reactive CN	P	247.994	µg/l	7/10/2019 13:11:39	
CCB2	S	Reactive CN	P	0.8099	µg/l	7/10/2019 13:11:40	