

Sample ID	Result	Std. Dev.	RSD	Mode	ALT
=====					
0.0PPM	15380			TOC	
0.5PPM	44911			TOC	
1.0PPM	75077			TOC	
2.0PPM.....	132535...	..	...	TOC	..
5.0PPM	307165			TOC	
10.0PPM	582330			TOC	
20.0PPM.....	1132911...	..	...	TOC	..
ICV	10.1993			TOC	
ICB	0.3351			TOC	
CCV.....	10.2763...	..	...	TOC	..
CCB	0.3381			TOC	
LB104631BLW	0.2956	0.0196	6.63	TOC	
LB104631BSW.....	10.6902...	0.1154..	1.08...	TOC	..
K4250-01X5	13.2146	0.1531	1.16	TOC	
K4250-02X5	5.0878	0.2800	5.50	TOC	
K4250-01DUPX5.....	13.2369...	0.1009..	0.76...	TOC	..
K4250-01MSX5	13.3729	0.2191	1.64	TOC	
K4250-01MSDX5	12.9716	0.0449	0.35	TOC	
CCV.....	10.9363...	..	...	TOC	..
CCB	0.3251			TOC	

Method ID	Sample Type	Vial	Timestamp	Message
=====	=====	=====	=====	=====
TOC 0 - 20 ppmC	TOC Standard	1	2019/08/15 14:20	
TOC 0 - 20 ppmC	TOC Standard	2	2019/08/15 14:33	
TOC 0 - 20 ppmC	TOC Standard	3	2019/08/15 14:46	
TOC 0 - 20 ppmC	...TOC Standard	.. 4..	2019/08/15 14:59	..
TOC 0 - 20 ppmC	TOC Standard	5	2019/08/15 15:12	
TOC 0 - 20 ppmC	TOC Standard	6	2019/08/15 15:26	
TOC 0 - 20 ppmC	...TOC Standard	.. 7..	2019/08/15 15:40	..
TOC 0 - 20 ppmC	Sample	8	2019/08/15 15:54	
TOC 0 - 20 ppmC	Sample	9	2019/08/15 16:07	
TOC 0 - 20 ppmC	...Sample	.. 10..	2019/08/16 08:49	..
TOC 0 - 20 ppmC	Sample	11	2019/08/16 09:02	
TOC 0 - 20 ppmC	Sample	12	2019/08/16 09:25	
TOC 0 - 20 ppmC	...Sample	.. 13..	2019/08/16 09:56	..
TOC 0 - 20 ppmC	Sample	14	2019/08/16 10:28	
TOC 0 - 20 ppmC	Sample	15	2019/08/16 10:54	
TOC 0 - 20 ppmC	...Sample	.. 16..	2019/08/16 11:25	..
TOC 0 - 20 ppmC	Sample	17	2019/08/16 11:56	
TOC 0 - 20 ppmC	Sample	18	2019/08/16 12:26	
TOC 0 - 20 ppmC	...Sample	.. 10..	2019/08/16 12:45	..
TOC 0 - 20 ppmC	Sample	11	2019/08/16 13:33	

```
=====
Sample ID:   0.0PPM                      Mode:      TOC
Method:      TOC 0 - 20 ppmC             Filename:   08151344
Cal. Curve:  TOC WATER 0-20PPM          Timestamp: 2019/08/15 14:20
Operator ID: AP AK                      Sample Type: TOC Standard
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1			15380	0.202	0.401	115

```
=====
```

```
Sample ID:   0.5PPM                      Mode:      TOC
Method:      TOC 0 - 20 ppmC             Filename:   08151344
Cal. Curve:  TOC WATER 0-20PPM          Timestamp: 2019/08/15 14:33
Operator ID: AP AK                      Sample Type: TOC Standard
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1			44911	0.209	0.406	127

```
=====
```

```
Sample ID:   1.0PPM                      Mode:      TOC
Method:      TOC 0 - 20 ppmC             Filename:   08151344
Cal. Curve:  TOC WATER 0-20PPM          Timestamp: 2019/08/15 14:46
Operator ID: AP AK                      Sample Type: TOC Standard
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1			75077	0.214	0.411	131

```
=====
```

```
Sample ID:   2.0PPM                      Mode:      TOC
Method:      TOC 0 - 20 ppmC             Filename:   08151344
Cal. Curve:  TOC WATER 0-20PPM          Timestamp: 2019/08/15 14:59
Operator ID: AP AK                      Sample Type: TOC Standard
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1			132535	0.180	0.378	139

```
=====
```

```
Sample ID:   5.0PPM                      Mode:      TOC
Method:      TOC 0 - 20 ppmC             Filename:   08151344
Cal. Curve:  TOC WATER 0-20PPM          Timestamp: 2019/08/15 15:12
Operator ID: AP AK                      Sample Type: TOC Standard
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1			307165	0.151	0.351	153

```
=====
```

```
Sample ID:   10.0PPM                     Mode:      TOC
Method:      TOC 0 - 20 ppmC             Filename:   08151344
Cal. Curve:  TOC WATER 0-20PPM          Timestamp: 2019/08/15 15:26
Operator ID: AP AK                      Sample Type: TOC Standard
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1			582330	0.154	0.352	173

```
=====
```

```
Sample ID:   20.0PPM                     Mode:      TOC
Method:      TOC 0 - 20 ppmC             Filename:   08151344
Cal. Curve:  TOC WATER 0-20PPM          Timestamp: 2019/08/15 15:40
Operator ID: AP AK                      Sample Type: TOC Standard
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1			1132911	0.137	0.337	201

Sample ID:	ICV	Mode:	TOC
Method:	TOC 0 - 20 ppmC	Filename:	08151344
Cal. Curve:	TOC WATER 0-20PPM	Timestamp:	2019/08/15 15:54
Operator ID:	AP AK	Sample Type:	Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	10.1993	5.0997	569474	0.045	0.245	160

Sample ID:	ICB	Mode:	TOC
Method:	TOC 0 - 20 ppmC	Filename:	08151344
Cal. Curve:	TOC WATER 0-20PPM	Timestamp:	2019/08/15 16:07
Operator ID:	AP AK	Sample Type:	Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	0.3351	0.1675	18708	0.014	0.214	115

Sample ID:	CCV	Mode:	TOC
Method:	TOC 0 - 20 ppmC	Filename:	08160813
Cal. Curve:	TOC WATER 0-20PPM	Timestamp:	2019/08/16 08:49
Operator ID:	AP AK	Sample Type:	Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	10.2763	5.1382	573772	0.102	0.299	163

Sample ID:	CCB	Mode:	TOC
Method:	TOC 0 - 20 ppmC	Filename:	08160813
Cal. Curve:	TOC WATER 0-20PPM	Timestamp:	2019/08/16 09:02
Operator ID:	AP AK	Sample Type:	Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	0.3381	0.1691	18879	0.109	0.306	117

Sample ID:	LB104631BLW	Mode:	TOC
Method:	TOC 0 - 20 ppmC	Filename:	08160813
Cal. Curve:	TOC WATER 0-20PPM	Timestamp:	2019/08/16 09:25
Operator ID:	AP AK	Sample Type:	Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	0.2677	0.1339	14947	0.052	0.251	116
2	0.2974	0.1487	16607	0.056	0.255	117
3	0.3050	0.1525	17029	0.086	0.284	115
4	0.3124	0.1562	17444	0.104	0.303	117

<<<Statistics>>> Mean: 0.2956 Std Dev: 0.0196 RSD: 6.63

Sample ID:	LB104631BSW	Mode:	TOC
Method:	TOC 0 - 20 ppmC	Filename:	08160931
Cal. Curve:	TOC WATER 0-20PPM	Timestamp:	2019/08/16 09:56
Operator ID:	AP AK	Sample Type:	Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	10.8075	5.4037	603430	0.077	0.277	176

2	10.5314	5.2657	588017	0.179	0.378	162
3	10.7178	5.3589	598422	0.197	0.395	164
4	10.7041	5.3521	597659	0.154	0.353	170
<<<Statistics>>> Mean: 10.6902 Std Dev: 0.1154 RSD: 1.08						
=====						
Sample ID: K4250-01X5				Mode: TOC		
Method: TOC 0 - 20 ppmC				Filename: 08160931		
Cal. Curve: TOC WATER 0-20PPM				Timestamp: 2019/08/16 10:28		
Operator ID: AP AK				Sample Type: Sample		
Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	13.0767	6.5384	730132	0.243	0.442	227
2	13.4325	6.7162	749994	0.249	0.448	249
3	13.1925	6.5962	736594	0.308	0.508	233
4	13.1565	6.5782	734584	0.315	0.514	230
<<<Statistics>>> Mean: 13.2146 Std Dev: 0.1531 RSD: 1.16						
=====						
Sample ID: K4250-02X5				Mode: TOC		
Method: TOC 0 - 20 ppmC				Filename: 08160931		
Cal. Curve: TOC WATER 0-20PPM				Timestamp: 2019/08/16 10:54		
Operator ID: AP AK				Sample Type: Sample		
Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	5.3925	2.6963	301089	0.166	0.366	165
2	5.2408	2.6204	292617	0.239	0.439	167
3	4.9429	2.4715	275985	0.323	0.520	152
4	4.7749	2.3874	266603	0.332	0.531	149
<<<Statistics>>> Mean: 5.0878 Std Dev: 0.2800 RSD: 5.50						
=====						
Sample ID: K4250-01DUPX5				Mode: TOC		
Method: TOC 0 - 20 ppmC				Filename: 08160931		
Cal. Curve: TOC WATER 0-20PPM				Timestamp: 2019/08/16 11:25		
Operator ID: AP AK				Sample Type: Sample		
Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	13.1446	6.5723	733920	0.268	0.468	237
2	13.3098	6.6549	743146	0.311	0.511	242
3	13.3375	6.6687	744691	0.351	0.549	239
4	13.1557	6.5778	734540	0.387	0.586	231
<<<Statistics>>> Mean: 13.2369 Std Dev: 0.1009 RSD: 0.76						
=====						
Sample ID: K4250-01MSX5				Mode: TOC		
Method: TOC 0 - 20 ppmC				Filename: 08160931		
Cal. Curve: TOC WATER 0-20PPM				Timestamp: 2019/08/16 11:56		
Operator ID: AP AK				Sample Type: Sample		
Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	13.0529	6.5264	728800	0.364	0.564	215
2	13.4154	6.7077	749041	0.268	0.467	224
3	13.4856	6.7428	752962	0.252	0.451	225
4	13.5377	6.7688	755868	0.260	0.459	226
<<<Statistics>>> Mean: 13.3729 Std Dev: 0.2191 RSD: 1.64						
=====						
Sample ID: K4250-01MSDX5				Mode: TOC		

```
Filename:      08160931
Timestamp:     2019/08/16 12:26
Sample Type:   Sample
```

```
<<<Statistics>>>      Mean:   12.9716      Std Dev:   0.0449      RSD: 0.35
```

```
Mode:          TOC
Filename:      08161232
Timestamp:     2019/08/16 12:45
Sample Type:   Sample
```

```
Mode:          TOC
Filename:      08161321
Timestamp:     2019/08/16 13:33
Sample Type:   Sample
```

Cal. Curve ID: TOC WATER 0-20PPM
Created: 2019/08/15 15:41
Calibration Factor (m): 1.117e+05
Y Intercept (b): 20098
r-squared: 0.99987

Standard ID	Y Raw Data	X Expected ug C	Measured ug C	Message	Date & Time
0.0PPM	15380	0.000	-0.042		2019/08/15 14:20
0.5PPM	44911	0.250	0.222		2019/08/15 14:33
1.0PPM	75077	0.500	0.492		2019/08/15 14:46
2.0PPM	132535	1.000	1.007		2019/08/15 14:59
5.0PPM	307165	2.500	2.571		2019/08/15 15:12
10.0PPM	582330	5.000	5.035		2019/08/15 15:26
20.0PPM	1132911	10.000	9.965		2019/08/15 15:40