

LB104928

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : CT

Instrument ID : Konelab

9/3/2019 9:46

Test: Phenolics- ug/L

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.990	0.0	0.124	
ICB1	0.002	0.0	0.016	
CCV1	0.987	0.0	0.123	
CCB1	0.001	0.0	0.016	
PB122712BL	-0.009	0.0	0.015	
PB122712BS	1.002	0.0	0.125	
K4555-23	0.001	0.0	0.016	
K4555-23DUP	0.005	0.0	0.017	
K4555-23MS	0.693	0.0	0.092	
K4555-23MSD	0.683	0.0	0.090	
CCV2	0.982	0.0	0.123	
CCB2	-0.000	0.0	0.016	

N	12
Mean	0.445
SD	0.4762
CV%	107.08

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : CT

Instrument ID : Konelab

9/3/2019 8:25


Test Phenolics-

Accepted 9/3/2019 8:25

Factor 9.191
Bias 0.016

Coeff. of det. 0.999995

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.022	0.0502	0.0500	
2	Phenol-2	0.027	0.1020	0.1000	
3	Phenol-2	0.043	0.2499	0.2500	
4	Phenol-2	0.070	0.4970	0.5000	
5	Phenol-2	0.125	1.0005	1.0000	
6	Phenol-2	0.234	2.0004	2.0000	

Aquakem v. 7.2AQ1

Results from time period:

Tue Sep 03 08:24:44 2019

Tue Sep 03 09:13:26 2019

CT
9/3/19

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and tin Stat
0.05PPM	A	Phenolics-	P	0.0502	mg/l	9/3/2019 8:24:44
0.1PPM	A	Phenolics-	P	0.102	mg/l	9/3/2019 8:24:45
0.25PPM	A	Phenolics-	P	0.2499	mg/l	9/3/2019 8:24:46
0.50PPM	A	Phenolics-	P	0.497	mg/l	9/3/2019 8:24:47
1.0PPM	A	Phenolics-	P	1.0005	mg/l	9/3/2019 8:24:48
2.0PPM	A	Phenolics-	P	2.0004	mg/l	9/3/2019 8:24:49
ICV1	S	Phenolics-	P	0.9903	mg/l	9/3/2019 8:52:10
ICB1	S	Phenolics-	P	0.0021	mg/l	9/3/2019 8:52:11
CCV1	S	Phenolics-	P	0.9866	mg/l	9/3/2019 8:52:12
CCB1	S	Phenolics-	P	0.0008	mg/l	9/3/2019 8:52:13
PB122712BL	S	Phenolics-	P	-0.0089	mg/l	9/3/2019 8:52:14
PB122712BS	S	Phenolics-	P	1.0023	mg/l	9/3/2019 8:52:15
K4555-23	S	Phenolics-	P	0.0012	mg/l	9/3/2019 8:52:16
K4555-23DUP	S	Phenolics-	P	0.0054	mg/l	9/3/2019 8:52:17
K4555-23MS	S	Phenolics-	P	0.693	mg/l	9/3/2019 9:13:23
K4555-23MSD	S	Phenolics-	P	0.683	mg/l	9/3/2019 9:13:24
CCV2	S	Phenolics-	P	0.9815	mg/l	9/3/2019 9:13:25
CCB2	S	Phenolics-	P	0	mg/l	9/3/2019 9:13:26