

LB105600

Calibration results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : 

Instrument ID : Konelab

10/16/2019 15:55


Test Phenolics-

Accepted 10/16/2019 15:55

Factor 9.02
Bias 0.011

Coeff. of det. 0.999846

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.017	0.0541	0.0500	
2	Phenol-2	0.023	0.1042	0.1000	
3	Phenol-2	0.039	0.2516	0.2500	
4	Phenol-2	0.067	0.4994	0.5000	
5	Phenol-2	0.120	0.9821	1.0000	
6	Phenol-2	0.234	2.0086	2.0000	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : CT

Instrument ID : Konelab

10/16/2019 17:13

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.970 ✓	0.0	0.119	
ICB1	0.002 ✓	0.0	0.012	
CCV1	0.972 ✓	0.0	0.119	
CCB1	0.002 ✓	0.0	0.012	
PB123758BL	-0.000	0.0	0.011	
PB123758BS	0.971 ✓	0.0	0.119	
K5111-10	0.008	0.0	0.012	
K5111-11	0.046	0.0	0.016	
CCV2	0.974 ✓	0.0	0.119	
CCB2	0.002 ✓	0.0	0.012	

N	10
Mean	0.395
SD	0.4969
CV%	125.93

CT
10/16/19

Aquakem v. 7.2AQ1

Results from time period:

Wed Oct 16 15:30:15 2019

Wed Oct 16 17:10:18 2019

Sample Id	Sam/Ctr/c/	Test short nam	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0541	mg/l	10/16/2019 15:53:42	
0.1PPM	A	Phenolics-	P	0.1042	mg/l	10/16/2019 15:53:43	
0.25PPM	A	Phenolics-	P	0.2516	mg/l	10/16/2019 15:53:44	
0.50PPM	A	Phenolics-	P	0.4994	mg/l	10/16/2019 15:53:45	
1.0PPM	A	Phenolics-	P	0.9821	mg/l	10/16/2019 15:53:46	
2.0PPM	A	Phenolics-	P	2.0086	mg/l	10/16/2019 15:53:47	
ICV1	S	Phenolics-	P	0.9704	mg/l	10/16/2019 16:20:13	
ICB1	S	Phenolics-	P	0.0018	mg/l	10/16/2019 16:20:15	
CCV1	S	Phenolics-	P	0.9719	mg/l	10/16/2019 16:20:17	
CCB1	S	Phenolics-	P	0.0022	mg/l	10/16/2019 16:20:19	
PB123758BL	S	Phenolics-	P	-0.0003	mg/l	10/16/2019 16:20:20	
PB123758BS	S	Phenolics-	P	0.9712	mg/l	10/16/2019 16:20:21	
K5111-10	S	Phenolics-	P	0.0079	mg/l	10/16/2019 17:10:09	
K5111-11	S	Phenolics-	P	0.0457	mg/l	10/16/2019 17:10:11	
CCV2	S	Phenolics-	P	0.9736	mg/l	10/16/2019 17:10:17	
CCB2	S	Phenolics-	P	0.0018	mg/l	10/16/2019 17:10:18	