


7471B

LB106350
INSTRUMENT ID CV1

Linear

 μ Abs.:

50254



A= 0.0000e+000

B= 1.9664e-004

C= -1.3788e-002

Rho= 0.9989180

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.026	0.026	204	0.000	204				
0.2	0.200	0.206	0.006	1117	0.0 %	1117				
2.5	2.500	2.548	0.048	13026	0.0 %	13026				
5.0	5.000	4.748	-0.252	24213	0.0 %	24213				
7.5	7.500	7.804	0.304	39758	0.0 %	39758				
10.0	10.000	9.868	-0.132	50254	0.0 %	50254				

LB106350 INSTRUMENT ID CV1

Method: 7471B		Operator: Admin		Date of Analysis: 18 Nov 2019 08:42:20					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1		204	-	0.0000	7471B	PPB	18 Nov 2019 09:00:07	S	Std
0.2 - 1		1117	-	0.2000	7471B	PPB	18 Nov 2019 09:02:22	S	Std
2.5 - 1		13026	-	2.5000	7471B	PPB	18 Nov 2019 09:04:37	S	Std
5.0 - 1		24213	-	5.0000	7471B	PPB	18 Nov 2019 09:06:53	S	Std
7.5 - 1		39758	-	7.5000	7471B	PPB	18 Nov 2019 09:09:10	S	Std
10.0 - 1		50254	-	10.0000	7471B	PPB	18 Nov 2019 09:11:26	S	Std
ICV09 - 1	ICV09	18914	3.7055	-	7471B	PPB	18 Nov 2019 09:15:14	U	SMPL
ICB09 - 1	ICB09	131	0.0120	-	7471B	PPB	18 Nov 2019 09:17:30	U	SMPL
CCV70 - 1	CCV70	24514	4.8067	-	7471B	PPB	18 Nov 2019 09:19:45	U	SMPL
CCB70 - 1	CCB70	176	0.0208	-	7471B	PPB	18 Nov 2019 09:22:00	U	SMPL
CRA - 1	CRA	1119	0.2063	-	7471B	PPB	18 Nov 2019 09:24:16	U	SMPL
HighStd - 1	HighStd	50233	9.8642	-	7471B	PPB	18 Nov 2019 09:26:32	U	SMPL
ChkStd - 1	ChkStd	36619	7.1871	-	7471B	PPB	18 Nov 2019 09:28:47	U	SMPL
PB124889BL - 1	PBS	-5	-0.0148	-	7471B	PPB	18 Nov 2019 09:31:05	U	SMPL
PB124889BS - 1	LCSS	20580	4.0331	-	7471B	PPB	18 Nov 2019 09:33:23	U	SMPL
K5669-01 - 1	CL-02-111519	1813	0.3427	-	7471B	PPB	18 Nov 2019 09:35:39	U	SMPL
K5669-01DUP - 1	CL-02-111519DUP	1952	0.3701	-	7471B	PPB	18 Nov 2019 09:37:56	U	SMPL
K5669-01MS - 1	CL-02-111519MS	23018	4.5125	-	7471B	PPB	18 Nov 2019 09:40:13	U	SMPL
K5669-01MSD - 1	CL-02-111519MSD	22716	4.4531	-	7471B	PPB	18 Nov 2019 09:42:31	U	SMPL
K5676-16 - 1	SU-01-111519	2426	0.4633	-	7471B	PPB	18 Nov 2019 09:44:47	U	SMPL
CCV71 - 1	CCV71	24774	4.8578	-	7471B	PPB	18 Nov 2019 09:47:03	U	SMPL
CCB71 - 1	CCB71	27	-0.0085	-	7471B	PPB	18 Nov 2019 09:49:19	U	SMPL
K5792-03 - 1	3	24617	4.8270	-	7471B	PPB	18 Nov 2019 09:51:35	U	SMPL
K5864-02 - 1	SB-01-COMP	5019	0.9732	-	7471B	PPB	18 Nov 2019 09:53:50	U	SMPL
K5864-04 - 1	SB-02-COMP	2244	0.4275	-	7471B	PPB	18 Nov 2019 09:56:08	U	SMPL
K5864-06 - 1	SB-03-COMP	6958	1.3545	-	7471B	PPB	18 Nov 2019 09:58:24	U	SMPL
K5864-08 - 1	SB-04-COMP	7799	1.5198	-	7471B	PPB	18 Nov 2019 10:00:40	U	SMPL
K5864-10 - 1	SB-05-COMP	1045	0.1917	-	7471B	PPB	18 Nov 2019 10:02:57	U	SMPL
K5865-23 - 1	2-(16-19)	111	0.0080	-	7471B	PPB	18 Nov 2019 10:05:13	U	SMPL
K5876-01 - 1	1-7TH-FL-GYM-NE-CORNER	3155	0.6066	-	7471B	PPB	18 Nov 2019 10:07:30	U	SMPL
K5876-02 - 1	2-7TH-FL-GYM-SE-CORNER	2768	0.5305	-	7471B	PPB	18 Nov 2019 10:09:48	U	SMPL
K5905-01 - 1	RT-3037	2580	0.4935	-	7471B	PPB	18 Nov 2019 10:12:03	U	SMPL
CCV72 - 1	CCV72	25667	5.0334	-	7471B	PPB	18 Nov 2019 10:14:18	U	SMPL
CCB72 - 1	CCB72	49	-0.0042	-	7471B	PPB	18 Nov 2019 10:16:34	U	SMPL
K5906-01 - 1	NA-SP	2484	0.4747	-	7471B	PPB	18 Nov 2019 10:18:50	U	SMPL
K5908-01 - 1	B-8D-(12-18)	6897	1.3425	-	7471B	PPB	18 Nov 2019 10:21:06	U	SMPL
K5908-02 - 1	B-8-1-(6-12)	14613	2.8597	-	7471B	PPB	18 Nov 2019 10:23:21	U	SMPL
K5908-03 - 1	B-8-2-(6-12)	9288	1.8126	-	7471B	PPB	18 Nov 2019 10:25:37	U	SMPL
K5908-04 - 1	B-8-3-(6-12)	6279	1.2209	-	7471B	PPB	18 Nov 2019 10:27:53	U	SMPL
PB124890BL - 1	PBS	107	0.0073	-	7471B	PPB	18 Nov 2019 10:30:09	U	SMPL
PB124890BS - 1	LCSS	21004	4.1165	-	7471B	PPB	18 Nov 2019 10:32:27	U	SMPL
K5865-01 - 1	4-(2-4)	3151	0.6058	-	7471B	PPB	18 Nov 2019 10:34:43	U	SMPL
K5865-01DUP - 1	4-(2-4)DUP	2944	0.5651	-	7471B	PPB	18 Nov 2019 10:37:01	U	SMPL
K5865-01MS - 1	4-(2-4)MS	23260	4.5601	-	7471B	PPB	18 Nov 2019 10:39:16	U	SMPL
CCV73 - 1	CCV73	27659	5.4251	-	7471B	PPB	18 Nov 2019 10:41:32	U	SMPL
CCB73 - 1	CCB73	1	-0.0136	-	7471B	PPB	18 Nov 2019 10:43:48	U	SMPL
K5865-01MSD - 1	4-(2-4)MSD	20453	4.0081	-	7471B	PPB	18 Nov 2019 10:46:04	U	SMPL
K5865-02 - 1	4-(4-6)	64	-0.0012	-	7471B	PPB	18 Nov 2019 10:48:20	U	SMPL
K5865-03 - 1	4-(6-8)	94	0.0047	-	7471B	PPB	18 Nov 2019 10:50:36	U	SMPL
K5865-04 - 1	4-(8-10)	127	0.0112	-	7471B	PPB	18 Nov 2019 10:52:52	U	SMPL
K5865-05 - 1	4-(10-12)	152	0.0161	-	7471B	PPB	18 Nov 2019 10:55:09	U	SMPL
K5865-06 - 1	4-(12-14)	85	0.0029	-	7471B	PPB	18 Nov 2019 10:57:25	U	SMPL
K5865-07 - 1	4-(14-17)	126	0.0110	-	7471B	PPB	18 Nov 2019 10:59:42	U	SMPL
K5865-08 - 1	3-(2-4)	3987	0.7702	-	7471B	PPB	18 Nov 2019 11:01:59	U	SMPL
K5865-09 - 1	3-(4-6)	-15	-0.0167	-	7471B	PPB	18 Nov 2019 11:04:16	U	SMPL
K5865-10 - 1	3-(6-8)	4493	0.8697	-	7471B	PPB	18 Nov 2019 11:06:31	U	SMPL
CCV74 - 1	CCV74	27072	5.3097	-	7471B	PPB	18 Nov 2019 11:08:47	U	SMPL
CCB74 - 1	CCB74	116	0.0090	-	7471B	PPB	18 Nov 2019 11:11:03	U	SMPL
K5865-11 - 1	3-(8-10)	3935	0.7600	-	7471B	PPB	18 Nov 2019 11:13:20	U	SMPL
K5865-12 - 1	3-(10-12)	3082	0.5923	-	7471B	PPB	18 Nov 2019 11:15:36	U	SMPL
K5865-13 - 1	3-(12-14)	225	0.0305	-	7471B	PPB	18 Nov 2019 11:17:52	U	SMPL
K5865-14 - 1	3-(14-16)	4606	0.8919	-	7471B	PPB	18 Nov 2019 11:20:08	U	SMPL
K5865-15 - 1	3-(16-18)	564	0.0971	-	7471B	PPB	18 Nov 2019 11:22:24	U	SMPL
K5865-16 - 1	2-(2-4)	3085	0.5929	-	7471B	PPB	18 Nov 2019 11:24:41	U	SMPL
K5865-17 - 1	2-(4-6)	341	0.0533	-	7471B	PPB	18 Nov 2019 11:26:58	U	SMPL
K5865-18 - 1	2-(6-8)	132	0.0122	-	7471B	PPB	18 Nov 2019 11:29:15	U	SMPL
K5865-19 - 1	2-(8-10)	112	0.0082	-	7471B	PPB	18 Nov 2019 11:31:32	U	SMPL
K5865-20 - 1	2-(10-12)	169	0.0194	-	7471B	PPB	18 Nov 2019 11:33:48	U	SMPL
CCV75 - 1	CCV75	24545	4.8128	-	7471B	PPB	18 Nov 2019 11:36:04	U	SMPL
CCB75 - 1	CCB75	126	0.0110	-	7471B	PPB	18 Nov 2019 11:38:21	U	SMPL
K5865-21 - 1	2-(12-14)	200	0.0255	-	7471B	PPB	18 Nov 2019 11:40:37	U	SMPL
K5865-22 - 1	2-(14-16)	248	0.0350	-	7471B	PPB	18 Nov 2019 11:42:53	U	SMPL
PB124891BL - 1	PBS	147	0.0151	-	7471B	PPB	18 Nov 2019 11:45:09	U	SMPL
PB124891BS - 1	LCSS	18790	3.6811	-	7471B	PPB	18 Nov 2019 11:47:25	U	SMPL
K5904-01 - 1	1-(2-4)	178	0.0212	-	7471B	PPB	18 Nov 2019 11:49:41	U	SMPL
K5904-01DUP - 1	1-(2-4)DUP	224	0.0303	-	7471B	PPB	18 Nov 2019 11:51:58	U	SMPL
K5904-01MS - 1	1-(2-4)MS	19241	3.7698	-	7471B	PPB	18 Nov 2019 11:54:14	U	SMPL
K5904-01MSD - 1	1-(2-4)MSD	18227	3.5704	-	7471B	PPB	18 Nov 2019 11:56:31	U	SMPL
K5904-02 - 1	1-(4-6)	54	-0.0032	-	7471B	PPB	18 Nov 2019 11:58:48	U	SMPL
K5904-03 - 1	1-(6-8)	77	0.0014	-	7471B	PPB	18 Nov 2019 12:01:05	U	SMPL
CCV76 - 1	CCV76	27817	5.4562	-	7471B	PPB	18 Nov 2019 12:03:21	U	SMPL
CCB76 - 1	CCB76	53	-0.0034	-	7471B	PPB	18 Nov 2019 12:05:38	U	SMPL
K5904-04 - 1	1-(8-10)	189	0.0234	-	7471B	PPB	18 Nov 2019 12:07:55	U	SMPL
K5904-05 - 1	1-(10-12)	121	0.0100	-	7471B	PPB	18 Nov 2019 12:10:12	U	SMPL
K5904-06 - 1	1-(12-14)	89	0.0037	-	7471B	PPB	18 Nov 2019 12:12:28	U	SMPL
K5904-07 - 1	1-(14-16)	123	0.0104	-	7471B	PPB	18 Nov 2019 12:14:44	U	SMPL
K5904-08 - 1	1-(16-18)	248	0.0350	-	7471B	PPB	18 Nov 2019 12:17:01	U	SMPL
K5904-09 - 1	1-(18-20)	163	0.0183	-	7471B	PPB	18 Nov 2019 12:19:17	U	SMPL
K5904-10 - 1	11-(0-2)	670	0.1180	-	7471B	PPB	18 Nov 2019 12:21:34	U	SMPL
K5904-11 - 1	11-(2-4)	5078	0.9848	-	7471B	PPB	18 Nov 2019 12:23:51	U	SMPL
K5904-12 - 1	11-(4-7)	4978	0.9651	-	7471B	PPB	18 Nov 2019 12:26:08	U	SMPL
K5904-13 - 1	12-(0-2)	1516	0.2843	-	7471B	PPB	18 Nov 2019 12:28:24	U	SMPL
CCV77 - 1	CCV77	27681	5.4295	-	7471B	PPB	18 Nov 2019 12:30:41	U	SMPL

LB106350 INSTRUMENT ID CV1

Method: 7471B		Operator: Admin		Date of Analysis: 18 Nov 2019 08:42:20				
Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
CCB77 - 1	CCB77	-15	-0.0167	-7471B	PPB	18 Nov 2019 12:32:57	U	SMPL
K5904-14 - 1	12-(2-4)	2379	0.4540	-7471B	PPB	18 Nov 2019 12:35:15	U	SMPL
K5904-15 - 1	12-(4-6)	15953	3.1233	-7471B	PPB	18 Nov 2019 12:37:31	U	SMPL
K5904-16 - 1	13-(0-2)	87102	17.1142	-7471B	PPB	18 Nov 2019 12:39:48	U	SMPL
K5904-17 - 1	13-(2-5)	19774	3.8746	-7471B	PPB	18 Nov 2019 12:42:04	U	SMPL
K5904-18 - 1	14-(0-2)	8576	1.6726	-7471B	PPB	18 Nov 2019 12:44:23	U	SMPL
K5904-19 - 1	14-(2-4)	2641	0.5055	-7471B	PPB	18 Nov 2019 12:46:39	U	SMPL
PB124892BL - 1	PBS	11	-0.0116	-7471B	PPB	18 Nov 2019 12:48:56	U	SMPL
PB124892BS - 1	LCSS	20637	4.0443	-7471B	PPB	18 Nov 2019 12:51:13	U	SMPL
K5665-01 - 1	HR-05-111419	3183	0.6121	-7471B	PPB	18 Nov 2019 12:53:31	U	SMPL
K5665-01DUP - 1	HR-05-111419DUP	3146	0.6048	-7471B	PPB	18 Nov 2019 12:55:47	U	SMPL
CCV78 - 1	CCV78	25400	4.9809	-7471B	PPB	18 Nov 2019 13:05:11	U	SMPL
CCB78 - 1	CCB78	-5	-0.0148	-7471B	PPB	18 Nov 2019 13:07:27	U	SMPL
K5665-01MS - 1	HR-05-111419MS	20663	4.0494	-7471B	PPB	18 Nov 2019 13:14:26	U	SMPL
K5665-01MSD - 1	HR-05-111419MSD	20602	4.0374	-7471B	PPB	18 Nov 2019 13:16:42	U	SMPL
K5671-06 - 1	OR-03-111419	2636	0.5046	-7471B	PPB	18 Nov 2019 13:18:59	U	SMPL
K5838-01 - 1	UC-TP	40820	8.0132	-7471B	PPB	18 Nov 2019 13:21:16	U	SMPL
K5839-01 - 1	PHIPPS-BH-NW-01	-137	-0.0407	-7471B	PPB	18 Nov 2019 13:23:33	U	SMPL
K5839-03 - 1	PHIPPS-BH-NW-02	276035	54.2664	-7471B	PPB	18 Nov 2019 13:25:49	U	SMPL
K5839-05 - 1	PHIPPS-BH-NW-03	-265	-0.0659	-7471B	PPB	18 Nov 2019 13:28:05	U	SMPL
K5839-07 - 1	PHIPPS-BH-SW-01	5729	1.1128	-7471B	PPB	18 Nov 2019 13:30:30	U	SMPL
K5839-09 - 1	PHIPPS-BH-SW-02	5965	1.1592	-7471B	PPB	18 Nov 2019 13:32:47	U	SMPL
K5840-01 - 1	PHIPPS-BH-SW-03	534	0.0912	-7471B	PPB	18 Nov 2019 13:35:04	U	SMPL
CCV79 - 1	CCV79	25254	4.9522	-7471B	PPB	18 Nov 2019 13:37:21	U	SMPL
CCB79 - 1	CCB79	166	0.0189	-7471B	PPB	18 Nov 2019 13:39:37	U	SMPL
K5840-03 - 1	PHIPPS-BH-NE-01	31679	6.2157	-7471B	PPB	18 Nov 2019 13:41:54	U	SMPL
K5840-05 - 1	PHIPPS-BH-NE-02	3733	0.7203	-7471B	PPB	18 Nov 2019 13:44:11	U	SMPL
K5840-07 - 1	PHIPPS-BH-NE-03	2839	0.5445	-7471B	PPB	18 Nov 2019 13:46:29	U	SMPL
K5840-09 - 1	PHIPPS-BH-SE-01	12350	2.4147	-7471B	PPB	18 Nov 2019 13:48:46	U	SMPL
K5840-11 - 1	PHIPPS-BH-SE-02	2163	0.4115	-7471B	PPB	18 Nov 2019 13:51:02	U	SMPL
K5840-13 - 1	PHIPPS-BH-SE-03	245	0.0344	-7471B	PPB	18 Nov 2019 13:53:19	U	SMPL
K5862-03 - 1	TP-13A-S	1320	0.2458	-7471B	PPB	18 Nov 2019 13:55:36	U	SMPL
K5863-05 - 1	TP-1-COMP	6006	1.1672	-7471B	PPB	18 Nov 2019 13:57:53	U	SMPL
K5669-01LX5 - 1		321	0.0493	-7471B	PPB	18 Nov 2019 14:00:10	U	SMPL
K5865-01LX5 - 1		553	0.0950	-7471B	PPB	18 Nov 2019 14:02:26	U	SMPL
CCV80 - 1	CCV80	25710	5.0419	-7471B	PPB	18 Nov 2019 14:04:44	U	SMPL
CCB80 - 1	CCB80	214	0.0283	-7471B	PPB	18 Nov 2019 14:07:01	U	SMPL
K5904-01LX5 - 1		74	0.0008	-7471B	PPB	18 Nov 2019 14:09:18	U	SMPL
K5665-01LX5 - 1		660	0.1160	-7471B	PPB	18 Nov 2019 14:11:34	U	SMPL
K5904-16DLX3 - 1		29893	5.8644	-7471B	PPB	18 Nov 2019 14:13:51	U	SMPL
K5839-03DLX10 - 1		30641	6.0115	-7471B	PPB	18 Nov 2019 14:28:10	U	SMPL
CCV81 - 1	CCV81	25993	5.0975	-7471B	PPB	18 Nov 2019 14:30:27	U	SMPL
CCB81 - 1	CCB81	206	0.0267	-7471B	PPB	18 Nov 2019 14:32:46	U	SMPL