

LB106466

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AL

Instrument ID : Konelab

11/25/2019 16:12

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	94.767	0.0	0.099	
ICB1	0.542	0.0	0.002	
CCV1	238.661	0.0	0.247	
CCB1	0.287	0.0	0.002	
PB125084BL	0.181	0.0	0.002	
PB125084BS	95.944	0.0	0.101	
LOWPB125084	10.963	0.0	0.013	
HIGHPB125084	487.045	0.0	0.503	
K5663-01	1.389	0.0	0.003	
K5663-01DUP	0.769	0.0	0.003	
K5663-01MS	41.259	0.0	0.044	
K5663-01MSD	36.781	0.0	0.040	
K5672-01	0.789	0.0	0.003	
K5676-31	1.452	0.0	0.003	
CCV2	241.702	0.0	0.251	
CCB2	0.176	0.0	0.002	
K5977-02	7.518	0.0	0.010	
K5977-03	0.386	0.0	0.002	
K5977-06	0.890	0.0	0.003	
K5977-07	0.381	0.0	0.002	
K5977-10	2.795	0.0	0.005	
K5977-11	1.022	0.0	0.003	
K6001-03	0.220	0.0	0.002	
PB125067BL	0.326	0.0	0.002	
PB125067BS	95.841	0.0	0.100	
K5959-13	0.747	0.0	0.003	
CCV3	240.211	0.0	0.249	
CCB3	0.082	0.0	0.002	
K5959-13DUP	0.554	0.0	0.002	
K5959-13MS	35.819	0.0	0.039	
K5959-13MSD	33.454	0.0	0.036	
CCV4	240.327	0.0	0.249	
CCB4	0.365	0.0	0.002	

N	33
Mean	57.989
SD	110.4114
CV%	190.40

Aquakem v. 7.2AQ1

Results from time period:

Mon Nov 25 13:38:34 2019

Mon Nov 25 15:17:06 2019

AO

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	-0.0214	µg/l	11/25/2019 14:03:39	
5.OPPBCN	A	Total CN	P	5.2843	µg/l	11/25/2019 14:03:40	
10PPBCN	A	Total CN	P	10.5848	µg/l	11/25/2019 14:03:41	
50PPBCN	A	Total CN	P	50.4141	µg/l	11/25/2019 14:03:42	
100PPBCN	A	Total CN	P	99.4613	µg/l	11/25/2019 14:03:43	
250PPBCN	A	Total CN	P	248.4503	µg/l	11/25/2019 14:03:44	
500PPBCN	A	Total CN	P	500.8266	µg/l	11/25/2019 14:03:45	
ICV1	S	Total CN	P	94.7671	µg/l	11/25/2019 14:57:37	
ICB1	S	Total CN	P	0.5423	µg/l	11/25/2019 14:57:38	
CCV1	S	Total CN	P	238.6606	µg/l	11/25/2019 14:57:39	
CCB1	S	Total CN	P	0.2867	µg/l	11/25/2019 14:57:40	
PB125084BL	S	Total CN	P	0.181	µg/l	11/25/2019 14:57:41	
PB125084BS	S	Total CN	P	95.9442	µg/l	11/25/2019 14:57:42	
LOWPB125084	S	Total CN	P	10.9628	µg/l	11/25/2019 14:57:43	
HIGHPB125084	S	Total CN	P	487.0448	µg/l	11/25/2019 14:57:44	
K5663-01	S	Total CN	P	1.3888	µg/l	11/25/2019 14:57:45	
K5663-01DUP	S	Total CN	P	0.7689	µg/l	11/25/2019 14:57:46	
K5663-01MS	S	Total CN	P	41.2587	µg/l	11/25/2019 14:57:47	
K5663-01MSD	S	Total CN	P	36.7806	µg/l	11/25/2019 15:05:11	
K5672-01	S	Total CN	P	0.7892	µg/l	11/25/2019 15:05:12	
K5676-31	S	Total CN	P	1.4518	µg/l	11/25/2019 15:05:13	
CCV2	S	Total CN	P	241.7024	µg/l	11/25/2019 15:05:14	
CCB2	S	Total CN	P	0.1764	µg/l	11/25/2019 15:05:15	
K5977-02	S	Total CN	P	7.5176	µg/l	11/25/2019 15:05:16	
K5977-03	S	Total CN	P	0.3857	µg/l	11/25/2019 15:05:17	
K5977-06	S	Total CN	P	0.8896	µg/l	11/25/2019 15:05:18	
K5977-07	S	Total CN	P	0.3813	µg/l	11/25/2019 15:05:19	
K5977-10	S	Total CN	P	2.7955	µg/l	11/25/2019 15:05:20	
K5977-11	S	Total CN	P	1.0217	µg/l	11/25/2019 15:05:21	
K6001-03	S	Total CN	P	0.2202	µg/l	11/25/2019 15:12:44	
PB125067BL	S	Total CN	P	0.3255	µg/l	11/25/2019 15:12:45	
PB125067BS	S	Total CN	P	95.8415	µg/l	11/25/2019 15:12:46	
K5959-13	S	Total CN	P	0.7471	µg/l	11/25/2019 15:12:47	
CCV3	S	Total CN	P	240.2106	µg/l	11/25/2019 15:12:48	
CCB3	S	Total CN	P	0.0818	µg/l	11/25/2019 15:12:49	
K5959-13DUP	S	Total CN	P	0.5539	µg/l	11/25/2019 15:12:50	
K5959-13MS	S	Total CN	P	35.8194	µg/l	11/25/2019 15:12:51	
K5959-13MSD	S	Total CN	P	33.4541	µg/l	11/25/2019 15:12:52	
CCV4	S	Total CN	P	240.3275	µg/l	11/25/2019 15:12:53	
CCB4	S	Total CN	P	0.365	µg/l	11/25/2019 15:12:54	

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AK

Instrument ID : Konelab

11/25/2019 14:06

Test Total CN


Accepted 11/25/2019 14:06

Factor 972.1

Bias 0.002

Coeff. of det. 0.999981

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.002	-0.0214	0.0000	
2	5.0PPBCN	0.007	5.2843	5.0000	
3	10PPBCN	0.013	10.5848	10.0000	
4	50PPBCN	0.054	50.4141	50.0000	
5	100PPBCN	0.104	99.4613	100.0000	
6	250PPBCN	0.257	248.4503	250.0000	
7	500PPBCN	0.517	500.8266	500.0000	