

LB106572

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AR

Instrument ID : Konelab

12/3/2019 13:00

Test: *Reactive*
~~Total~~ CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.984	0.0	0.103	
ICB1	-0.256	0.0	0.002	
CCV1	244.433	0.0	0.255	
CCB1	-0.099	0.0	0.002	
PB125198BL	-0.394	0.0	0.002	
K5676-38	-0.388	0.0	0.002	
K5676-38DUP	-0.559	0.0	0.002	
K6090-04	-0.320	0.0	0.002	
CCV2	246.462	0.0	0.257	
CCB2	-0.201	0.0	0.002	

N	10
Mean	58.666
SD	103.1049
CV%	175.75

Aquakem v. 7.2AQ1

Results from time period:

Tue Dec 03 12:49:37 2019

Tue Dec 03 12:49:46 2019



Sample Id	Sam/Ctr/c/	Test short nam	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.6813	µg/l	12/3/2019 11:14:24	
5.OPPBCN	A	Reactive CN	P	4.6543	µg/l	12/3/2019 11:14:25	
10PPBCN	A	Reactive CN	P	9.967	µg/l	12/3/2019 11:14:26	
50PPBCN	A	Reactive CN	P	47.1832	µg/l	12/3/2019 11:14:27	
100PPBCN	A	Reactive CN	P	101.7114	µg/l	12/3/2019 11:14:28	
250PPBCN	A	Reactive CN	P	254.4438	µg/l	12/3/2019 11:14:29	
500PPBCN	A	Reactive CN	P	497.7216	µg/l	12/3/2019 11:14:30	
ICV1	S	Reactive CN	P	97.9845	µg/l	12/3/2019 12:49:37	
ICB1	S	Reactive CN	P	-0.2559	µg/l	12/3/2019 12:49:38	
CCV1	S	Reactive CN	P	244.433	µg/l	12/3/2019 12:49:39	
CCB1	S	Reactive CN	P	-0.0992	µg/l	12/3/2019 12:49:40	
PB125198BL	S	Reactive CN	P	-0.3944	µg/l	12/3/2019 12:49:41	
K5676-38	S	Reactive CN	P	-0.3882	µg/l	12/3/2019 12:49:42	
K5676-38DUP	S	Reactive CN	P	-0.5588	µg/l	12/3/2019 12:49:43	
K6090-04	S	Reactive CN	P	-0.3196	µg/l	12/3/2019 12:49:44	
CCV2	S	Reactive CN	P	246.4622	µg/l	12/3/2019 12:49:45	
CCB2	S	Reactive CN	P	-0.2013	µg/l	12/3/2019 12:49:46	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : G

Instrument ID : Konelab

12/3/2019 11:15


Test Total CN

Accepted 12/3/2019 11:15

Factor 968.5
Bias 0.002

Coeff. of det. 0.999823

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.001	-0.6813	0.0000	
2	5.0PPBCN	0.007	4.6543	5.0000	
3	10PPBCN	0.012	9.9670	10.0000	
4	50PPBCN	0.051	47.1832	50.0000	
5	100PPBCN	0.107	101.7114	100.0000	
6	250PPBCN	0.265	254.4438	250.0000	
7	500PPBCN	0.516	497.7216	500.0000	