

Clear table

IC-2 300.0
12-10-19 90


ident	concentra tion F-	concentrati on CL-	concentrati on NO2	concentrati on BR-	concentrati on NO3	concentrati on HPO4	concentrati on SO4	file name	date time	Initial wt/ Final Vol	Analyst
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2019-12-02	12/2/19 10:56		CT/AP
STD2	0.2720	0.4060	0.3910	1.2920	0.3260	0.1450	2.2400	2019-12-02	12/2/19 11:25		CT/AP
STD3	0.3960	0.6620	0.6250	2.1250	0.5300	0.8360	3.4100	2019-12-02	12/2/19 11:54		CT/AP
STD4	0.8280	1.2340	1.1520	3.9940	1.0030	2.1520	6.0950	2019-12-02	12/2/19 12:23		CT/AP
STD5	1.9260	2.8420	2.6260	9.0000	2.2820	5.1270	13.8610	2019-12-02	12/2/19 12:53		CT/AP
STD6	3.8710	5.8910	5.5700	18.7720	4.7480	9.8150	28.9730	2019-12-02	12/2/19 13:22		CT/AP
STD7	5.1060	7.7170	7.7300	25.7100	6.4440	12.5700	38.6390	2019-12-02	12/2/19 13:51		CT/AP
ICV	1.9790	2.9590	2.9160	9.9090	2.4610	5.4570	14.5220	2019-12-02	12/2/19 15:18		CT/AP
ICB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2019-12-02	12/2/19 15:48		CT/AP
CCV	1.9950	2.7840	2.8700	9.7430	2.4690	5.0810	14.6950	2019-12-09	12/9/19 10:25		CT/AP
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2019-12-09	12/9/19 10:57		CT/AP
LB10664BLW	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2019-12-09	12/9/19 11:26		CT/AP
LB10664BSW	1.9380	2.6690	2.7340	9.2860	2.3610	5.5200	14.1920	2019-12-09	12/9/19 11:55		CT/AP
K6211-01	0.0000	0.0000	0.0000	0.0000	0.2700	0.0000	0.0000	2019-12-09	12/9/19 12:24		CT/AP
K6209-01	0.0000	0.0000	0.0000	0.0000	0.0910	0.0000	0.0000	2019-12-09	12/9/19 12:54		CT/AP
K6209-01DUP	0.0000	0.0000	0.0000	0.0000	0.0970	0.0000	0.0000	2019-12-09	12/9/19 13:23		CT/AP
K6209-01MS	1.9340	2.7420	2.8170	9.4640	2.4600	4.9930	14.3280	2019-12-09	12/9/19 13:52		CT/AP
K6209-01MSD	1.8190	2.6140	2.6940	9.0560	2.3520	4.7930	13.6700	2019-12-09	12/9/19 14:21		CT/AP
CCV	1.9850	2.7930	2.8770	9.7250	2.4770	5.6750	14.5310	2019-12-09	12/9/19 14:50		CT/AP
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2019-12-09	12/9/19 15:19		CT/AP

CALIBRATION OF COMPONENT F-

IC-2 cal

Method: AnionsIC2-120219.mtw
Equation: $Q = 0.0752034 \cdot A + 1.81152$
RSD: 4.391 %
Correlation coefficient: 0.999185

12-02-19
AK



K3 = 0 K2 = 0 K1 = 0.0752034 K0 = 1.81152

Base: Area

Ref.channel: chl

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1		0	0	0	0		0
2	2.296	44.86	0.2	20	5.516		
3	4.329	80.87	0.4	20	5.516		
4	8.943	195	0.8	20	5.516		
5	24.22	488.7	2	20	5.516		
6	53.88	1010	4	20	5.516		
7	71.73	1334	5	20	5.516		

CALIBRATION OF COMPONENT CL-

Method: AnionsIC2-120219.mtw
 Equation: $Q = 0.120863 \cdot A + 1.2447$
 RSD: 6.200 %
 Correlation coefficient: 0.998374



K3 = 0 K2 = 0 K1 = 0.120863 K0 = 1.2447

Base: Area

Ref.channel: ch1

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	3.158	55.35	0.3	20	8.443		
3	5.833	99.99	0.6	20	8.443		
4	11.23	193.8	1.2	20	8.443		
5	26.29	453.2	3	20	8.443		
6	56.28	947	6	20	8.443		
7	75.7	1267	7.5	20	8.443		

CALIBRATION OF COMPONENT NO2

Method: AnionsIC2-120219.mtw
 Equation: $Q = 0.0576239 \cdot A + 2.62969$
 RSD: 6.907 %
 Correlation coefficient: 0.997982




K3 = 0 K2 = 0 K1 = 0.0576239 K0 = 2.62969
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	5.575	104.3	0.3	20	10.18		
3	10.16	190.2	0.6	20	10.18		
4	19.71	377.2	1.2	20	10.18		
5	46.07	909.8	3	20	10.18		
6	94.95	1963	6	20	10.18		
7	122.7	2637	7.5	20	10.18		

CALIBRATION OF COMPONENT BR-

Method: AnionsIC2-120219.mtw
 Equation: $Q = 0.30011 \cdot A + 4.35517$
 RSD: 6.082 %
 Correlation coefficient: 0.998435




K3 = 0 K2 = 0 K1 = 0.30011 K0 = 4.35517
 Base: Area
 Ref.channel: chl
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1		0	0	0	0		0
2	3.041		71.85	1	20		13.07
3	5.607		131.9	2	20		13.07
4	11.13		262.5	4	20		13.07
5	26.41		608.2	10	20		13.07
6	56.7		1271	20	20		13.07
7	76.24		1699	25	20		13.07

±□¿\$

CALIBRATION OF COMPONENT NO3

Method: AnionsIC2-120219.mtw
 Equation: $Q = 0.0512792 \cdot A + 1.26972$
 RSD: 6.621 %
 Correlation coefficient: 0.998146




K3 = 0 K2 = 0 K1 = 0.0512792 K0 = 1.26972
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used File
1	0	0	0	0	0	0
2	3.627	106.8	0.25	20	15.04	
3	6.583	190.8	0.5	20	15.04	
4	12.97	378.3	1	20	15.04	
5	30.31	884	2.5	20	15.04	
6	64.26	1848	5	20	15.04	
7	86.84	2489	6.25	20	15.04	

CALIBRATION OF COMPONENT HPO4

Method: AnionsIC2-120219.mtw
 Equation: $Q = 0.0895479 \cdot A - 10.5374$
 RSD: 4.024 %
 Correlation coefficient: 0.999316



K3 = 0 K2 = 0 K1 = 0.0895479 K0 = -10.5374

Base: Area

Ref.channel: ch1

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	3.632	182.8	0.5	20	20.37		
3	6.723	334	1	20	20.37		
4	12.66	614.6	2	20	20.37		
5	27.3	1271	5	20	20.37		
6	51.79	2303	10	20	20.37		
7	66.92	2925	12.5	20	20.37		

CALIBRATION OF COMPONENT SO4

Method: AnionsIC2-120219.mtw
 Equation: $Q = 0.180193 \cdot A + 9.63188$
 RSD: 6.659 %
 Correlation coefficient: 0.998124



K3 = 0 K2 = 0 K1 = 0.180193 K0 = 9.63188
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		
2	4.314	178.7	1.5	20	23.43		
3	7.703	321	3	20	23.43		
4	15.05	625.1	6	20	23.43		
5	35.66	1486	15	20	23.43		
6	76.51	3155	30	20	23.43		
7	103.2	4235	37.5	20	23.43		

Report date: 12/4/2019 10:47:14 AM
Printed by: wet


Ident: STD1
Analysis from: 12/2/2019 10:56:36 AM
File: _2019-12-02_

Last save: 12/2/2019 11:23:07 AM

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23817

Last save: 12/2/2019 9:08:

SAMPLE:
: CT/AP
Vial number: 2
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 12/4/2019 10:47:28 AM
Printed by: wet


Ident: STD2
Analysis from: 12/2/2019 11:25:43 AM
File: _2019-12-02_

Last save: 12/2/2019 11:52:13 AM

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23818

Last save: 12/2/2019 11:23

SAMPLE:
: CT/AP
Vial number: 3
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.45	0.233	2.29	8.95	44.857	6.02
2	8.35	0.205	3.16	12.31	55.349	7.43
3	10.09	0.246	5.58	21.74	104.339	14.01
4	13.02	0.301	3.04	11.86	71.853	9.65
5	14.99	0.354	3.62	14.13	106.768	14.34
6	19.95	0.649	3.63	14.16	182.847	24.55
7	22.95	0.573	4.31	16.81	178.743	24.00
7	26.50	0.366	25.63	99.95	744.757	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/4/2019 10:47:42 AM
Printed by: wet


Ident: STD3
Analysis from: 12/2/2019 11:54:51 AM
File: _2019-12-02_

Last save: 12/2/2019 12:21:21 PM

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23819

Last save: 12/2/2019 11:52

SAMPLE:
: CT/AP
Vial number: 4
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.45	0.223	4.33	9.22	80.870	6.00
2	8.34	0.204	5.83	12.43	99.989	7.41
3	10.06	0.245	10.16	21.65	190.168	14.10
4	12.99	0.298	5.61	11.94	131.936	9.78
5	14.94	0.350	6.58	14.02	190.757	14.14
6	19.93	0.639	6.72	14.32	333.952	24.76
7	22.93	0.572	7.70	16.41	320.970	23.80
7	26.50	0.362	46.94	99.99	1348.642	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/4/2019 10:47:54 AM
Printed by: wet


Ident: STD4
Analysis from: 12/2/2019 12:23:59 PM
File: _2019-12-02_

Last save: 12/2/2019 12:50:29 PM

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23820

Last save: 12/2/2019 12:21

SAMPLE:
: CT/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.45	0.225	8.94	9.75	194.998	7.37
2	8.33	0.204	11.23	12.25	193.786	7.32
3	10.06	0.251	19.71	21.49	377.229	14.25
4	12.96	0.296	11.14	12.14	262.482	9.92
5	14.90	0.350	12.97	14.14	378.281	14.29
6	19.92	0.625	12.65	13.80	614.631	23.22
7	22.93	0.570	15.05	16.41	625.148	23.62
7	26.50	0.360	91.68	100.00	2646.555	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/4/2019 10:48:40 AM
Printed by: wet


Ident: STD5
Analysis from: 12/2/2019 12:53:08 PM
File: _2019-12-02_

Last save: 12/2/2019 1:19:38 PM

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23821

Last save: 12/2/2019 12:50

SAMPLE:
: CT/AP
Vial number: 6
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.45	0.219	24.22	11.20	488.743	8.01
2	8.33	0.204	26.29	12.16	453.221	7.43
3	10.04	0.255	46.07	21.30	909.837	14.91
4	12.94	0.292	26.41	12.21	608.236	9.97
5	14.83	0.351	30.31	14.02	884.022	14.49
6	19.90	0.599	27.30	12.62	1270.514	20.82
7	22.93	0.565	35.66	16.49	1486.471	24.36
7	26.50	0.355	216.25	100.00	6101.043	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/4/2019 10:48:56 AM
Printed by: wet


Ident: STD6
Analysis from: 12/2/2019 1:22:16 PM
File: _2019-12-02_

Last save: 12/2/2019 1:48:47 PM

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23822

Last save: 12/2/2019 1:19:

SAMPLE:
: CT/AP
Vial number: 7
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.45	0.207	53.88	11.86	1009.951	8.08
2	8.32	0.200	56.28	12.39	947.013	7.58
3	10.01	0.264	94.94	20.89	1962.950	15.71
4	12.90	0.286	56.70	12.48	1271.405	10.17
5	14.76	0.347	64.26	14.14	1848.352	14.79
6	19.86	0.573	51.78	11.40	2302.756	18.43
7	22.91	0.555	76.51	16.84	3155.054	25.25
7	26.50	0.347	454.36	100.00	12497.482	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/4/2019 10:49:05 AM
Printed by: wet


Ident: STD7
Analysis from: 12/2/2019 1:51:25 PM
File: _2019-12-02_

Last save: 12/2/2019 2:17:55 PM

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23823

Last save: 12/2/2019 1:48:

SAMPLE:
: CT/AP
Vial number: 8
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.46	0.207	71.73	11.89	1333.776	8.04
2	8.33	0.199	75.70	12.55	1266.716	7.64
3	10.00	0.277	122.69	20.34	2637.313	15.90
4	12.90	0.285	76.24	12.64	1698.834	10.24
5	14.74	0.346	86.83	14.39	2488.596	15.00
6	19.86	0.564	66.92	11.09	2925.006	17.64
7	22.92	0.552	103.17	17.10	4235.156	25.54
7	26.50	0.347	603.27	100.00	16585.396	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/4/2019 10:50:23 AM
Printed by: wet


Ident: ICV
Analysis from: 12/2/2019 3:18:52 PM
File: _2019-12-02_

Last save: 12/4/2019 10:50:14 AM

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23826

Last save: 12/2/2019 2:17:

SAMPLE:
: CT/AP
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.45	0.217	24.75	10.96	502.172	7.82
2	8.33	0.204	27.81	12.31	479.323	7.46
3	10.04	0.261	48.00	21.25	966.604	15.05
4	12.94	0.293	27.84	12.33	645.858	10.05
5	14.83	0.353	31.90	14.12	935.064	14.56
6	19.91	0.603	28.58	12.65	1336.367	20.80
7	22.94	0.572	37.01	16.38	1558.420	24.26
7	26.50	0.358	225.90	100.00	6423.808	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/4/2019 10:50:33 AM
Printed by: wet


Ident: ICB
Analysis from: 12/2/2019 3:48:02 PM
File: _2019-12-02_

Last save: 12/4/2019 10:50:15 AM

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23827

Last save: 12/2/2019 2:17:

SAMPLE:
: CT/AP
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD


Report date: 12/9/2019 4:52:00 PM
Printed by: wet

Ident: CCV
Analysis from: 12/9/2019 10:25:03 AM
File: 2019-12-09

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23894

SAMPLE:

CT/AP
Vial number: 2
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.46	0.222	24.52	10.39	506.435	8.03
2	8.36	0.199	28.48	12.06	450.453	7.14
3	10.08	0.256	49.06	20.78	950.444	15.06
4	13.01	0.286	28.76	12.18	634.778	10.06
5	14.93	0.351	33.02	13.99	938.071	14.87
6	19.84	0.542	29.35	12.43	1252.391	19.85
7	22.83	0.502	42.87	18.16	1577.580	25.00
7	26.50	0.337	236.07	100.00	6310.151	100.00


This report has been created by IC Net
METROHM LTD

Report date: 12/9/2019 4:52:08 PM
Printed by: wet

Ident: CCB
Analysis from: 12/9/2019 10:57:31 AM
File: 2019-12-09

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23895
Last save: 12/9/2019 11:24:02 AM
Last save: 12/2/2019 2:17:

SAMPLE: CT/AP
Vial number: 3
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 12/9/2019 4:52:16 PM
Printed by: wet


Ident: LB106664BLW
Analysis from: 12/9/2019 11:26:39 AM
File: 2019-12-09

Last save: 12/9/2019 4:51:53 PM

Method: AnionsIC2-120219.mtw

Run operator: wet
Analysis number: 23896

SAMPLE: CT/AP
Vial number: 4
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks


This report has been created by IC Net
METROHM LTD

Report date: 12/9/2019 4:52:25 PM
Printed by: wet

Ident: LB106664BSW
Analysis from: 12/9/2019 11:55:48 AM
File: 2019-12-09

Method: AnIonsIC2-120219.mtw
Run operator: wet
Analysis number: 23897
Last save: 12/9/2019 4:51:53 PM
Last save: 12/2/2019 2:17:

SAMPLE: CT/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.46	0.223	23.75	10.40	491.264	7.93
2	8.33	0.199	27.19	11.90	431.318	6.96
3	10.04	0.254	46.51	20.36	903.130	14.57
4	12.95	0.282	27.81	12.18	604.350	9.75
5	14.84	0.346	32.04	14.03	895.988	14.46
6	19.86	0.562	30.63	13.41	1350.587	21.79
7	22.86	0.516	40.50	17.73	1521.773	24.55
7	26.50	0.340	228.43	100.00	6198.410	100.00


This report has been created by IC Net
METROHM LTD

Report date: 12/9/2019 4:53:29 PM
Printed by: wet

Ident: K6211-01
Analysis from: 12/9/2019 12:24:58 PM
File: 2019-12-09

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23898
Method: Last save: 12/2/2019 2:17:
Last save: 12/9/2019 4:53:26 PM

SAMPLE: CT/AP
Vial number: 7
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention	Width/2	Height	Height	Area	Area
7	26.50	0.050	2.57	99.96	80.539	100.00
7	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
5	14.97	0.352	2.57	99.96	80.539	100.00
4	0.00	0.000	0.00	0.00	0.000	0.00
3	0.00	0.000	0.00	0.00	0.000	0.00
2	0.00	0.000	0.00	0.00	0.000	0.00
1	0.00	0.000	0.00	0.00	0.000	0.00
	min	min	mV	%	mV*sec	%
	Retention	Width/2	Height	Height	Area	Area

This report has been created by IC Net
METROHM LTD


Report date: 12/9/2019 4:53:57 PM
Printed by: wet

Ident: K6209-01
Analysis from: 12/9/2019 12:54:07 PM

File: 2019-12-09

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23899
Last save: 12/9/2019 4:53:54 PM
Last save: 12/2/2019 2:17:

SAMPLE: :
Vial number: 8
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
7	26.50	0.051	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
5	14.99	0.354	0.43	99.87	10.872	100.00
4	0.00	0.000	0.00	0.00	0.000	0.00
3	0.00	0.000	0.00	0.00	0.000	0.00
2	0.00	0.000	0.00	0.00	0.000	0.00
1	0.00	0.000	0.00	0.00	0.000	0.00


Report date: 12/9/2019 4:56:06 PM
Printed by: wet

Ident: K6209-01DUP
Analysis from: 12/9/2019 1:23:16 PM
File: 2019-12-09

Method: AnIonsIC2-120219.mtw
Run operator: wet
Analysis number: 23900
SAMPLE: CT/AP
Vial number: 9
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000

Last save: 12/9/2019 4:54:46 PM

Last save: 12/2/2019 2:17:



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
7	26.50	0.052	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
5	14.98	0.364	0.44	99.66	12.962	100.00
4	0.00	0.000	0.00	0.00	0.000	0.00
3	0.00	0.000	0.00	0.00	0.000	0.00
2	0.00	0.000	0.00	0.00	0.000	0.00
1	0.00	0.000	0.00	0.00	0.000	0.00


This report has been created by IC Net
METROHM LTD

Report date: 12/9/2019 4:56:25 PM
Printed by: wet

Ident: K6209-01MS
Analysis from: 12/9/2019 1:52:24 PM
File: 2019-12-09

Method: AnIonsIC2-120219.mtw
Run operator: wet
Analysis number: 23901
Method: Last save: 12/2/2019 2:17:
Last save: 12/9/2019 2:18:54 PM

SAMPLE: :
CT/AP
Vial number: 10
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention	Width/2	Height	Height	Area	Area
1	5.46	0.224	23.91	10.48	490.307	7.93
2	8.33	0.201	27.66	12.13	443.450	7.17
3	10.03	0.253	48.26	21.16	931.921	15.06
4	12.91	0.285	28.22	12.37	616.208	9.96
5	14.79	0.348	33.09	14.51	934.542	15.11
6	19.89	0.576	27.29	11.97	1232.812	19.93
7	22.89	0.532	39.64	17.38	1536.823	24.84
7	26.50	0.345	228.08	100.00	6186.061	100.00

This report has been created by IC Net
METROHM LTD


Report date: 12/9/2019 4:56:34 PM
Printed by: wet

Ident: K6209-01MSD
Analysis from: 12/9/2019 2:21:32 PM
File: 2019-12-09

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23902
SAMPLE: CT/AP
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000

Last save: 12/9/2019 2:48:03 PM

Last save: 12/2/2019 2:17:



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.45	0.220	22.37	10.29	459.629	7.78
2	8.32	0.200	26.45	12.16	422.196	7.15
3	10.02	0.252	46.25	21.26	889.232	15.06
4	12.89	0.284	26.99	12.41	588.982	9.97
5	14.77	0.347	31.64	14.55	892.702	15.12
6	19.88	0.580	26.14	12.02	1188.207	20.12
7	22.87	0.534	37.64	17.31	1463.812	24.79
7	26.50	0.345	217.48	100.00	5904.761	100.00


This report has been created by IC Net
METROHM LTD

Report date: 12/9/2019 4:56:44 PM
Printed by: wet

Ident: GCV
Analysis from: 12/9/2019 2:50:44 PM
File: 2019-12-09

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23903
Last save: 12/2/2019 2:17:15 PM

SAMPLE: CT/AP
Vial number: 2
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.46	0.222	24.48	10.44	503.741	7.84
2	8.33	0.201	28.23	12.04	451.836	7.03
3	10.03	0.257	48.76	20.80	952.900	14.82
4	12.91	0.284	28.92	12.34	633.574	9.86
5	14.78	0.347	33.36	14.23	941.160	14.64
6	19.89	0.579	30.61	13.06	1385.243	21.55
7	22.89	0.535	40.05	17.08	1559.367	24.26
7	26.50	0.346	234.42	100.00	6427.821	100.00

Report date: 12/9/2019 4:56:53 PM
Printed by: wet

Ident: CCB
Analysis from: 12/9/2019 3:19:54 PM
File: 2019-12-09

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23904

SAMPLE: :
Vial number: 3
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000

Last save: 12/9/2019 3:46:25 PM
Last save: 12/2/2019 2:17:



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD