

CB106858

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

Instrument ID : Konelab

12/19/2019 12:59

Test: Ammonia-N

mg/L

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.946	0.0	0.176	
ICB1	0.005	0.0	0.039	
CCV1	0.941	0.0	0.175	
CCB1	0.008	0.0	0.039	
PB125557BL	-0.000	0.0	0.038	
PB125557BS	0.928	0.0	0.173	
K6315-03	0.522	0.0	0.114	
K6315-03DUP	0.536	0.0	0.116	
K6315-03MS	1.389	0.0	0.241	
K6315-03MS	1.353	0.0	0.235	
K6315-03MSD	1.540	0.0	0.263	
K6315-03MSD	1.383	0.0	0.240	
K6316-03	0.378	0.0	0.093	
K6347-01	0.100	0.0	0.053	
K6347-02	0.117	0.0	0.055	
K6347-03	0.093	0.0	0.052	
CCV2	0.931	0.0	0.174	
CCB2	0.010	0.0	0.040	
K6369-01	0.099	0.0	0.053	
K6347-04	0.047	0.0	0.045	
K6369-02	0.271	0.0	0.078	
K6369-03	0.281	0.0	0.079	
K6369-04	1.289	0.0	0.226	
CCV3	0.940	0.0	0.175	
CCB3	0.015	0.0	0.040	

AK
12/19/19

N	25
Mean	0.565
SD	0.5409
CV%	95.76

Aquakem v. 7.2AQ1

Results from time period:

Thu Dec 19 11:11:31 2019

Thu Dec 19 12:35:47 2019

AR

Sample Id	Sam/Ctr/c/	Test short Test type	Result	Result unit	Result date and time	Stat
0.1PPM	A	Ammonia-I P	0.0936	mg/l	12/19/2019 11:39:45	
0.2PPM	A	Ammonia-I P	0.2023	mg/l	12/19/2019 11:39:46	
0.4PPM	A	Ammonia-I P	0.4054	mg/l	12/19/2019 11:39:47	
1.0PPM	A	Ammonia-I P	0.9858	mg/l	12/19/2019 11:39:48	
1.3PPM	A	Ammonia-I P	1.3538	mg/l	12/19/2019 11:39:49	
2.0PPM	A	Ammonia-I P	1.9925	mg/l	12/19/2019 11:39:50	
ICV1	S	Ammonia-I P	0.9464	mg/l	12/19/2019 12:20:11	
ICB1	S	Ammonia-I P	0.0052	mg/l	12/19/2019 12:20:12	
CCV1	S	Ammonia-I P	0.9414	mg/l	12/19/2019 12:20:13	
CCB1	S	Ammonia-I P	0.0081	mg/l	12/19/2019 12:20:14	
PB125557BL	S	Ammonia-I P	-0.0004	mg/l	12/19/2019 12:20:15	
PB125557BS	S	Ammonia-I P	0.9279	mg/l	12/19/2019 12:20:16	
K6315-03	S	Ammonia-I P	0.5218	mg/l	12/19/2019 12:20:17	
K6315-03DUP	S	Ammonia-I P	0.5359	mg/l	12/19/2019 12:20:18	
K6315-03MS	S	Ammonia-I P	1.3889	mg/l	12/19/2019 12:20:19	
K6315-03MSD	S	Ammonia-I P	1.383	mg/l	12/19/2019 12:20:22	
K6316-03	S	Ammonia-I P	0.3781	mg/l	12/19/2019 12:30:52	
K6347-01	S	Ammonia-I P	0.0999	mg/l	12/19/2019 12:30:53	
K6347-02	S	Ammonia-I P	0.1166	mg/l	12/19/2019 12:30:54	
K6347-03	S	Ammonia-I P	0.0931	mg/l	12/19/2019 12:30:55	
CCV2	S	Ammonia-I P	0.9311	mg/l	12/19/2019 12:30:56	
CCB2	S	Ammonia-I P	0.0098	mg/l	12/19/2019 12:30:57	
K6369-01	S	Ammonia-I P	0.0988	mg/l	12/19/2019 12:30:58	
K6347-04	S	Ammonia-I P	0.0468	mg/l	12/19/2019 12:30:59	
K6369-02	S	Ammonia-I P	0.2707	mg/l	12/19/2019 12:31:00	
K6369-03	S	Ammonia-I P	0.2812	mg/l	12/19/2019 12:31:01	
K6369-04	S	Ammonia-I P	1.2893	mg/l	12/19/2019 12:31:02	
CCV3	S	Ammonia-I P	0.9396	mg/l	12/19/2019 12:31:03	
CCB3	S	Ammonia-I P	0.0154	mg/l	12/19/2019 12:35:46	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : *fl*

Instrument ID : Konelab

12/19/2019 11:44


Test Ammonia-N

Accepted 12/19/2019 11:44

Factor 6.86
Bias 0.038

Coeff. of det. 0.999728

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	NH3-2PPM	0.052	0.0936	0.1000	
2	NH3-2PPM	0.068	0.2023	0.2000	
3	NH3-2PPM	0.097	0.4054	0.4000	
4	NH3-2PPM	0.182	0.9858	1.0000	
5	NH3-2PPM	0.235	1.3538	1.3333	
6	NH3-2PPM	0.329	1.9925	2.0000	