

IC-2 300.0

12/26/19

Clear table

ident	concentra tion F-	concentrati on CL-	concentrati on NO2	concentrati on BR-	concentrati on NO3	concentrati on HPO4	concentrati on SO4	file name	date time	Initial wt/ Final Vol	Analyst
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2019-12-19_	12/19/19 15:26		CT/AP
STD2	0.2500	0.3990	0.3800	1.2380	0.3170	0.3550	2.0020	2019-12-19_	12/19/19 15:56		CT/AP
STD3	0.4400	0.6420	0.6580	2.2270	0.5560	0.9870	3.3430	2019-12-19_	12/19/19 16:25		CT/AP
STD4	0.8110	1.1640	1.1660	3.8820	0.9740	2.0690	5.8260	2019-12-19_	12/19/19 16:54		CT/AP
STD5	1.9030	2.8750	2.8740	9.5760	2.3840	5.1810	14.2170	2019-12-19_	12/19/19 17:23		CT/AP
STD6	3.8700	5.8950	5.9010	19.6930	4.9110	9.9480	29.3600	2019-12-19_	12/19/19 17:52		CT/AP
STD7	5.1270	7.6250	7.6200	25.3850	6.3580	12.4600	38.2530	2019-12-19_	12/19/19 18:21		CT/AP
ICV	2.0360	2.9970	2.9950	9.9910	2.4820	5.3130	14.8340	2019-12-19_	12/19/19 19:49		CT/AP
ICB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2019-12-19_	12/19/19 20:18		CT/AP
CCV	1.8980	3.0210	2.9300	9.7430	2.5400	4.9830	14.2260	2019-12-24_	12/24/19 8:49		CT/AP
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2019-12-24_	12/24/19 9:26		CT/AP
LB106919BLW	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2019-12-24_	12/24/19 9:55		CT/AP
LB106919BSW	1.8430	3.0090	2.9560	9.6810	2.5390	5.1140	14.2240	2019-12-24_	12/24/19 10:24		CT/AP
K6434-03	0.0000	253.7440	0.0000	0.0000	0.0000	0.0000	25.7630	2019-12-24_	12/24/19 15:42		CT/AP
K6434-03DUP	0.0000	253.1120	0.0000	0.0000	0.0000	0.0000	25.6010	2019-12-24_	12/24/19 16:11		CT/AP
K6434-03MS	3.1960	237.0540	1.5110	10.2060	2.5420	10.7090	38.2860	2019-12-24_	12/24/19 16:40		CT/AP
K6434-03MSD	3.1580	236.2070	1.5460	10.6450	2.7140	11.5800	38.8250	2019-12-24_	12/24/19 17:09		CT/AP
CCV	2.0060	3.0590	2.9200	9.8250	2.6490	7.1840	14.1310	2019-12-24_	12/24/19 17:38		CT/AP
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2019-12-24_	12/24/19 18:07		CT/AP


CALIBRATION OF COMPONENT F-

Method: AnionsIC2-121919.mtw
 Equation: $Q = 0.0755513 \cdot A + 1.93672$
 RSD: 5.227 %
 Correlation coefficient: 0.998845

IC2 CAL

12-19-19

AL



K3 = 0 K2 = 0 K1 = 0.0755513 K0 = 1.93672

Base: Area

Ref.channel: chl

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	1.602	40.47	0.2	20	5.516		
3	3.353	90.84	0.4	20	5.516		
4	7.546	189	0.8	20	5.516		
5	20.92	478.2	2	20	5.516		
6	45.31	998.7	4	20	5.516		
7	62.11	1332	5	20	5.516		

CALIBRATION OF COMPONENT CL-

Method: AnionsIC2-121919.mtw
 Equation: $Q = 0.140776 \cdot A + 1.87523$
 RSD: 3.786 %
 Correlation coefficient: 0.999394



K3 = 0 K2 = 0 K1 = 0.140776 K0 = 1.87523

Base: Area

Ref.channel: ch1

ISTD:

Formula: Linear


Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	2.573	43.41	0.3	20	8.443		
3	5.183	77.84	0.6	20	8.443		
4	10.01	152	1.2	20	8.443		
5	26.33	395.2	3	20	8.443		
6	55.7	824.2	6	20	8.443		
7	72.45	1070	7.5	20	8.443		

ŷ

CALIBRATION OF COMPONENT NO2

Method: AnionsIC2-121919.mtw
 Equation: $Q = 0.0627248 \cdot A + 2.30587$
 RSD: 3.646 %
 Correlation coefficient: 0.999438



K3 = 0 K2 = 0 K1 = 0.0627248 K0 = 2.30587

Base: Area

Ref.channel: ch1

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	4.615	84.53	0.3	20	10.18		
3	9.322	173.1	0.6	20	10.18		
4	17.96	334.9	1.2	20	10.18		
5	46.4	879.7	3	20	10.18		
6	93.12	1845	6	20	10.18		
7	120	2393	7.5	20	10.18		

CALIBRATION OF COMPONENT BR-

Method: AnionsIC2-121919.mtw
 Equation: $Q = 0.316728 \cdot A + 7.75699$
 RSD: 3.568 %
 Correlation coefficient: 0.999462



K3 = 0 K2 = 0 K1 = 0.316728 K0 = 7.75699

Base: Area

Ref.channel: ch1

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	2.506	53.65	1	20	13.07		
3	5.233	116.1	2	20	13.07		
4	10.12	220.6	4	20	13.07		
5	26.95	580.2	10	20	13.07		
6	57.56	1219	20	20	13.07		
7	75.01	1578	25	20	13.07		

CALIBRATION OF COMPONENT NO3

Method: AnionsIC2-121919.mtw
 Equation: $Q = 0.0567155 \cdot A + 1.9162$
 RSD: 3.923 %
 Correlation coefficient: 0.999349



K3 = 0 K2 = 0 K1 = 0.0567155 K0 = 1.9162

Base: Area

Ref.channel: chl

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	3.046	77.91	0.25	20	15.04		
3	6.249	162.2	0.5	20	15.04		
4	11.98	309.8	1	20	15.04		
5	31.2	806.9	2.5	20	15.04		
6	65.71	1698	5	20	15.04		
7	85.55	2208	6.25	20	15.04		

CALIBRATION OF COMPONENT HPO4

Method: AnionsIC2-121919.mtw
 Equation: $Q = 0.0942908 \cdot A - 3.46159$
 RSD: 2.433 %
 Correlation coefficient: 0.999750



K3 = 0 K2 = 0 K1 = 0.0942908 K0 = -3.46159

Base: Area

Ref.channel: chl

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	2.415	112	0.5	20	20.37		
3	5.344	246	1	20	20.37		
4	10.51	475.5	2	20	20.37		
5	26.2	1136	5	20	20.37		
6	51.59	2147	10	20	20.37		
7	65.47	2680	12.5	20	20.37		

CALIBRATION OF COMPONENT SO4

Method: AnionsIC2-121919.mtw
 Equation: $Q = 0.181366 \cdot A + 12.4427$
 RSD: 4.550 %
 Correlation coefficient: 0.999125



K3 = 0 K2 = 0 K1 = 0.181366 K0 = 12.4427

Base: Area

Ref.channel: chl

ISTD:

Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	3.868	152.1	1.5	20	23.43		
3	7.718	300.1	3	20	23.43		
4	14.7	573.8	6	20	23.43		
5	38.7	1499	15	20	23.43		
6	82.78	3169	30	20	23.43		
7	108.7	4150	37.5	20	23.43		

Report date: 12/23/2019 11:14:07 AM
Printed by: wet


Ident: STD1
Analysis from: 12/19/2019 3:26:59 PM
File: _2019-12-19_

Last save: 12/20/2019 9:09:00 AM

Method: AnionsIC2-121919.mtw
Run operator: wet
Analysis number: 24042

Last save: 12/19/2019 6:19

SAMPLE:
: CT/AP
Vial number: 2
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 12/23/2019 11:14:24 AM
Printed by: wet


Ident: STD2
Analysis from: 12/19/2019 3:56:07 PM
File: _2019-12-19_

Last save: 12/20/2019 9:09:00 AM

Method: AnionsIC2-121919.mtw
Run operator: wet
Analysis number: 24043

Last save: 12/19/2019 6:19

SAMPLE:
: CT/AP
Vial number: 3
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.46	0.274	1.60	7.76	40.468	7.17
2	8.35	0.206	2.57	12.46	43.408	7.69
3	10.08	0.245	4.61	22.37	84.529	14.98
4	13.03	0.294	2.51	12.15	53.654	9.51
5	15.00	0.350	3.04	14.76	77.906	13.81
6	19.89	0.594	2.41	11.70	112.029	19.86
7	22.87	0.536	3.87	18.75	152.123	26.97
7	26.50	0.357	20.62	99.96	564.118	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/23/2019 11:14:38 AM
Printed by: wet


Ident: STD3
Analysis from: 12/19/2019 4:25:15 PM
File: _2019-12-19_

Last save: 12/20/2019 9:09:01 AM

Method: AnionsIC2-121919.mtw
Run operator: wet
Analysis number: 24044

Last save: 12/19/2019 6:19

SAMPLE:
: CT/AP
Vial number: 4
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.46	0.273	3.35	7.90	90.838	7.79
2	8.34	0.204	5.18	12.22	77.838	6.67
3	10.08	0.250	9.32	21.98	173.143	14.85
4	13.00	0.295	5.23	12.34	116.126	9.96
5	14.96	0.350	6.25	14.74	162.230	13.91
6	19.90	0.588	5.34	12.60	246.023	21.09
7	22.88	0.533	7.72	18.20	300.076	25.73
7	26.50	0.356	42.40	99.99	1166.274	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/23/2019 11:14:51 AM
Printed by: wet


Ident: STD4
Analysis from: 12/19/2019 4:54:22 PM
File: _2019-12-19_

Last save: 12/20/2019 9:09:01 AM

Method: AnionsIC2-121919.mtw
Run operator: wet
Analysis number: 24045

Last save: 12/19/2019 6:19

SAMPLE:
: CT/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.47	0.258	7.54	9.11	188.985	8.38
2	8.35	0.204	10.01	12.09	151.994	6.74
3	10.07	0.248	17.96	21.68	334.935	14.85
4	12.98	0.292	10.12	12.22	220.644	9.78
5	14.91	0.350	11.98	14.46	309.778	13.73
6	19.90	0.579	10.50	12.68	475.545	21.08
7	22.89	0.531	14.70	17.75	573.814	25.44
7	26.50	0.352	82.82	99.99	2255.695	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/23/2019 11:15:01 AM
Printed by: wet


Ident: STD5
Analysis from: 12/19/2019 5:23:31 PM
File: _2019-12-19_

Last save: 12/20/2019 9:09:01 AM

Method: AnionsIC2-121919.mtw
Run operator: wet
Analysis number: 24046

Last save: 12/19/2019 6:19

SAMPLE:
: CT/AP
Vial number: 6
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.47	0.243	20.92	9.65	478.173	8.28
2	8.34	0.200	26.32	12.15	395.195	6.84
3	10.04	0.253	46.40	21.41	879.705	15.23
4	12.95	0.286	26.95	12.44	580.168	10.05
5	14.84	0.350	31.19	14.40	806.924	13.97
6	19.87	0.555	26.20	12.09	1135.752	19.67
7	22.87	0.524	38.70	17.86	1499.142	25.96
7	26.50	0.345	216.68	100.00	5775.059	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/23/2019 11:15:11 AM
Printed by: wet


Ident: STD6
Analysis from: 12/19/2019 5:52:39 PM
File: _2019-12-19_

Last save: 12/20/2019 9:09:01 AM

Method: AnionsIC2-121919.mtw
Run operator: wet
Analysis number: 24047

Last save: 12/19/2019 6:19

SAMPLE:
: CT/AP
Vial number: 7
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.48	0.240	45.31	10.03	998.717	8.39
2	8.34	0.198	55.70	12.33	824.179	6.93
3	10.03	0.269	93.12	20.61	1844.817	15.50
4	12.93	0.281	57.56	12.74	1219.057	10.24
5	14.79	0.350	65.71	14.55	1698.186	14.27
6	19.85	0.535	51.59	11.42	2146.735	18.04
7	22.87	0.515	82.77	18.32	3169.042	26.63
7	26.50	0.341	451.76	100.00	11900.734	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/23/2019 11:15:25 AM
Printed by: wet


Ident: STD7
Analysis from: 12/19/2019 6:21:47 PM
File: _2019-12-19_

Last save: 12/20/2019 9:09:59 AM

Method: AnionsIC2-121919.mtw
Run operator: wet
Analysis number: 24048

Last save: 12/19/2019 6:19

SAMPLE:
: CT/AP
Vial number: 8
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.49	0.240	62.11	10.54	1331.550	8.64
2	8.34	0.198	72.44	12.29	1069.966	6.94
3	10.02	0.270	120.03	20.37	2392.969	15.53
4	12.94	0.280	75.01	12.73	1578.430	10.24
5	14.78	0.350	85.55	14.52	2208.140	14.33
6	19.83	0.528	65.47	11.11	2679.590	17.39
7	22.86	0.513	108.66	18.44	4149.674	26.93
7	26.50	0.340	589.27	100.00	15410.320	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/23/2019 11:16:05 AM
Printed by: wet


Ident: ICV
Analysis from: 12/19/2019 7:49:10 PM
File: _2019-12-19_

Last save: 12/20/2019 9:10:00 AM

Method: AnionsIC2-121919.mtw
Run operator: wet
Analysis number: 24051

Last save: 12/19/2019 6:19

SAMPLE:
: CT/AP
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.48	0.251	22.06	9.93	513.278	8.52
2	8.39	0.204	27.02	12.17	412.466	6.85
3	10.13	0.258	47.48	21.38	918.295	15.25
4	13.11	0.294	27.35	12.32	606.385	10.07
5	15.05	0.362	31.45	14.16	841.437	13.97
6	19.88	0.560	26.76	12.05	1163.557	19.32
7	22.88	0.531	39.94	17.99	1567.230	26.02
7	26.50	0.351	222.06	100.00	6022.648	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/23/2019 11:16:16 AM
Printed by: wet


Ident: ICB
Analysis from: 12/19/2019 8:18:19 PM
File: _2019-12-19_

Last save: 12/20/2019 9:10:00 AM

Method: AnionsIC2-121919.mtw
Run operator: wet
Analysis number: 24052

Last save: 12/19/2019 6:19

SAMPLE:
:
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 12/26/2019 10:01:40 AM
Printed by: wet


Ident: CCV
Analysis from: 12/24/2019 8:49:31 AM
File: _2019-12-24_

Last save: 12/24/2019 9:16:02 AM

Method: AnionsIC2-121919.mtw
Run operator: wet
Analysis number: 24091

Last save: 12/20/2019 11:5

SAMPLE:
: CT/AP
Vial number: 2
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.48	0.269	19.40	9.07	476.789	8.17
2	8.38	0.207	25.86	12.09	415.804	7.12
3	10.12	0.267	44.63	20.87	897.434	15.38
4	13.09	0.295	26.44	12.36	590.763	10.12
5	15.02	0.366	30.49	14.26	861.741	14.77
6	19.83	0.533	26.19	12.25	1093.630	18.74
7	22.86	0.503	40.83	19.09	1500.112	25.70
7	26.50	0.349	213.83	99.99	5836.274	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/26/2019 10:01:52 AM
Printed by: wet


Ident: CCB
Analysis from: 12/24/2019 9:26:40 AM
File: _2019-12-24_

Last save: 12/24/2019 9:53:10 AM

Method: AnionsIC2-121919.mtw
Run operator: wet
Analysis number: 24092

Last save: 12/20/2019 11:5

SAMPLE:
: CT/AP
Vial number: 3
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 12/26/2019 10:02:03 AM
Printed by: wet


Ident: LB106919BLW
Analysis from: 12/24/2019 9:55:48 AM
File: _2019-12-24_

Last save: 12/26/2019 10:01:23 AM

Method: AnionsIC2-121919.mtw
Run operator: wet
Analysis number: 24093

Last save: 12/20/2019 11:5

SAMPLE:
: CT/AP
Vial number: 4
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 12/26/2019 10:02:13 AM
Printed by: wet


Ident: LB106919BSW
Analysis from: 12/24/2019 10:24:56 AM
File: _2019-12-24_

Last save: 12/26/2019 10:01:23 AM

Method: AnionsIC2-121919.mtw
Run operator: wet
Analysis number: 24094

Last save: 12/20/2019 11:5

SAMPLE:
: CT/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.48	0.262	19.50	9.02	462.168	7.90
2	8.36	0.206	26.09	12.07	414.106	7.08
3	10.08	0.264	45.41	21.01	905.901	15.48
4	13.00	0.290	26.82	12.40	586.808	10.03
5	14.91	0.359	31.03	14.35	861.665	14.72
6	19.85	0.537	26.68	12.34	1121.539	19.16
7	22.88	0.505	40.65	18.80	1499.884	25.63
7	26.50	0.346	216.17	100.00	5852.070	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/26/2019 10:03:21 AM
Printed by: wet


Ident: K6434-0103 ^{Ak} 12/26/19
Analysis from: 12/24/2019 3:42:05 PM
File: _2019-12-24_

Last save: 12/26/2019 10:03:20 AM

Method: AnionsIC2-121919.mtw
Run operator: wet
Analysis number: 24097

Last save: 12/20/2019 11:5

SAMPLE:
: CT/AP
Vial number: 6
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.55	0.359	1464.85	95.03	36036.100	92.86
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	22.97	0.484	76.60	4.97	2772.364	7.14
7	26.50	0.120	1541.45	100.00	38808.464	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/26/2019 10:03:48 AM
Printed by: wet


Ident: K6434-01DUP
Analysis from: 12/24/2019 4:11:14 PM
File: _2019-12-24_

Last save: 12/26/2019 10:03:47 AM

Method: AnionsIC2-121919.mtw
Run operator: wet
Analysis number: 24098

Last save: 12/20/2019 11:5

SAMPLE:
: CT/AP
Vial number: 7
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.54	0.358	1463.13	95.09	35946.326	92.88
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	22.99	0.487	75.47	4.90	2754.577	7.12
7	26.50	0.121	1538.59	100.00	38700.903	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/26/2019 10:04:02 AM
Printed by: wet


Ident: K6434-01MS
Analysis from: 12/24/2019 4:40:24 PM
File: _2019-12-24_

Last save: 12/24/2019 5:06:54 PM

Method: AnionsIC2-121919.mtw
Run operator: wet
Analysis number: 24099

Last save: 12/20/2019 11:5

SAMPLE:
: CT/AP
Vial number: 8
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.50	0.380	25.38	1.47	820.287	1.91
2	8.52	0.336	1451.66	84.06	33665.005	78.49
3	10.09	0.257	27.15	1.57	444.964	1.04
4	13.13	0.275	29.35	1.70	619.973	1.45
5	15.01	0.350	31.90	1.85	862.616	2.01
6	19.98	0.665	43.81	2.54	2308.181	5.38
7	22.96	0.482	117.75	6.82	4153.313	9.68
7	26.50	0.392	1727.01	100.00	42874.339	99.97

This report has been created by IC Net
METROHM LTD

Report date: 12/26/2019 10:04:20 AM
Printed by: wet


Ident: K6434-01MSD
Analysis from: 12/24/2019 5:09:33 PM
File: _2019-12-24_

Last save: 12/24/2019 5:36:03 PM

Method: AnionsIC2-121919.mtw
Run operator: wet
Analysis number: 24100

Last save: 12/20/2019 11:5

SAMPLE:
: CT/AP
Vial number: 9
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.51	0.382	25.00	1.44	810.263	1.88
2	8.51	0.335	1449.84	83.54	33544.691	77.81
3	10.08	0.251	28.48	1.64	456.264	1.06
4	13.11	0.274	30.71	1.77	647.680	1.50
5	14.98	0.350	33.52	1.93	923.409	2.14
6	19.97	0.637	48.90	2.82	2492.967	5.78
7	22.97	0.484	118.96	6.85	4212.794	9.77
7	26.50	0.388	1735.40	100.00	43088.067	99.94

This report has been created by IC Net
METROHM LTD

Report date: 12/26/2019 10:04:31 AM
Printed by: wet


Ident: CCV
Analysis from: 12/24/2019 5:38:48 PM
File: _2019-12-24_

Last save: 12/24/2019 6:05:18 PM

Method: AnionsIC2-121919.mtw
Run operator: wet
Analysis number: 24101

Last save: 12/20/2019 11:5

SAMPLE:
: CT/AP
Vial number: 2
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.48	0.269	20.47	9.03	505.371	7.94
2	8.36	0.205	26.49	11.69	421.314	6.62
3	10.06	0.259	45.86	20.24	894.220	14.04
4	12.98	0.291	27.14	11.98	595.920	9.36
5	14.88	0.358	31.60	13.95	900.342	14.14
6	19.89	0.571	34.15	15.07	1560.534	24.51
7	22.92	0.504	40.91	18.05	1489.671	23.40
7	26.50	0.351	226.62	100.00	6367.372	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/26/2019 10:04:43 AM
Printed by: wet


Ident: CCB
Analysis from: 12/24/2019 6:07:57 PM
File: _2019-12-24_

Last save: 12/24/2019 6:34:27 PM

Method: AnionsIC2-121919.mtw
Run operator: wet
Analysis number: 24102

Last save: 12/20/2019 11:5

SAMPLE:
: CT/AP
Vial number: 3
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD