

LB106989

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AL

Instrument ID : Konelab

12/31/2019 12:32

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.058	0.0	0.111	
ICB1	-0.002	0.0	0.009	
CCV1	0.997	0.0	0.105	
CCB1	0.005	0.0	0.010	
PB125808BL	-0.003	0.0	0.009	
PB125808BS	1.050	0.0	0.111	
K6434-03	-0.006	0.0	0.009	
K6434-03DUP	0.005	0.0	0.010	
K6434-03MS	0.890	0.0	0.095	
K6434-03MSD	0.921	0.0	0.098	
CCV2	0.994	0.0	0.105	
CCB2	-0.003	0.0	0.009	

N	12
Mean	0.492
SD	0.5169
CV%	105.02

AD

Aquakem v. 7.2AQ1

Results from time period:

Tue Dec 31 11:12:52 2019

Tue Dec 31 12:20:33 2019

Sample Id	Sam/Ctr/c	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0452	mg/l	12/31/2019 11:46:49	
0.1PPM	A	Phenolics-	P	0.0933	mg/l	12/31/2019 11:46:50	
0.25PPM	A	Phenolics-	P	0.2418	mg/l	12/31/2019 11:46:51	
0.50PPM	A	Phenolics-	P	0.4955	mg/l	12/31/2019 11:46:52	
1.0PPM	A	Phenolics-	P	1.0433	mg/l	12/31/2019 11:46:53	
2.0PPM	A	Phenolics-	P	1.9809	mg/l	12/31/2019 11:46:54	
ICV1	S	Phenolics-	P	1.0584	mg/l	12/31/2019 12:17:45	
ICB1	S	Phenolics-	P	-0.0024	mg/l	12/31/2019 12:17:46	
CCV1	S	Phenolics-	P	0.997	mg/l	12/31/2019 12:17:47	
CCB1	S	Phenolics-	P	0.0045	mg/l	12/31/2019 12:17:48	
PB125808BL	S	Phenolics-	P	-0.003	mg/l	12/31/2019 12:17:49	
PB125808BS	S	Phenolics-	P	1.0499	mg/l	12/31/2019 12:17:50	
K6434-03	S	Phenolics-	P	-0.0061	mg/l	12/31/2019 12:17:51	
K6434-03DUP	S	Phenolics-	P	0.0052	mg/l	12/31/2019 12:17:52	
K6434-03MS	S	Phenolics-	P	0.8901	mg/l	12/31/2019 12:17:53	
K6434-03MSD	S	Phenolics-	P	0.9211	mg/l	12/31/2019 12:17:54	
CCV2	S	Phenolics-	P	0.9943	mg/l	12/31/2019 12:17:55	
CCB2	S	Phenolics-	P	-0.0029	mg/l	12/31/2019 12:17:56	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AL

Instrument ID : Konelab

12/31/2019 11:55


Test Phenolics-

Accepted 12/31/2019 11:55

Factor 10.36
Bias 0.009

Coeff. of det. 0.999142

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.014	0.0452	0.0500	
2	Phenol-2	0.018	0.0933	0.1000	
3	Phenol-2	0.033	0.2418	0.2500	
4	Phenol-2	0.057	0.4955	0.5000	
5	Phenol-2	0.110	1.0433	1.0000	
6	Phenol-2	0.200	1.9809	2.0000	