

LB107084

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

Instrument ID : Konelab

1/8/2020 14:25

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	92.984	0.0	0.096	
ICB1	0.608	0.0	0.002	
CCV1	237.830	0.0	0.242	
CCB1	0.565	0.0	0.002	
PB125956BL	0.440	0.0	0.002	
L1047-02	0.449	0.0	0.002	
L1047-02DUP	0.446	0.0	0.002	
PB125958BL	0.403	0.0	0.002	
L1013-06	0.453	0.0	0.002	
L1013-06DUP	0.429	0.0	0.002	
L1016-04	0.496	0.0	0.002	
CCV2	253.475	0.0	0.258	
CCB2	0.554	0.0	0.002	

N	13
Mean	45.318
SD	92.5395
CV%	204.20

Aquakem v. 7.2AQ1

AK

Results from time period:

Wed Jan 08 14:20:25 2020

Wed Jan 08 14:23:49 2020

Sample Id	Sam/Ctr/c/	Test short Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive Cl P	0.5913	µg/l	1/8/2020 12:18:51	
5.OPPBCN	A	Reactive Cl P	5.6388	µg/l	1/8/2020 12:18:52	
10PPBCN	A	Reactive Cl P	10.3637	µg/l	1/8/2020 12:18:53	
50PPBCN	A	Reactive Cl P	49.3304	µg/l	1/8/2020 12:18:54	
100PPBCN	A	Reactive Cl P	99.7033	µg/l	1/8/2020 12:18:55	
250PPBCN	A	Reactive Cl P	248.5199	µg/l	1/8/2020 12:18:56	
500PPBCN	A	Reactive Cl P	500.8527	µg/l	1/8/2020 12:18:57	
ICV1	S	Reactive Cl P	92.9843	µg/l	1/8/2020 14:20:25	
ICB1	S	Reactive Cl P	0.6076	µg/l	1/8/2020 14:20:26	
CCV1	S	Reactive Cl P	237.8297	µg/l	1/8/2020 14:20:27	
CCB1	S	Reactive Cl P	0.5651	µg/l	1/8/2020 14:20:28	
PB125956BL	S	Reactive Cl P	0.4401	µg/l	1/8/2020 14:20:29	
L1047-02	S	Reactive Cl P	0.4487	µg/l	1/8/2020 14:20:30	
L1047-02DUP	S	Reactive Cl P	0.4458	µg/l	1/8/2020 14:20:31	
PB125958BL	S	Reactive Cl P	0.4031	µg/l	1/8/2020 14:20:32	
L1013-06	S	Reactive Cl P	0.453	µg/l	1/8/2020 14:20:33	
L1013-06DUP	S	Reactive Cl P	0.4289	µg/l	1/8/2020 14:20:34	
L1016-04	S	Reactive Cl P	0.4963	µg/l	1/8/2020 14:20:35	
CCV2	S	Reactive Cl P	253.4748	µg/l	1/8/2020 14:23:48	
CCB2	S	Reactive Cl P	0.5539	µg/l	1/8/2020 14:23:49	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

Instrument ID : Konelab

1/8/2020 14:27


Test Ammonia-N

Accepted 1/6/2020 10:06

Factor 6.769
Bias 0.039

Coeff. of det. 0.999293

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	NH3-2PPM	0.052	0.0912	0.1000	
2	NH3-2PPM	0.068	0.1969	0.2000	
3	NH3-2PPM	0.098	0.4031	0.4000	
4	NH3-2PPM	0.185	0.9926	1.0000	
5	NH3-2PPM	0.241	1.3704	1.3333	
6	NH3-2PPM	0.331	1.9791	2.0000	