

LB 107278


7471B

Instrument ID: CV1

Linear

 μ Abs.:

50851



A= 0.0000e+000

B= 1.9669e-004

C= -7.1168e-002

Rho= 0.9999296

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.015	-0.015	284	0.000	284				
0.2	0.200	0.177	-0.023	1260	0.0 %	1260				
2.5	2.500	2.499	-0.001	13068	0.0 %	13068				
5.0	5.000	5.051	0.051	26042	0.0 %	26042				
7.5	7.500	7.558	0.058	38786	0.0 %	38786				
10.0	10.000	9.931	-0.069	50851	0.0 %	50851				

LB107278 INSTRUMENT IDICV1

Method: 7471B

Operator: Admin

Date of Analysis: 21 Jan 2020 12:22:06

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	SD.0	284	-	0.0000	7471B	PPB	21 Jan 2020 12:29:44	S	Std
0.2 - 1	SD.2	1260	-	0.2000	7471B	PPB	21 Jan 2020 12:32:00	S	Std
2.5 - 1	SD.5	13068	-	2.5000	7471B	PPB	21 Jan 2020 12:34:17	S	Std
5.0 - 1	SD.0	26042	-	5.0000	7471B	PPB	21 Jan 2020 12:36:34	S	Std
7.5 - 1	SD.5	38786	-	7.5000	7471B	PPB	21 Jan 2020 12:38:50	S	Std
10.0 - 1	SD.0	50851	-	10.0000	7471B	PPB	21 Jan 2020 12:41:08	S	Std
ICV75 - 1	ICV75	20174	3.8969	-7471B	PPB	21 Jan 2020 13:20:43	U	SMPL	
ICB75 - 1	ICB75	215	-0.0289	-7471B	PPB	21 Jan 2020 13:22:58	U	SMPL	
CCV82 - 1	CCV82	26421	5.1256	-7471B	PPB	21 Jan 2020 13:25:14	U	SMPL	
CCB82 - 1	CCB82	230	-0.0259	-7471B	PPB	21 Jan 2020 13:27:29	U	SMPL	
CRA - 1	CRA	1247	0.1741	-7471B	PPB	21 Jan 2020 13:29:45	U	SMPL	
HighStd - 1	HighStd	51388	10.0364	-7471B	PPB	21 Jan 2020 13:31:59	U	SMPL	
ChkStd - 1	ChkStd	33875	6.5917	-7471B	PPB	21 Jan 2020 13:34:14	U	SMPL	
PB126228BL - 1	PB126228BL	189	-0.0340	-7471B	PPB	21 Jan 2020 13:36:32	U	SMPL	
PB126228BS - 1	PB126228BS	19335	3.7318	-7471B	PPB	21 Jan 2020 13:38:50	U	SMPL	
L1004-10 - 1	BU-02-101720	9231	1.7445	-7471B	PPB	21 Jan 2020 13:41:07	U	SMPL	
L1004-10DUP - 1	BU-02-101720DUP	9680	1.8328	-7471B	PPB	21 Jan 2020 13:43:25	U	SMPL	
L1004-10MS - 1	BU-02-101720MS	32985	6.4167	-7471B	PPB	21 Jan 2020 13:45:41	U	SMPL	
L1004-10MSD - 1	BU-02-101720MSD	33602	6.5380	-7471B	PPB	21 Jan 2020 13:47:58	U	SMPL	
L1014-06 - 1	OR-02-011720	7387	1.3818	-7471B	PPB	21 Jan 2020 13:50:13	U	SMPL	
CCV83 - 1	CCV83	25110	4.8677	-7471B	PPB	21 Jan 2020 13:52:30	U	SMPL	
CCB83 - 1	CCB83	198	-0.0322	-7471B	PPB	21 Jan 2020 13:54:45	U	SMPL	
L1164-06 - 1	QVD-SB-01-0-52	1365	0.1973	-7471B	PPB	21 Jan 2020 13:57:01	U	SMPL	
L1164-07 - 1	QVD-SB-0203-0-11	4162	0.7475	-7471B	PPB	21 Jan 2020 13:59:16	U	SMPL	
L1164-08 - 1	QVD-SB-BC-0-5	3881	0.6922	-7471B	PPB	21 Jan 2020 14:01:32	U	SMPL	
L1181-05 - 1	93OAK-AVE-COMP	2773	0.4743	-7471B	PPB	21 Jan 2020 14:03:47	U	SMPL	
L1187-01 - 1	STOCKPILE	26882	5.2163	-7471B	PPB	21 Jan 2020 14:06:03	U	SMPL	
L1188-01 - 1	B2-0-15	1057	0.1367	-7471B	PPB	21 Jan 2020 14:08:19	U	SMPL	
L1188-03 - 1	B2-15-30	1271	0.1788	-7471B	PPB	21 Jan 2020 14:10:35	U	SMPL	
L1190-02 - 1	TSP-02	49707	9.7057	-7471B	PPB	21 Jan 2020 14:12:52	U	SMPL	
L1004-10LX5 - 1		1451	0.2142	-7471B	PPB	21 Jan 2020 14:15:09	U	SMPL	
L1004-10A - 1		29856	5.8012	-7471B	PPB	21 Jan 2020 14:54:12	U	SMPL	
CCV84 - 1	CCV84	25456	4.9358	-7471B	PPB	21 Jan 2020 14:56:27	U	SMPL	
CCB84 - 1	CCB84	203	-0.0312	-7471B	PPB	21 Jan 2020 14:58:43	U	SMPL	