

LB107293


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

40424



A= 0.0000e+000

B= 2.5097e-004

C= -7.0394e-002

Rho= 0.9997467

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.027	-0.027	171	0.000	171				
0.2	0.200	0.180	-0.020	996	0.0 %	996				
2.5	2.500	2.629	0.129	10757	0.0 %	10757				
5.0	5.000	4.975	-0.025	20103	0.0 %	20103				
7.5	7.500	7.369	-0.131	29641	0.0 %	29641				
10.0	10.000	10.075	0.075	40424	0.0 %	40424				

LB107293 INSTRUMENT ID :CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 Jan 2020 14:47:17

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	171	-	0.0000	7470A	PPB	22 Jan 2020 15:32:58	S	Std
0.2 - 1	50.2	996	-	0.2000	7470A	PPB	22 Jan 2020 15:35:05	S	Std
2.5 - 1	52.5	10757	-	2.5000	7470A	PPB	22 Jan 2020 15:37:13	S	Std
5.0 - 1	55.0	20103	-	5.0000	7470A	PPB	22 Jan 2020 15:39:21	S	Std
7.5 - 1	57.5	29641	-	7.5000	7470A	PPB	22 Jan 2020 15:41:28	S	Std
10.0 - 1	50.0	40424	-	10.0000	7470A	PPB	22 Jan 2020 15:43:34	S	Std
ICV81 - 1	ICV81	16725	4.1271	-	7470A	PPB	22 Jan 2020 15:46:38	U	SMPL
ICB81 - 1	ICB81	780	0.1254	-	7470A	PPB	22 Jan 2020 15:48:44	U	SMPL
CCV90 - 1	CCV90	21225	5.2565	-	7470A	PPB	22 Jan 2020 15:50:49	U	SMPL
CCB90 - 1	CCB90	804	0.1314	-	7470A	PPB	22 Jan 2020 15:52:56	U	SMPL
CRA - 1	CRA	991	0.1783	-	7470A	PPB	22 Jan 2020 15:57:43	U	SMPL
HighStd - 1	HighStd	39164	9.7587	-	7470A	PPB	22 Jan 2020 15:59:49	U	SMPL
ChkStd - 1	ChkStd	27291	6.7789	-	7470A	PPB	22 Jan 2020 16:01:55	U	SMPL
PB126272BL - 1	PBW	779	0.1251	-	7470A	PPB	22 Jan 2020 16:04:01	U	SMPL
PB126272BS - 1	LCSW	16271	4.0132	-	7470A	PPB	22 Jan 2020 16:06:07	U	SMPL
L1004-15 - 1	BU-03-012020	1637	0.3404	-	7470A	PPB	22 Jan 2020 16:08:14	U	SMPL
L1004-15DUP - 1	BU-03-012020DUP	1696	0.3553	-	7470A	PPB	22 Jan 2020 16:10:21	U	SMPL
L1004-15MS - 1	BU-03-012020MS	17228	4.2534	-	7470A	PPB	22 Jan 2020 16:12:29	U	SMPL
L1004-15MSD - 1	BU-03-012020MSD	17856	4.4110	-	7470A	PPB	22 Jan 2020 16:14:36	U	SMPL
L1154-02 - 1	B4-0-15	2143	0.4674	-	7470A	PPB	22 Jan 2020 16:16:43	U	SMPL
CCV91 - 1	CCV91	20489	5.0718	-	7470A	PPB	22 Jan 2020 16:18:49	U	SMPL
CCB91 - 1	CCB91	586	0.0767	-	7470A	PPB	22 Jan 2020 16:20:55	U	SMPL
L1154-04 - 1	B4-15-20	2386	0.5284	-	7470A	PPB	22 Jan 2020 16:25:48	U	SMPL
L1154-06 - 1	B3-0-15	2140	0.4667	-	7470A	PPB	22 Jan 2020 16:27:54	U	SMPL
L1154-08 - 1	B3-15-20	2318	0.5114	-	7470A	PPB	22 Jan 2020 16:30:00	U	SMPL
L1155-03 - 1	RBR-200028	2098	0.4561	-	7470A	PPB	22 Jan 2020 16:32:07	U	SMPL
L1155-06 - 1	RBR-200052	2125	0.4629	-	7470A	PPB	22 Jan 2020 16:34:13	U	SMPL
L1210-07 - 1	Q012-WC1	2378	0.5264	-	7470A	PPB	22 Jan 2020 16:36:20	U	SMPL
L1211-02 - 1	B1-0-15	2479	0.5518	-	7470A	PPB	22 Jan 2020 16:38:27	U	SMPL
L1211-04 - 1	B1-15-30	1802	0.3819	-	7470A	PPB	22 Jan 2020 16:40:34	U	SMPL
PB126273BL - 1	PBW	447	0.0418	-	7470A	PPB	22 Jan 2020 16:42:42	U	SMPL
PB126273BS - 1	LCSW	16192	3.9934	-	7470A	PPB	22 Jan 2020 16:44:48	U	SMPL
CCV92 - 1	CCV92	19395	4.7972	-	7470A	PPB	22 Jan 2020 16:46:54	U	SMPL
CCB92 - 1	CCB92	1	-0.0701	-	7470A	PPB	22 Jan 2020 16:49:00	U	SMPL
L1213-02 - 1	TP-1	1996	0.4305	-	7470A	PPB	22 Jan 2020 16:51:07	U	SMPL
L1213-02DUP - 1	TP-1DUP	2022	0.4371	-	7470A	PPB	22 Jan 2020 16:53:13	U	SMPL
L1213-02MS - 1	TP-1MS	17318	4.2760	-	7470A	PPB	22 Jan 2020 16:55:19	U	SMPL
L1213-02MSD - 1	TP-1MSD	17408	4.2985	-	7470A	PPB	22 Jan 2020 16:57:25	U	SMPL
L1213-04 - 1	TP-2	2168	0.4737	-	7470A	PPB	22 Jan 2020 16:59:32	U	SMPL
L1216-09 - 1	SB-01-COMP	427	0.0368	-	7470A	PPB	22 Jan 2020 17:01:39	U	SMPL
L1216-10 - 1	SB-02-COMP	976	0.1746	-	7470A	PPB	22 Jan 2020 17:03:46	U	SMPL
L1216-11 - 1	SB-03-COMP	940	0.1655	-	7470A	PPB	22 Jan 2020 17:05:54	U	SMPL
L1216-12 - 1	SB-04-COMP	1050	0.1931	-	7470A	PPB	22 Jan 2020 17:08:01	U	SMPL
PB126272TB - 1	PB126226TB	447	0.0418	-	7470A	PPB	22 Jan 2020 17:10:08	U	SMPL
CCV93 - 1	CCV93	19654	4.8622	-	7470A	PPB	22 Jan 2020 17:12:14	U	SMPL
CCB93 - 1	CCB93	370	0.0225	-	7470A	PPB	22 Jan 2020 17:14:20	U	SMPL
PB126273TB - 1	PB126261TB	448	0.0420	-	7470A	PPB	22 Jan 2020 17:16:26	U	SMPL
L1004-15LX5 - 1		340	0.0149	-	7470A	PPB	22 Jan 2020 17:18:33	U	SMPL
L1213-02LX5 - 1		418	0.0345	-	7470A	PPB	22 Jan 2020 17:20:39	U	SMPL
CCV94 - 1	CCV94	19598	4.8482	-	7470A	PPB	22 Jan 2020 17:22:45	U	SMPL
CCB94 - 1	CCB94	445	0.0413	-	7470A	PPB	22 Jan 2020 17:24:51	U	SMPL